

The order of this data package is as follows:

1. Chain-of-Custody/Lab Request
2. Copies of field COCs
3. Validation Report
4. Laboratory analysis

Comments:

[illegible]

[illegible]

Shipping Classification Determination Checklist

Page 6 of 6

Sampling Plan ID/Name: R-48, R-58, R-63 11669

COC: 2018-1645

TEST - Explosives				YES	NO
Samples collected from a WFO area? (TAS -8, 9, 11, 16, 37, 14, 15, 36, 22, 39, 40, and 49)				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Field Test for Explosives Results				YES	NO NA
HE SPOT test result positive. If YES - Do not transport.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
TEST - Chemical Preservation				YES	NO
Samples are chemically preserved?				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Field Team Member Statement				YES	NO NA
Chemical preservation exceeds limits given 40 CFR 136, Table II - Required Containers, Preservation Techniques and Holding Times (footnote 3). If YES - Do not ship.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
TEST - Field Screen				YES	NO
The sample has field screening measurements of alpha and beta activity?				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Sample Activity (dpm/100cm ²)	Shipment Activity (dpm*g/100cm ²)	Sampled Location		YES	NO NA
Alpha detectable	AND Alpha ≥ 160,000	AT TA-1 and adjacent hillsides, TA-21, Acid Canyon, MDA C at TA-50, Area G at TA-54, TA-48, or TA-49		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Alpha ≥ 125	AND Alpha ≥ 1,250,000	AT other locations		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Beta ≥ 1,500	AND Beta ≥ 15,000,000	AT any location		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
The sample Alpha ≥ 16,000,000 dpm*g/100cm ² or Beta ≥ 160,000,000 dpm*g/100cm ² . If YES - Do not ship.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
On the external surface of the sample container, alpha activity ≥ 24 dpm/cm ² , beta activity ≥ 240 dpm/cm ² , or surface activity ≥ 0.5 mR/hr. If YES - Do not ship.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
The sample is tentatively identified as DOT hazard Class 7 (Radioactive). The shipment is labeled <i>Radioactive Material, Excepted Package - Limited Quantity Material - UN2910</i> , based on field screening measurements of alpha and beta activity.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
TEST - Location				YES	NO
Prior analytical measurements of radioactive isotopes are available?				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Sample Activity (pCi/g)	Shipment Activity (pCi)			YES	NO NA
Am-241 ≥ 27 pCi/g	AND Am-241 ≥ 270,000 pCi Total			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Cs-137 ≥ 270 pCi/g	AND Cs-137 ≥ 270,000 pCi Total			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Pu-238 ≥ 27 pCi/g	AND Pu-238 ≥ 270,000 pCi Total			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Pu-239/240 ≥ 27 pCi/g	AND Pu-239/240 ≥ 270,000 pCi Total			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Th-228 ≥ 27 pCi/g	AND Th-228 ≥ 270,000 pCi Total			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
U-234 ≥ 270 pCi/g	AND U-234 ≥ 1,600,000,000 pCi Total			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
U-238 ≥ 270 pCi/g	AND U-238 ≥ unlimited			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
H-3 ≥ 27,000,000 pCi/g	AND H-3 ≥ 27,000,000,000 pCi Total			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Am-241, Pu-238, Pu-239/240, or Th 228 ≥ 27,000,000 pCi; or Cs-137 ≥ 270,000,000 pCi or U-234 ≥ 160,000,000 pCi; or H-3 ≥ 1 Ci. If YES - Do not ship.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
The sample is tentatively identified as DOT hazard Class 7 (Radioactive). The shipment is labeled <i>Radioactive Material, Excepted Package - Limited Quantity of Material - UN2910</i> , based on prior analytical measurements of radioactive isotopes.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
TEST - AK				YES	NO NA
The shippers documented knowledge of the sample positively identifies appropriate labeling.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
Documented Field Team Member Statement				YES	NO NA
The sample is tentatively identified as DOT hazard Class 7 (Radioactive). The shipment is labeled <i>Radioactive Material, Excepted Package - Limited Quantity of Material - UN2910</i> , and the sample is submitted to ARS or RP for hazard classification analysis.				<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>

These samples do not meet the criteria for classification in any hazard class according to regulation OSHA 29 CFR 1910.1200. The sample(s) contained in this shipment have been assigned a tentative proper DOT shipping name, hazard class, identification number, and packing group, based on the shipper's knowledge of the sample:

Hazard Assessment Completed	Date/Time
(Printed Name) Allison Stantfield (Signature) <i>Allison Stantfield</i>	2/7/18 1440

Hazard Assessment Reviewed	Date/Time
(Printed Name) <i>Renee Smith</i> (Signature) <i>Renee Smith</i>	2/7/18 1440

DATA VALIDATION REPORT

Chain Of Custody No. 2018-1648

1. Distribution Of Samples In EDD.

SDG	Analytical Method	Regular Samples	Field Duplicates	Trip Blanks	Field Blanks	Equipment Blanks
ARS1-18-00398	Generic:Low_Level_Tritium	3				

SDG	Analytical Method	Analysis Lot ID	Prep Lot ID	Regular Samples	Field Duplicates	Trip Blanks	Field Blanks	Equipment Blanks	Method Blanks	Matrix Spikes	Matrix Spike Dups	Analytical Spikes	Post-Digestion Spikes	Lab Control Samples	Lab Control Sample Dups	Blank Spike	Blank Spike Dups	Lab Duplicates	Storage Blanks	Preparation Blanks	Reagent Blanks
ARS1-18-00398	Generic:Low_Level_Tritium	ARS1-B18-	ARS1-B18-	3					1					1	1						

2. Distribution Of Analytes In EDD.

Analytical Method	Analytical Method Category	Field Sample ID	Lab Sample ID	Sample Purpose	Target Analytes	Surrogates	Spiked Compounds	TICS
Generic:Low_Level_Tritium	RAD	CAWA-18-129	ARS1-B18-00376-07	PEB	1	0	0	0
Generic:Low_Level_Tritium	RAD	CAWA-18-72	ARS1-B18-00376-04	REG	1	0	0	0
Generic:Low_Level_Tritium	RAD	CAWA-18-75	ARS1-B18-00376-05	REG	1	0	0	0
Generic:Low_Level_Tritium	RAD	CAWA-18-78	ARS1-B18-00376-06	REG	1	0	0	0
Generic:Low_Level_Tritium	RAD	LCS	ARS1-B18-00376-01	LCS	0	0	1	0
Generic:Low_Level_Tritium	RAD	LCSD	ARS1-B18-00376-02	LCSD	0	0	1	0
Generic:Low_Level_Tritium	RAD	MB	ARS1-B18-00376-03	MB	1	0	0	0

3. Are any analytes missing?

No.

4. Were any holding times exceeded?

No.

5. Any contaminants in blanks?

Only results shown in Section 13 'Display Flagged Data' are current as of this report generation. All other sections are valid for the date the COC data was inserted into EIM, and may have changed due to data updates in the intervening time.

DATA VALIDATION REPORT

No.

6. Any surrogate recoveries outside the control limits?

No.

7. Any MS/MSD recoveries or RPDs outside the control limits?

No.

8. Any LCS/LCSD or BS/BSD recoveries or RPDs outside the control limits?

LCS Lab Sample	LCSD Lab	Analytical Method	Parameter Name	Lab Lot ID	Analysis	Sample Matrix	LCS Spike Recovery	LCSD Spike Recovery	Upper Limit	Lower Limit	Upper Rejection Limit	Lower Rejection Limit	RPD	RPD Limit
ARS1-B18-00376-01	ARS1-B18-00376-02	Generic:Low_Level_Tritium	Tritium	ARS1-B18-00376	04-11-2018	W	50.000	77.000	120.00	80.000		10	40.984	

9. Any Field Duplicate RPDs outside the desired limits?

No.

10. Any Lab Duplicate RPDs outside the desired limits?

No.

11. Any required reporting limits exceeded?

No.

Only results shown in Section 13 'Display Flagged Data' are current as of this report generation. All other sections are valid for the date the COC data was inserted into EIM, and may have changed due to data updates in the intervening time.

DATA VALIDATION REPORT

12. Additional Validator's Comments.

13. Display Flagged Data.

Location ID	COC Number	Field Sample ID	Sample Purpose	Analysis Type Code	Analytical Suite	Analytical Method	Parameter Name	Lab Qualifier	Validation Qualifier	Validation Reason Codes	Detect Flag	Lab Result	Lab Units	Report Result	Report Units	Report MDA	Report Uncertainty	Lab Matrix	Sample Date	Percent	Analysis Lot ID	Validation Status Code	Use Flag
R-63	2018-1648	CAWA-18-129	PEB	INIT	RAD	Generic:Low_Level_Tritium	Tritium	QU*	U	R5	N	-0.893	pCi/L	-0.893	pCi/L	2.964	0.872	W	02/07/2018		ARS1-B18-00376	VAL	Y
R-48	2018-1648	CAWA-18-72	REG	INIT	RAD	Generic:Low_Level_Tritium	Tritium	QU*	U	R5	N	-3.252	pCi/L	-3.252	pCi/L	3.129	0.996	W	02/07/2018		ARS1-B18-00376	VAL	Y
R-58	2018-1648	CAWA-18-75	REG	INIT	RAD	Generic:Low_Level_Tritium	Tritium	QU*	U	R5	N	-1.837	pCi/L	-1.837	pCi/L	3.101	0.927	W	02/07/2018		ARS1-B18-00376	VAL	Y
R-63	2018-1648	CAWA-18-78	REG	INIT	RAD	Generic:Low_Level_Tritium	Tritium	QU*	U	R5	N	-0.358	pCi/L	-0.358	pCi/L	2.924	0.860	W	02/07/2018		ARS1-B18-00376	VAL	Y

Reason Code

Description

R5

Analyte is not detected because the amount reported is less than the MDC.

14. Usable Result Count.

Field Sample ID	Location ID	Sample Purpose	Analytical Method	No. Unuseable Records	Total Records
CAWA-18-129	R-63	PEB	Generic:Low_Level_Tritium	0	1
CAWA-18-72	R-48	REG	Generic:Low_Level_Tritium	0	1
CAWA-18-75	R-58	REG	Generic:Low_Level_Tritium	0	1
CAWA-18-78	R-63	REG	Generic:Low_Level_Tritium	0	1



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory

Request Number: 2018-1648

SDG: ARS1-18-00398



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

**Los Alamos National Laboratory
Request: 2018-1648**

Original COC

[illegible]

3 of 82



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

**Los Alamos National Laboratory
Request: 2018-1648**

Case Narrative



ARS International, LLC

Laboratory Analysis Report

ARS1-18-00398

Prepared for:

Los Alamos National Laboratory

**Nita Patel
P.O. Box 1663
MS M992
Los Alamos, NM 87545**

**npatel@lanl.gov
sherwoods@lanl.gov**

**Phone: 505-665-9273
Fax: 505-665-9972**

Project Manager Review

Notes: ARS International, LLC assumes no liability for the use or the interpretation of any analytical results provided other than the cost of the analysis itself. Reproduction of this report in less than full requires the written consent of the client.

Contact Person: Questions regarding this analytical report should be addressed to:

**Project Manager
ProjectManagers@amrad.com**

**Phone: 225.381.2991
Fax: 225.381.2996**



April 18, 2018

Nita Patel
Sherri Sherwood
Los Alamos National Laboratory
505-665-9273
npatel@lanl.gov

ARS SDG: **ARS1-18-00398**
COC Number: **2018-1648**
Charge Code: **ADEP**

Dear Nita,

On February 12, 2018, ARS Aleut Analytical, LLC received four (4) samples to be analyzed for Enriched H-3.

The samples were processed and counted using the appropriate equipment and techniques for these types of analyses. Results of all the analyses are attached in the data package.

The client and QA/QC samples were counted with a count time sufficient to meet quality control parameters for counting equipment and were within acceptance criteria and statistical sound detection limits.

If you have any questions, please do not hesitate to call at 255.381.2991 or email ProjectManagers@amrad.com.

Sincerely,

Susan Leese
Project Management
ARS Aleut Analytical, LLC



**PROJECT SAMPLE IDENTIFICATION
CROSS-REFERENCE
TO ARS SAMPLE LABORATORY IDs**

Client Sample ID	ARS Aleut Analytical Sample ID
2018-1648	ARS1-18-00398-001
2018-1648	ARS1-18-00398-002
2018-1648	ARS1-18-00398-003
2018-1648	ARS1-18-00398-004

SAMPLE RECEIPT/PREP

The samples arrived in good condition. The samples were screened for radioactive contamination as per procedure ARS-062 "Sample Receiving". Turnaround time was set at 40 calendar days.

ANALYTICAL METHODS

Enriched H-3 analysis was performed using ARS-040, "Tritium Assay in Water Samples Using Electrolytic Enrichment".

H-3 screening analysis was performed using ARS-054, "Tritium in Water (EPA 906.0)".

ANALYTICAL RESULTS

ARS1-B18-00376: LCS/LCSD recoveries are below the required 80-120% range. After technical review, data is being reported and qualified with a "Q".

RPD is elevated therefore data is qualified with a "*".



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

ARS Aleut Analytical Laboratory Management's Comments:

"I certify that this sample data package is in compliance with SOW requirements, both technically and for completeness, other than the conditions detailed above. Release of the data contained in this sample data package and the computer-readable EDD, as applicable, submitted on diskette or by modem, has been authorized by the Laboratory Manager or the Manager's designee, as verified by the following signature."

"I certify that this electronic image and all hardcopies produced from this image accurately represent the data and is in compliance with client specific requirements, both technically and for completeness, other than the conditions detailed above or in the sample data package narrative. Release, by submission through email, the data contained in this electronic image and the computer-readable EDD (as applicable), has been authorized by the laboratory Manager/Technical Director or the Manager's designee."

Signature

Laboratory Management, ARS Aleut Analytical

Title

4-23-18

Date



Notes (Case Narrative):

General Comments:

- 1.0) Soil and Sludge analysis are reported on a wet basis or an as received basis unless otherwise indicated.
- 2.0) Modified analysis procedures are procedures that are modified to meet the certain specifications. An example may be the use of a water method to analyze a solid matrix due to the lack of an officially recognized procedure for the analysis of the solid matrix. Modified analyses are indicated by the subsequent addition of "m" to the procedure number (i.e. 900.0M).
- 3.0) All NIOSH method results are reported without blank corrections applied.

Radiochemistry Comments:

- 1.0) All MDA/MDC values are calculated on a sample specific basis.
- 2.0) Data in this report are within the limits of uncertainty specified in the reference method unless otherwise specified.
- 3.0) Total activity is actually total gamma activity and is determined utilizing the prominent gamma emitters from the naturally occurring radioactive decay chains and other prominent radioactive nuclides. Total activity may be lower than the actual total activity due to the extent of secular equilibrium achieved in the various decay chains at the time of analysis. The total activity is not representative of nuclides that emit solely alpha or beta particles.
- 4.0) Ra-228 is determined via secular equilibrium with its daughter, Actinium 228 (Gamma Spectroscopy only).
- 5.0) U-238 is determined via secular equilibrium with its daughter, Thorium 234 (Gamma Spectroscopy only).
- 6.0) All gamma spectroscopy was performed utilizing high purity germanium detectors (**HPGe**).
- 7.0) ARS makes every attempt to match sample density to calibrated density; however, in some cases, it is not practical or possible to do so and data results may be affected (Gamma Spectroscopy only).
- 8.0) Gamma spectroscopy results are calculated values based on the **ORTEC**[®] GammaVision ENV32 Analysis Engine.
- 9.0) ACLASS DOD and ISO 17025 certification applies only to the following analytes and methods: Gross Alpha and Gross Beta (EPA 900, SM7110B&C, SW846 9310); Radium 226 (EPA 903, EPA 903.1, SM 7500 Ra-B, SW846 9315); Radium 228 (EPA 904, SM 7500 Ra-B SW846 9320); Iodine-131(EPA 901.1); Uranium by ICPMS (EPA 200.8); Strontium 89/90 (EPA 905, Eichrom SRW01, HASL 300 Sr-03-RC); Tritium (EPA 906, EPA 906M); Gamma Emitters (EPA 901.1, SM7120B, HASL 300 Ga-01-R); Americium-241, Curium 242/244, Plutonium 239/240 and 241, Thorium 228/230/232, Uranium 234/233 and 238 (Eichrom ACW03 VBS); Lead 210 (HASL 300 Pb-01-RC, Eichrom OTW01); Polonium 210 (HASL 300 Po-01-RC, HASL 300 Po-02-RC); Technetium-99 (Eichrom TCW02, Eichrom TCS01M).

Definitions:

CRDL	Contract Required Detection Limit
CSU	Combined Standard Uncertainty
DLC	Decision Level Concentration (ANSI N42.23) or critical level
DO	Duplicate Original
DUP	Method Duplicate
LCS/LCSD	Laboratory Control Sample/Laboratory Control Sample Duplicate
MDA	Minimum Detectable Activity
MDC	(Minimum Detectable Concentration) minimum concentration of the analyte that ARS can detect utilizing the specific analysis
MBL	Method Blank
MS/MSD	Matrix Spike/Matrix Spike Duplicate
N/A	Not Applicable
NP	Not Provided
NR	Not Referenced
LOD	Limit of Detection
LOQ	Limit of Quantitation
MCL	Maximum Contaminant Level

Data Qualifiers:

B	The analyte is found in both the associated method blank and the sample. This flag indicates probable blank contamination.
D	Sample analysis accomplished through dilution.
J	The reported result is an estimated value above the limit of detection but outside of quantitation range (e.g., matrix interference was observed).
Q	One or more quality control criteria failed (e.g., LCS recovery, surrogate spike recovery, or CCV recovery).
U	Activity is below the MDC, MDA, MDL, or LOD
N	The analyte is a tentatively identified compound using mass spectrometry or any non-customer requested compounds that are tentatively identified.
*	LCS/LCSD or MS/MSD fails RPD criteria.
S	Spike
SC	Subcontracted out to another qualified laboratory
H	Holding time exceeded
E	Exceeds MCL
**	Reporting Limit is higher than MCL; Target cannot be detected



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory

Low Level Tritium by Low Level Liquid Scintillation Counting



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • FAX (225) 381-2996

ARS Sample Delivery Group: ARS1-18-00398

Client Sample ID: CAWA-18-72

Sample Collection Date: 02/07/18

Sample Matrix: Aqueous

Percent Solids: N/A

Request or PO Number: 2018-1648

ARS Sample ID: ARS1-18-00398-001

Date Received: 02/12/18

Report Date: 04/13/18

Radiochemistry

Analysis Description	Analysis Results	CSU +/-1s	MDC	DLC	CRDL	Qual	Analysis Units	Method	Analysis Date/Time	Analysis Technician	Tracer/Chem Recovery
Enriched H-3	-3.252	0.996	3.129	1.522	3.221	QU*	pCi/L	ARS-040/	04/11/18 20:41	MMORGAN	N/A

Notes: ARS Aleut Analytical, LLC assumes no liability for the use or interpretation of any analytical results provided other than the cost of the analysis itself. Reproduction of this report in less than full requires the written consent of the client.

LELAP Certificate# 01949



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • FAX (225) 381-2996

ARS Sample Delivery Group: ARS1-18-00398

Client Sample ID: CAWA-18-75

Sample Collection Date: 02/07/18

Sample Matrix: Aqueous

Percent Solids: N/A

Request or PO Number: 2018-1648

ARS Sample ID: ARS1-18-00398-002

Date Received: 02/12/18

Report Date: 04/13/18

Radiochemistry

Analysis Description	Analysis Results	CSU +/-1s	MDC	DLC	CRDL	Qual	Analysis Units	Method	Analysis Date/Time	Analysis Technician	Tracer/Chem Recovery
Enriched H-3	-1.837	0.927	3.101	1.508	3.221	QU*	pCi/L	ARS-040/	04/12/18 2:23	MMORGAN	N/A

Notes: ARS Aleut Analytical, LLC assumes no liability for the use or interpretation of any analytical results provided other than the cost of the analysis itself. Reproduction of this report in less than full requires the written consent of the client.

LELAP Certificate# 01949



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • FAX (225) 381-2996

ARS Sample Delivery Group: ARS1-18-00398

Client Sample ID: CAWA-18-78

Sample Collection Date: 02/07/18

Sample Matrix: Aqueous

Percent Solids: N/A

Request or PO Number: 2018-1648

ARS Sample ID: ARS1-18-00398-003

Date Received: 02/12/18

Report Date: 04/13/18

Radiochemistry

Analysis Description	Analysis Results	CSU +/-1s	MDC	DLC	CRDL	Qual	Analysis Units	Method	Analysis Date/Time	Analysis Technician	Tracer/Chem Recovery
Enriched H-3	-0.358	0.860	2.924	1.422	3.221	QU*	pCi/L	ARS-040/	04/12/18 8:06	MMORGAN	N/A

Notes: ARS Aleut Analytical, LLC assumes no liability for the use or interpretation of any analytical results provided other than the cost of the analysis itself. Reproduction of this report in less than full requires the written consent of the client.

LELAP Certificate# 01949



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • FAX (225) 381-2996

ARS Sample Delivery Group: ARS1-18-00398

Client Sample ID: CAWA-18-129

Sample Collection Date: 02/07/18

Sample Matrix: Aqueous

Percent Solids: N/A

Request or PO Number: 2018-1648

ARS Sample ID: ARS1-18-00398-004

Date Received: 02/12/18

Report Date: 04/13/18

Radiochemistry

Analysis Description	Analysis Results	CSU +/-1s	MDC	DLC	CRDL	Qual	Analysis Units	Method	Analysis Date/Time	Analysis Technician	Tracer/Chem Recovery
Enriched H-3	-0.893	0.872	2.964	1.442	3.221	QU*	pCi/L	ARS-040/	04/12/18 13:49	MMORGAN	N/A

Notes: ARS Aleut Analytical, LLC assumes no liability for the use or interpretation of any analytical results provided other than the cost of the analysis itself. Reproduction of this report in less than full requires the written consent of the client.

LELAP Certificate# 01949



QC Results per Analytical Batch

Analytical Batch	ARS1-B18-00376
SDG	ARS1-18-00398
Analysis	Low Level Tritium by Electrolytic Enrichment
Analysis Test Method	ARS-040/
Analysis Code	LSC-LLH3-AQ
Report Units	pCi/L

Acceptable QC Performance Ranges

QC Sample Type	Performance Items and Ranges
Laboratory Control Sample	Recovery (%): > 80 < 120
Matrix Spike	Recovery (%): > 60 < 140
Duplicate	Replicate Error Ratio (RER): < 1
	Duplicate Error Ratio (DER): < 3
	Relative Percent Difference (RPD %): ≤ 25

Laboratory Control Sample			Analysis Date	04/11/18 09:14	Analysis Technician	MMORGAN	
Analysis Batch Sample ID	QC Type	Analyte	Results	CSU (1s)	Expected Value	LCS Rec (%)	MDC
ARS1-B18-00376-02	LCS	ENRICHED H-3	25.060	3.988	32.459	77.2	3.289

Duplicate RER/DER/RPD			Analysis Date	04/11/18 09:14	Analysis Technician	MMORGAN	
Analyte	Results LCS	CSU LCS (1s)	Results LCSD	CSU LCSD (1s)	RER	DER	RPD
ENRICHED H-3	16.536	2.737	25.060	3.988	1.268	1.762	41.0

Method Blank		Analysis Date	04/11/18 14:57	Analysis Technician	MMORGAN	
Analysis Batch Sample ID	QC Type	Analyte	Results	CSU (1s)	MDC	Qual
ARS1-B18-00376-03	MBL	ENRICHED H-3	-1.480	0.884	2.981	QU*

Notes: ARS Aleut Analytical, LLC assumes no liability for the use or interpretation of any analytical results provided other than the cost of the analysis itself. Reproduction of this report in less than full requires the written consent of the client.

LELAP Certificate# 01949



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory

**Low Level Tritium
by
Low Level Liquid
Scintillation Counting**

Laboratory Records

Analysis Batch ID ARS1-B18-00376												
Method		ARS-040		Analysis		LSC-LLH3-AQ		Matrix		AQ		
Description				Low Level Tritium by Electrolytic Enrichment								
ABatch Sample ID	Type	Blind Iso1	Blind Iso2	Blind Iso3	SDG	FR	Run	Prep Code	Filtered	Client ID	Group Name	Lab Deadline
ARS1-B18-00376-01	LCS											
ARS1-B18-00376-02	LCSD											
ARS1-B18-00376-03	MBL											
ARS1-B18-00376-04	TRG				ARS1-18-00398	001	1			CAWA-18-72		03/21/18
ARS1-B18-00376-05	TRG				ARS1-18-00398	002	1			CAWA-18-75		03/21/18
ARS1-B18-00376-06	TRG				ARS1-18-00398	003	1			CAWA-18-78		03/21/18
ARS1-B18-00376-07	TRG				ARS1-18-00398	004	1			CAWA-18-129		03/21/18
ARS1-B18-00376-08	TRG				ARS1-18-00402	001	1			CAWA-18-34		03/21/18
ARS1-B18-00376-09	TRG				ARS1-18-00403	001	1			CrPZ-2-18-151284		03/21/18
ARS1-B18-00376-10	TRG				ARS1-18-00403	002	1			CrPZ-5-18-151305		03/21/18
ARS1-B18-00376-11	TRG				ARS1-18-00404	001	1			CAWA-18-81		03/21/18

LCS Report

Analytical Batch: ARS1-B18-00376

Blind ID	ABatch Sample ID	Blind Group	Std ID	Isotope	Exp Addition (g)	Expected Value (pCi/g)	Empty Wt (g)	Gross Wt (g)	Net Wt (g)	Expected Value CT (pCi/g)	Mid Point Count Date	Known Value (pCi)	User ID	Mod Date
B-25085	ARS1-B18-00376-01	B-H3	S-0324	H-3	5	2.46675	17.0006	22.0082	5.0076	2.45537	04/11/2018	12.29552	MMORGAN	03/12/2018
B-25086	ARS1-B18-00376-02	B-H3	S-0324	H-3	5	2.46675	16.7401	21.7319	4.9918	2.45537	04/11/2018	12.25673	MMORGAN	03/12/2018

Tritium Assay in Water Samples Using Electrolytic Enrichment

Preparation Date: 03/22/2018 14:36
Prepared By: MMORGAN

Procedure Data		Type	SDG/Fraction	Tare Wt of Electrolysis Cell & Electrodes	Tare Wt Reservoir	Gross Weight of Sample Reservoir	Wt Sodium Peroxide	Gross Sample Added	Electrolysis Start Date & Time	Start AMP	Start Bath (C)	Electrolysis End Date/Time	End Bath (C)
AR51-B18-00376-01	LCS			324.8700	255.4000	631.2200	1.5000	375.8200	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-02	LCSD			326.2200	209.8400	587.4500	1.5000	377.6100	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-03	MBL			321.9400	226.2200	604.5700	1.5000	378.3500	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-04	TRG	ARS1-18-00398-001		325.5100	213.2500	589.2800	1.5000	376.0300	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-05	TRG	ARS1-18-00398-002		327.9900	214.2000	595.2700	1.5000	381.0700	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-06	TRG	ARS1-18-00398-003		317.1200	222.5400	601.3200	1.5000	378.7800	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-07	TRG	ARS1-18-00398-004		323.4400	219.3500	598.8000	1.5000	379.4500	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-08	TRG	ARS1-18-00402-001		323.8400	222.0800	598.6300	1.5000	376.5500	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-09	TRG	ARS1-18-00403-001		323.7000	232.8200	610.7800	1.5000	377.9600	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-10	TRG	ARS1-18-00403-002		323.4100	224.3600	604.2200	1.5000	379.8600	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000
AR51-B18-00376-11	TRG	ARS1-18-00404-001		320.2300	209.7000	585.6500	1.5000	375.9500	3/22/2018 3:36:00 PM	5.0000	2.0000	4/9/2018 11:00:00 AM	2.0000

Tritium Assay in Water Samples Using Electrolytic Enrichment

Procedure Data		End Wt of Cell + Resv + Sample	Gross Sample Recovered	Enrichment Factor	Tare Wt Cryo-distill flask	Gross Wt flask + Sample	Recovered Water	Tare Weight of LSC Vial	Vial + Sample	Net Sample	Gross Wt Vial + Dead Water If used	Net Dead Water Added
ABatch Sample ID	Type											
ARS1-B18-00376-01	LCS	595.3600	15.0900	24.9052	116.0800	129.9900	13.9100	6.6000	16.6700	10.0700	16.6700	0.0000
ARS1-B18-00376-02	LCSD	551.9300	15.8700	23.7940	125.6900	139.1300	13.4400	6.6800	16.7100	10.0300	16.7100	0.0000
ARS1-B18-00376-03	MBL	563.6000	15.4400	24.5045	109.4200	122.9700	13.5500	6.5500	16.5500	10.0000	16.5500	0.0000
ARS1-B18-00376-04	TRG	554.5600	15.8000	23.7994	126.3400	139.9800	13.6400	6.6300	16.6300	10.0000	16.6300	0.0000
ARS1-B18-00376-05	TRG	557.9000	15.7100	24.2565	129.8100	143.5400	13.7300	6.5700	16.5700	10.0000	16.5700	0.0000
ARS1-B18-00376-06	TRG	555.3500	15.6900	24.1415	120.9300	134.5800	13.6500	6.6900	16.7200	10.0300	16.7200	0.0000
ARS1-B18-00376-07	TRG	557.9300	15.1400	25.0627	122.8500	135.9500	13.1000	6.6000	16.6200	10.0200	16.6200	0.0000
ARS1-B18-00376-08	TRG	562.2200	16.3000	23.1012	128.5500	142.4400	13.8900	6.6000	16.6100	10.0100	16.6100	0.0000
ARS1-B18-00376-09	TRG	573.1500	16.6300	22.7276	99.9500	114.1300	14.1800	6.6200	16.6300	10.0100	16.6300	0.0000
ARS1-B18-00376-10	TRG	559.8200	12.0500	31.5237	110.4500	119.9300	9.4800	6.5600	15.9100	9.3500	17.2500	1.3400
ARS1-B18-00376-11	TRG	546.1500	16.2200	23.1782	106.8000	120.4500	13.6500	6.6300	16.6400	10.0100	16.6400	0.0000

ARS-040
Tritium Assay in Water Samples Using Electrolytic Enrichment

ARS Aleut Analytical, LLC
Baton Rouge Laboratory

Procedure Data					
ABatch Sample ID	Type	Tare Wt b/f Cocktail	Gross Wt Vial + Cocktail	Net Wt of Cocktail Added	
ARS1-B18-00376-01	LCS	16.6700	26.7600	10.0900	
ARS1-B18-00376-02	LCSD	16.7100	26.7200	10.0100	
ARS1-B18-00376-03	MBL	16.5500	26.6000	10.0500	
ARS1-B18-00376-04	TRG	16.6300	26.6800	10.0500	
ARS1-B18-00376-05	TRG	16.5700	26.5900	10.0200	
ARS1-B18-00376-06	TRG	16.7200	26.7500	10.0300	
ARS1-B18-00376-07	TRG	16.6200	26.6400	10.0200	
ARS1-B18-00376-08	TRG	16.6100	26.6400	10.0300	
ARS1-B18-00376-09	TRG	16.6300	26.6400	10.0100	
ARS1-B18-00376-10	TRG	17.2500	27.2500	10.0000	
ARS1-B18-00376-11	TRG	16.6400	26.6500	10.0100	

ARS-040
Tritium Assay in Water Samples Using Electrolytic Enrichment

ARS Aleut Analytical, LLC
Baton Rouge Laboratory

Reagent Amounts				
ABatch Sample ID	Type	SDG/Fraction	14.2.12 DISTILLAT - Ionize & add O to electrolysis - Sodium Peroxide (granular) Reagent Grade (g)	14.3.22 DISTILLATION - Add scint cocktail - Ultima Gold LLT Reagent Grade (mL)
ARS1-B18-00376-01	LCS		1.50	10.00
ARS1-B18-00376-02	LCSD		1.50	10.00
ARS1-B18-00376-03	MBL		1.50	10.00
ARS1-B18-00376-04	TRG	ARS1-18-00398-001	1.50	10.00
ARS1-B18-00376-05	TRG	ARS1-18-00398-002	1.50	10.00
ARS1-B18-00376-06	TRG	ARS1-18-00398-003	1.50	10.00
ARS1-B18-00376-07	TRG	ARS1-18-00398-004	1.50	10.00
ARS1-B18-00376-08	TRG	ARS1-18-00402-001	1.50	10.00
ARS1-B18-00376-09	TRG	ARS1-18-00403-001	1.50	10.00
ARS1-B18-00376-10	TRG	ARS1-18-00403-002	1.50	10.00
ARS1-B18-00376-11	TRG	ARS1-18-00404-001	1.50	10.00

Tritium Assay in Water Samples Using Electrolytic Enrichment

Reagent Tracking	
Procedure Section	Reagent ID
14.2.12 DISTILLAT - Ionize & add O to electrolysis	R17-00926
14.3.22 DISTILLATION - Add scint cocktail	R17-00665

Protocol# 8 - Low Level H3.lsa

User: ARS

Assay Definition

Assay Description:
LLH3 Assay in DPM Mode
Assay Type: DPM (Single)
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\ARS\Low Level H3\20180410_1319
Raw Results Path: C:\Packard\Tricarb\Results\ARS\Low Level H3\20180410_1319\20180410_1319.results
RTF File Name: C:\Packard\Tricarb\Results\ARS\Low Level H3\20180410_1319\LLH3.rtf
Comma-Delimited File Name: C:\Packard\Tricarb\Results\ARS\Low Level H3\20180410_1319\LLH3 Results.csv
Assay File Name: C:\Packard\Tricarb\Assays\Low Level H3.lsa

Count Conditions

Nuclide: Low Level H3
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set:
Low Energy: LLH3 10ml
Count Time (min): 330.00
Count Mode: Low Level
Assay Count Cycles: 1
Number of Vials/Sample: 1
Repeat Sample Count: 1
Calculate % Reference: Off

Background Subtract


Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL
A	2.0	18.6
B	0.0	2000.0
C	0.0	2000.0

Count Corrections

Static Controller: On	Luminescence Correction: Off	GCT: Off
Colored Samples: n/a	Heterogeneity Monitor: Off	PAC: Disabled
Coincidence Time (nsec): 18	Delay Before Burst (nsec): 75	PAC Strength: n/a
		Auxiliary Spectrum: n/a

Cycle 1 Results
Quench Curve Block Data



Date Acquired: 08/18/2017
 Date Modified:
 LLH3 10ml in A

tSIE/AEC	Count Efficiency (%)
380.12	26.09
307.45	23.20
245.66	20.44
188.72	17.73
151.46	14.93
138.74	13.42
120.63	11.00
99.95	9.01
90.27	7.44

P#	S#	SMPL ID	CPMA	DPM1	tSIE	Eff Nucl	In A	Count Time	DATE	TIME	MESSAGES
8	1	BACKGROUND	1.317	6.22	261.80	21.16		330.00	4/10/2018	7:02:30 PM	

Protocol# 8 - Low Level H3.lsa

8	2	B18-00376-01	2.942	13.88	262.76	21.20	330.00	4/11/2018	12:45:53 AM
8	3	B18-00376-02	3.619	17.32	255.95	20.90	330.00	4/11/2018	6:29:16 AM
8	4	B18-00376-03	1.167	5.48	265.20	21.31	330.00	4/11/2018	12:12:18 PM
8	5	B18-00376-04	1.003	4.76	259.85	21.07	330.00	4/11/2018	5:55:49 PM
8	6	B18-00376-05	1.138	5.45	255.40	20.87	330.00	4/11/2018	11:38:04 PM
8	7	B18-00376-06	1.280	5.77	284.28	22.16	330.00	4/12/2018	5:20:43 AM
8	8	B18-00376-07	1.226	5.83	258.85	21.03	330.00	4/12/2018	11:03:43 AM
8	9	B18-00376-08	2.016	9.57	259.55	21.06	330.00	4/12/2018	4:47:00 PM
8	10	B18-00376-09	7.944	37.88	257.70	20.97	330.00	4/12/2018	10:30:04 PM
8	11	B18-00376-10	1.805	8.77	248.81	20.58	330.00	4/13/2018	4:13:28 AM
8	12	B18-00376-11	1.223	5.86	255.40	20.87	330.00	4/13/2018	9:56:52 AM



LSC Instrument Data Transfer Report

\\Tricarb\ars\Low Level H3\20180410

Batch Sample ID		Non-BKG Samples Transferred		Samples Eligible To Save								
ARS1-B18-00376		11		11								
LJMS Batch Sample ID	LSC P#	LSC PID	LSC SW#	LSC SMPL_ID	LSC Count Date	LSC CPMA	LSC ISIE	LSC EFF	LSC Count Dur	Analyte Batch	LJMS SDG	LJMS Run
BKG	8		1	BACKGROUND	04/10/18 19:02	1.32	261.80	21.1600	330.00	ARS1-B18-00376		
ARS1-B18-00376-01	8		2	B18-00376-01	04/11/18 00:45	2.94	262.76	21.2000	330.00	ARS1-B18-00376		
ARS1-B18-00376-02	8		3	B18-00376-02	04/11/18 06:29	3.62	255.95	20.9000	330.00	ARS1-B18-00376		
ARS1-B18-00376-03	8		4	B18-00376-03	04/11/18 12:12	1.17	265.20	21.3100	330.00	ARS1-B18-00376		
ARS1-B18-00376-04	8		5	B18-00376-04	04/11/18 17:55	1.00	259.85	21.0700	330.00	ARS1-B18-00376	ARS1-18-00398	1
ARS1-B18-00376-05	8		6	B18-00376-05	04/11/18 23:38	1.14	255.40	20.8700	330.00	ARS1-B18-00376	ARS1-18-00398	1
ARS1-B18-00376-06	8		7	B18-00376-06	04/12/18 05:20	1.28	284.28	22.1600	330.00	ARS1-B18-00376	ARS1-18-00398	1
ARS1-B18-00376-07	8		8	B18-00376-07	04/12/18 11:03	1.23	258.85	21.0300	330.00	ARS1-B18-00376	ARS1-18-00398	1
ARS1-B18-00376-08	8		9	B18-00376-08	04/12/18 16:47	2.02	259.55	21.0600	330.00	ARS1-B18-00376	ARS1-18-00402	1
ARS1-B18-00376-09	8		10	B18-00376-09	04/12/18 22:30	7.94	257.70	20.9700	330.00	ARS1-B18-00376	ARS1-18-00403	1
ARS1-B18-00376-10	8		11	B18-00376-10	04/13/18 04:13	1.81	248.81	20.5800	330.00	ARS1-B18-00376	ARS1-18-00403	1
ARS1-B18-00376-11	8		12	B18-00376-11	04/13/18 09:56	1.22	255.40	20.8700	330.00	ARS1-B18-00376	ARS1-18-00404	1

ARS-040 Calculation Results			
ARS1-B18-00376			
ACF	1		
UCF	2.22		
Sys Error	0.15		

AnalysisCode	ABatchSampleID	Initial_Mass_sample_g	Mass_Na2O2_added_g	Final_mass_electrolyzed_sample_NaOH_g	Mass_equivalent_NaOH_g	Final_Mass_Electrolyzed_sample_g	VolumeFactor_X	Enrichment_Factor_Y
LSC-LLH3-AQ	ARS1-B18-00376-01	375.820	1.500	15.090	1.539	13.551	0.036	21.880
LSC-LLH3-AQ	ARS1-B18-00376-02	377.610	1.500	15.870	1.539	14.331	0.038	20.829
LSC-LLH3-AQ	ARS1-B18-00376-03	378.350	1.500	15.440	1.539	13.901	0.037	21.489
LSC-LLH3-AQ	ARS1-B18-00376-04	376.030	1.500	15.800	1.539	14.261	0.038	20.844
LSC-LLH3-AQ	ARS1-B18-00376-05	381.070	1.500	15.710	1.539	14.171	0.037	21.241
LSC-LLH3-AQ	ARS1-B18-00376-06	378.780	1.500	15.690	1.539	14.151	0.037	21.147
LSC-LLH3-AQ	ARS1-B18-00376-07	379.450	1.500	15.140	1.539	13.601	0.036	22.005
LSC-LLH3-AQ	ARS1-B18-00376-08	376.550	1.500	16.300	1.539	14.761	0.039	20.191
LSC-LLH3-AQ	ARS1-B18-00376-09	377.960	1.500	16.630	1.539	15.091	0.040	19.838
LSC-LLH3-AQ	ARS1-B18-00376-10	379.860	1.500	12.050	1.539	10.511	0.028	28.220
LSC-LLH3-AQ	ARS1-B18-00376-11	375.950	1.500	16.220	1.539	14.681	0.039	20.266

ARS-040 Calculation Results

ARS1-B18-00376

ACF	1
UCF	2.22
Sys Error	0.15

AnalysisCode	ABatchSampleID	Average_Sample_CPM	Bkg_CPM	tSIE	Detector_Eff_decimal	Aliquot	AliqUnits	Activity_reference_date	Start_Date_of_Count	Sample_Count	Duration_min
LSC-LLH3-AQ	ARS1-B18-00376-01	2.942	1.317	262.760	0.212	0.01007	L	4/27/2017	4/11/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-02	3.619	1.317	255.950	0.209	0.01003	L	4/27/2017	4/11/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-03	1.167	1.317	265.200	0.213	0.01000	L	3/22/2018	4/11/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-04	1.003	1.317	259.850	0.211	0.01000	L	2/7/2018	4/11/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-05	1.138	1.317	255.400	0.209	0.01000	L	2/7/2018	4/11/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-06	1.280	1.317	284.280	0.222	0.01003	L	2/7/2018	4/12/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-07	1.226	1.317	258.850	0.210	0.01002	L	2/7/2018	4/12/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-08	2.016	1.317	259.550	0.211	0.01001	L	2/6/2018	4/12/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-09	7.944	1.317	257.700	0.210	0.01001	L	2/1/2018	4/12/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-10	1.805	1.317	248.810	0.206	0.00935	L	2/1/2018	4/13/2018		330.000
LSC-LLH3-AQ	ARS1-B18-00376-11	1.223	1.317	255.400	0.209	0.01001	L	2/6/2018	4/13/2018		330.000

ARS-040 Calculation Results

ARS1-B18-00376

ACF	1
UCF	2.22
Sys Error	0.15

AnalysisCode	ABatchSampleID	Total_Bkg_Count	Duration_min	DF	Sample_Activity_Conc	Standard_Counting_Uncertainty	CU_1	CSU_1	CU_1_96	CSU_1_96	MDC	DLC	ActivityReportUnits
LSC-LLH3-AQ	ARS1-B18-00376-01	330.000	0.94766	0.94766	16.536	1.156	1.156	2.737	2.266	5.364	3.075	1.496	pCi
LSC-LLH3-AQ	ARS1-B18-00376-02	330.000	0.94766	0.94766	25.060	1.331	1.331	3.988	2.609	7.816	3.289	1.600	pCi
LSC-LLH3-AQ	ARS1-B18-00376-03	330.000	0.99692	0.99692	-1.480	0.856	0.856	0.884	1.678	1.733	2.981	1.450	pCi
LSC-LLH3-AQ	ARS1-B18-00376-04	330.000	0.99034	0.99034	-3.252	0.868	0.868	0.996	1.702	1.952	3.129	1.522	pCi
LSC-LLH3-AQ	ARS1-B18-00376-05	330.000	0.99019	0.99019	-1.837	0.885	0.885	0.927	1.735	1.817	3.101	1.508	pCi
LSC-LLH3-AQ	ARS1-B18-00376-06	330.000	0.99019	0.99019	-0.358	0.859	0.859	0.860	1.683	1.686	2.924	1.422	pCi
LSC-LLH3-AQ	ARS1-B18-00376-07	330.000	0.99019	0.99019	-0.893	0.861	0.861	0.872	1.688	1.708	2.964	1.442	pCi
LSC-LLH3-AQ	ARS1-B18-00376-08	330.000	0.98912	0.98912	7.472	1.074	1.074	1.552	2.105	3.043	3.230	1.571	pCi
LSC-LLH3-AQ	ARS1-B18-00376-09	330.000	0.98912	0.98912	72.474	1.832	1.832	11.024	3.591	21.608	3.304	1.607	pCi
LSC-LLH3-AQ	ARS1-B18-00376-10	330.000	0.98912	0.98912	4.093	0.816	0.816	1.021	1.599	2.001	2.534	1.233	pCi
LSC-LLH3-AQ	ARS1-B18-00376-11	330.000	0.98989	0.98989	-1.010	0.943	0.943	0.955	1.848	1.872	3.248	1.580	pCi


ARS-040 Calculation Results

ARS1-B18-00376

ACF	1
UCF	2.22
Sys Error	0.15

AnalysisCode	ABatchSampleID	AliquotReportUnits	UserID	ModDate
LSC-LLH3-AQ	ARS1-B18-00376-01	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-02	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-03	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-04	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-05	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-06	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-07	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-08	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-09	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-10	L	AMRAD\mmorgan	4/13/2018
LSC-LLH3-AQ	ARS1-B18-00376-11	L	AMRAD\mmorgan	4/13/2018

Technical Note Report

 ARS Aleut Analytical	ABATCH	ARS1-B18-00376		
	Analysis Code	LSC-LLH3-AQ		
	Procedure No	ARS-040		
	Matrix	AQ		
#		Date	Technical Note	User ID
1		04/13/2018 1:37 PM	LCS/LCSD recoveries are below the required 80-120% range. After technical review, data is being reported and qualified with a "Q". RPD is elevated therefore data is qualified with a "**".	SLEESE

Liquid Scintillation Count Log

Date	Time	ARS Sample I.D. Number	Batch Fraction Number	Liquid Scintillation File Number	Technician Initials	Notes Identifier
3/26/2018	11:00	B18-00370	4	2113	MM	
3/26/2018	11:00	B18-00370	5	2113	MM	
3/26/2018	11:00	B18-00370	6	2113	MM	
3/26/2018	11:00	B18-00370	8	2113	MM	
3/26/2018	11:00	B18-00370	9	2113	MM	
3/26/2018	11:00	B18-00370	10	2113	MM	
3/27/2018	15:00	SNC163	QA	QA	MM	
3/27/2018	15:00	Background	N/A	N/A	MM	
3/27/2018	15:00	B18-00683	4	650	MM	
3/27/2018	15:00	B18-00683	5	650	MM	
3/28/2018	14:30	SNC163	QA	QA	MM	
3/28/2018	16:00	Background	N/A	N/A	MM	
3/28/2018	16:00	B18-00305	1	951	MM	
3/28/2018	16:00	B18-00305	2	951	MM	
3/28/2018	16:00	B18-00305	3	951	MM	
3/28/2018	16:00	B18-00305	4	951	MM	
3/28/2018	16:00	B18-00305	5	951	MM	
3/28/2018	16:00	B18-00305	6	951	MM	
3/29/2018	8:19	SNC163	QA	QA	BJS	
4/4/2018	15:30	SNC163	QA	QA	MM	
4/4/2018	16:00	Background	NA	NA	MM	
4/4/2018	16:00	B18-00372	1	1854	MM	
4/4/2018	16:00	B18-00372	2	1854	MM	
4/4/2018	16:00	B18-00372	3	1854	MM	
4/4/2018	16:00	B18-00372	4	1854	MM	
4/4/2018	16:00	B18-00372	5	1854	MM	
4/4/2018	16:00	B18-00372	6	1854	MM	
4/4/2018	16:00	B18-00372	7	1854	MM	
4/4/2018	16:00	B18-00372	8	1854	MM	
4/4/2018	16:00	B18-00372	9	1854	MM	
4/4/2018	16:00	B18-00372	10	1854	MM	
4/6/2018	16:49	SNC163	QA	QA	BJS	
4/9/2018	10:00	SNC163	QA	QA	MM	
4/9/2018	10:00	B18-00763	4	1135	MM	
4/9/2018	10:00	B18-00763	5	1135	MM	
4/9/2018	10:00	B18-00763	6	1135	MM	
4/9/2018	10:00	B18-00763	7	1135	MM	
4/9/2018	10:00	B18-00763	8	1135	MM	
4/9/2018	10:00	B18-00763	9	1135	MM	
4/9/2018	10:00	B18-00763	10	1135	MM	
4/9/2018	10:00	B18-00763	11	1135	MM	
4/9/2018	10:00	B18-00763	12	1135	MM	
4/9/2018	10:00	B18-00763	13	1135	MM	
4/10/2018	13:00	SNC163	QA	QA	MM	
4/10/2018	14:00	Background	N/A	N/A	MM	
4/10/2018	14:00	B18-00376	1	1319	MM	
4/10/2018	14:00	B18-00376	2	1319	MM	
4/10/2018	14:00	B18-00376	3	1319	MM	
4/10/2018	14:00	B18-00376	4	1319	MM	
4/10/2018	14:00	B18-00376	5	1319	MM	
4/10/2018	14:00	B18-00376	6	1319	MM	

Liquid Scintillation Count Log

Date	Time	ARS Sample I.D. Number	Batch Fraction Number	Liquid Scintillation File Number	Technician Initials	Notes Identifier
4/10/2018	14:00	B18-00376	7	1319	MM	
4/10/2018	14:00	B18-00376	8	1319	MM	
4/10/2018	14:00	B18-00376	9	1319	MM	
4/10/2018	14:00	B18-00376	10	1319	MM	
4/10/2018	14:00	B18-00376	11	1319	MM	



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports





for

Los Alamos National Laboratory

**Low Level Tritium
by
Low Level Liquid
Scintillation Counting**

Control Charts

QC Chart




3H Efficiency

Total # pts : 114
Valid # pts : 30
Mean : 60.64
SD : 0.13

Date	Value	Include
Feb 13, 2018	60.56	X
Feb 14, 2018	60.57	X
Feb 14, 2018	60.75	X
Feb 16, 2018	60.79	X
Feb 22, 2018	60.76	X
Feb 25, 2018	60.75	X
Feb 26, 2018	60.65	X
Feb 28, 2018	60.69	X
Mar 03, 2018	60.63	X
Mar 06, 2018	60.67	X
Mar 11, 2018	60.76	X
Mar 12, 2018	60.52	X
Mar 14, 2018	60.69	X
Mar 14, 2018	60.65	X
Mar 17, 2018	60.73	X
Mar 19, 2018	60.63	X
Mar 21, 2018	60.74	X
Mar 22, 2018	60.80	X
Mar 23, 2018	60.63	X
Mar 26, 2018	60.85	X
Mar 28, 2018	60.53	X
Mar 28, 2018	60.71	X
Mar 30, 2018	60.48	X
Apr 01, 2018	60.42	X
Apr 04, 2018	60.63	X
Apr 04, 2018	60.71	X
Apr 07, 2018	60.59	X
Apr 09, 2018	60.56	X
Apr 10, 2018	60.29	X
Apr 13, 2018	60.45	X

3H Efficiency : 114
Total # pts : 30
Valid # pts : 60.64
Mean : 60.64
SD : 0.13




3H Background

Total # pts : 112
Valid # pts : 30
Mean : 1.89
SD : 0.17

Date	Value	Include
Feb 13, 2018	1.75	X
Feb 14, 2018	2.03	X
Feb 14, 2018	1.80	X
Feb 16, 2018	1.92	X
Feb 22, 2018	2.13	X
Feb 25, 2018	1.87	X
Feb 26, 2018	1.80	X
Feb 28, 2018	2.05	X
Mar 03, 2018	1.73	X
Mar 06, 2018	1.63	X
Mar 11, 2018	1.67	X
Mar 12, 2018	1.90	X
Mar 14, 2018	1.80	X
Mar 14, 2018	2.07	X
Mar 17, 2018	1.87	X
Mar 19, 2018	2.12	X
Mar 21, 2018	1.72	X
Mar 22, 2018	2.13	X
Mar 23, 2018	1.77	X
Mar 26, 2018	1.77	X
Mar 28, 2018	1.80	X
Mar 28, 2018	1.83	X
Mar 30, 2018	1.58	X
Apr 01, 2018	1.87	X
Apr 04, 2018	1.93	X
Apr 04, 2018	2.00	X
Apr 07, 2018	2.25	X
Apr 09, 2018	2.00	X
Apr 10, 2018	1.77	X
Apr 13, 2018	2.12	X

3H Background
Total # pts : 112
Valid # pts : 30
Mean : 1.89
SD : 0.17





2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory

Tritium- Screening by Low Level Liquid Scintillation Counting



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory

**Tritium-Screening
by
Low Level Liquid
Scintillation Counting**

Samples



Sample ID:	Count Time	CPMA	Background CPMA	Eff Nucl In A	Aliquot (grams)	ACTIVITY	units	MDA	Sample Must be analyzed as LSC-A-001
ARS1-B18-00339-04	120	0.982	1.274	21.72	10.00	-60.558	pCi/L	126.6423	NO
ARS1-B18-00339-05	120	1.081	1.274	21.74	10.00	-39.989	pCi/L	126.5258	NO
ARS1-B18-00339-06	120	0.968	1.274	22.32	10.00	-61.755	pCi/L	123.2379	NO
ARS1-B18-00339-07	120	1.107	1.274	22.1	10.00	-34.039	pCi/L	124.4647	NO
ARS1-B18-00339-08	120	1.189	1.274	21.83	10.00	-17.539	pCi/L	126.0041	NO
ARS1-B18-00339-09	120	2.125	1.274	21.77	10.00	176.083	pCi/L	126.9514	YES, analyze by LSC-A-001.
ARS1-B18-00339-10	120	2.751	1.274	22.1	10.00	301.048	pCi/L	124.4647	YES, analyze by LSC-A-001.
ARS1-B18-00339-11	120	1.127	1.274	22.16	10.00	-29.881	pCi/L	124.1277	NO
ARS1-B18-00339-12	120	1.062	1.274	21.85	10.00	-43.705	pCi/L	125.8888	NO
B18-00339-09 recount	120	1.536	1.118	20.77	10.00	90.654	pCi/L	124.3715	NO
B18-00339-10 recount	120	1.144	1.118	20.93	10.00	5.596	pCi/L	123.4207	NO
						If activity is > 150 pCi/L, contact client before running by electrolytic enrichment			



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory


**Tritium-Screening
by
Low Level Liquid
Scintillation Counting**

**Laboratory
Records**

Analytical Batch Report

Analysis Batch ID ARS1-B18-00339





ARS Alert Analytical

Method		ARS-054		Analysis		LSC-LLH3/SC-AQ		Matrix		AQ		
Description		Low Level Tritium Screening										
ABatch Sample ID	Type	Blind Iso1	Blind Iso2	Blind Iso3	SDG	FR	Run	Prep Code	Filtered	Client ID	Group Name	Lab Deadline
ARS1-B18-00339-01	LCS											
ARS1-B18-00339-02	LCS											
ARS1-B18-00339-03	MBL											
ARS1-B18-00339-04	TRG				ARS1-18-00398	001	1			CAWA-18-72		03/21/18
ARS1-B18-00339-05	TRG				ARS1-18-00398	002	1			CAWA-18-75		03/21/18
ARS1-B18-00339-06	TRG				ARS1-18-00398	003	1			CAWA-18-78		03/21/18
ARS1-B18-00339-07	TRG				ARS1-18-00398	004	1			CAWA-18-129		03/21/18
ARS1-B18-00339-08	TRG				ARS1-18-00402	001	1			CAWA-18-34		03/21/18
ARS1-B18-00339-09	TRG				ARS1-18-00403	001	1			CrPZ-2-18-151284		03/21/18
ARS1-B18-00339-10	TRG				ARS1-18-00403	002	1			CrPZ-5-18-151305		03/21/18
ARS1-B18-00339-11	TRG				ARS1-18-00404	001	1			CAWA-18-81		03/21/18
ARS1-B18-00339-12	TRG				ARS1-18-00415	003	1			GNO01-01.1802001-003 (USGS-1)		03/10/18

Assay Definition

Assay Description:
LLH3 Assay in DPM Mode
Assay Type: DPM (Single)
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\ARS\Low Level H3_3\20180216_1735
Raw Results Path: C:\Packard\Tricarb\Results\ARS\Low Level H3_3\20180216_1735\20180216_1735.results
RTF File Name: C:\Packard\Tricarb\Results\ARS\Low Level H3_3\20180216_1735\LLH3.rtf
Comma-Delimited File Name: C:\Packard\Tricarb\Results\ARS\Low Level H3_3\20180216_1735\LLH3 Results.csv
Assay File Name: C:\Packard\TriCarb\Assays\Low Level H3_3.1sa

Count Conditions

Nuclide: Low Level H3
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set:
Low Energy: LLH3 10ml
Count Time (min): 120.00
Count Mode: Low Level
Assay Count Cycles: 1
Number of Vials/Sample: 1
Repeat Sample Count: 1
Calculate % Reference: Off

Background Subtract

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL
A	2.0	18.6
B	0.0	2000.0
C	0.0	2000.0

46 of 82

Count Corrections


Static Controller: On
Colored Samples: n/a
Coincidence Time (nsec): 18
Luminescence Correction: Off
Heterogeneity Monitor: Off
Delay Before Burst (nsec): 75
GCT: Off
PAC: Disabled
PAC Strength: n/a
Auxiliary Spectrum: n/a

Cycle 1 Results

P#	S#	SMPL ID	CPMA	DPM1	tsIE	Eff Nucl	In A	Count	Time	DATE	TIME	MESSAGES
5	1	BACKGROUND	1.118	5.33	257.78		20.98	120.00	120.00	2/16/2018	7:46:55 PM	
5	2	B18-00339-09	1.536	7.39	253.15		20.77	120.00	120.00	2/16/2018	9:58:11 PM	
5	3	B17-00339-10	1.144	5.47	256.70		20.93	120.00	120.00	2/17/2018	12:09:21 AM	

SNC Protocol


Calibration Information
Software Version IC: 5.2
Software Version EC: 5.2
Instrument Model: Tri-Carb 6220TR/SL
Instrument Serial Number: SGL025170524
3H Chi Square: 16.67 Date Processed: 2/16/2018 5:35:13 PM
14C Chi Square: 15.58 Date Processed: 2/16/2018 5:35:13 PM
3H FOM (1-18.6 keV): 1927.88 Date Processed: 2/16/2018 5:35:13 PM
14C FOM (4-156 keV): 7333.73 Date Processed: 2/16/2018 5:35:13 PM
3H Efficiency (1-18.6 keV): 60.79 Date Processed: 2/16/2018 5:35:13 PM
14C Efficiency (4-156 keV): 92.50 Date Processed: 2/16/2018 5:35:13 PM
IPA Background Date Processed: 2/16/2018 5:35:13 PM
3H Background CPM (1-18.6 keV): 1.92 Date Processed: 2/16/2018 5:35:13 PM
14C Background CPM (4-156 keV): 1.17 Date Processed: 2/16/2018 5:35:13 PM
3H Calibration DPM: 276400
3H Reference Date: 1/26/2017
14C Calibration DPM: 118500

\\PACKARDJ170_NEW\Results\ARS\Low Low Level Tritium



LSC Instrument Data Transfer Report

\\Tricarb\ars\Low Level H3_3\20180211

 INTERNATIONAL			Batch Sample ID			Non-BKG Samples Transferred			Samples Eligible To Save			LSC 5		
			ARS1-B18-00339			2			2					
LIMS Batch Sample ID	LSC P#	LSC PID	LSC S#	LSC SMPL_ID	LSC Count Date	LSC CPMA	LSC SEIE	LSC EFF	LSC Count Dur	Analysis Batch	LIMS SDG	LIMS Run		
BKG	5		1	BACKGROUND	02/16/18 19:46	1.12	257.78	20.9800	120.00	ARS1-B18-00339				
ARS1-B17-00339-10	5		3	B17-00339-10	02/17/18 00:09	1.14	256.70	20.9300	120.00	ARS1-B17-00339	ARS1-17-00499	1		
ARS1-B18-00339-09	5		2	B18-00339-09	02/16/18 21:58	1.54	253.15	20.7700	120.00	ARS1-B18-00339	ARS1-18-00403	1		

Assay Definition-

Assay Description:

Low Level H3

Assay Type: DPM (Single)

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\ARS\Low Low Level Tritium 3\20180215_0746

Raw Results Path: C:\Packard\Tricarb\Results\ARS\Low Low Level Tritium 3\20180215_0746\20180215_0746.results

RTF File Name: C:\Packard\Tricarb\Results\ARS\Low Low Level Tritium 3\20180215_0746\Report1.rtf

Comma-Delimited File Name: C:\Packard\Tricarb\Results\ARS\Low Low Level Tritium 3\20180215_0746\LLH3 Results.csv

Assay File Name: C:\Packard\Tricarb\Assays\Low Low Level Tritium 3.lsa

Count Conditions-

Nuclide: H-3 LL

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set:

Low Energy: ARS LL H3 10

Count Time (min): 120.00

Count Mode: Low Level

Assay Count Cycles: 1

#Vials/Sample: 1

Repeat Sample Count: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: On - Any Region

Regions	LL	UL	2Sigma % Terminator
A	2.0	18.6	0.50
B	0.0	2000.0	0.00
C	0.0	2000.0	0.00

Count Corrections-

Static Controller: On

Colored Samples: Off

Coincidence Time (nsec): 18

Half Life-

Luminescence Correction: Off

Heterogeneity Monitor: Off

Delay Before Burst (nsec): 200

Half Life Correction: Off

Regions Half Life

Units Reference Date Reference Time

Protocol# 49 - Low Low Level Tritium 3.lsa

User: ARS

A
B
C

Cycle 1 Results

P#	S#	SMPL ID	CPMA	DPM1	tSIE	Eff	Nucl	In A	Count	Time	DATE	TIME	MESSAGES
49	1	BACKGROUND	1.274	5.814	210.93			21.92	120.00		2/15/2018	7:54:56 AM	
49	2	B18-00339-04	0.982	4.519	208.34			21.72	120.00		2/15/2018	10:04:55 AM	
49	3	B18-00339-05	1.081	4.970	208.56			21.74	120.00		2/15/2018	12:14:56 PM	
49	4	B18-00339-06	0.968	4.335	216.48			22.32	120.00		2/15/2018	2:24:58 PM	
49	5	B18-00339-07	1.107	5.009	213.44			22.10	120.00		2/15/2018	4:34:59 PM	
49	6	B18-00339-08	1.189	5.445	209.78			21.83	120.00		2/15/2018	6:45:00 PM	
49	7	B18-00339-09	2.125	9.761	208.91			21.77	120.00		2/15/2018	8:55:01 PM	
49	8	B18-00339-10	2.751	12.448	213.46			22.10	120.00		2/15/2018	11:05:03 PM	
49	9	B18-00339-11	1.127	5.086	214.18			22.16	120.00		2/16/2018	1:15:18 AM	
49	10	B18-00339-12	1.062	4.861	209.99			21.85	120.00		2/16/2018	3:25:25 AM	

ARS-054
Tritium in Water

ARS International
Baton Rouge Laboratory

Preparation Date: 02/13/2018 13:56
Prepared By: MMORGAN

Procedure Data						
ABatch Sample ID	Client ID	Parent	ICOC ID	Aliquot 1 Vol/Wt	Aliquot 1 Units	Aliquot 2 Vol/Wt
ARS1-B18-00339-01						
ARS1-B18-00339-02						
ARS1-B18-00339-03						
ARS1-B18-00339-04	CAWA-18-72		286557	0.0010 L		
ARS1-B18-00339-05	CAWA-18-75		286558	0.0010 L		
ARS1-B18-00339-06	CAWA-18-78		286559	0.0010 L		
ARS1-B18-00339-07	CAWA-18-129		286560	0.0010 L		
ARS1-B18-00339-08	CAWA-18-34		286561	0.0010 L		
ARS1-B18-00339-09	CrPZ-2-18-151284		286562	0.0010 L		
ARS1-B18-00339-10	CrPZ-5-18-151305		286563	0.0010 L		
ARS1-B18-00339-11	CAWA-18-81		286564	0.0010 L		
ARS1-B18-00339-12	GNO01-01.1802001-003 (USGS-1)		286566	0.0010 L		

Reagent Amounts			
ABatch Sample ID	Client ID	14.1.5 OPTIONAL AQ W/O DIST - Add scint cocktail - Ultima Gold LLT Reagent Grade (mL)	User ID
ARS1-B18-00339-01		10.00	
ARS1-B18-00339-02		10.00	
ARS1-B18-00339-03		10.00	
ARS1-B18-00339-04	CAWA-18-72	10.00	
ARS1-B18-00339-05	CAWA-18-75	10.00	
ARS1-B18-00339-06	CAWA-18-78	10.00	
ARS1-B18-00339-07	CAWA-18-129	10.00	
ARS1-B18-00339-08	CAWA-18-34	10.00	
ARS1-B18-00339-09	CrPZ-2-18-151284	10.00	
ARS1-B18-00339-10	CrPZ-5-18-151305	10.00	
ARS1-B18-00339-11	CAWA-18-81	10.00	
ARS1-B18-00339-12	GN001-01.1802001-003 (USGS-1)	10.00	

Reagent Tracking									
Reagent ID	Lot	Quantity	Unit	Expiry Date	Storage Location	Revised By	Revised Date	Revised Time	Revised
R17-00665									

Procedure Section

14.1.5 OPTIONAL AQ W/O DIST - Add scint cocktail

Low Level Tritium pH Checks

SDG#	Fraction	pH	Date	Analyst
ARSI-18-00398	001	8	2-13-18	MM
	002	8	2-13-18	MM
	003	8	2-13-18	MM
	004	8	2-13-18	MM
ARSI-18-00402	001	8	2-13-18	MM
ARSI-18-00403	001	8	2-13-18	MM
	002	8	2-13-18	MM
ARSI-18-00404	001	8	2-13-18	MM
ARSI-18-00405	003	8	2-13-18	MM
MM				
2-13-18				

Liquid Scintillation Count Log

Date	Time	ARS Sample I.D. Number	Batch Fraction Number	Liquid Scintillation File Number	Technician Initials	Notes Identifier
2/2/2018	13:45	B17-02662	7	2006	MM	
2/2/2018	13:45	B17-02662	8	2006	MM	
2/2/2018	13:45	B17-02662	9	2006	MM	
2/2/2018	13:45	B17-02662	10	2006	MM	
2/2/2018	13:45	B17-02662	11	2006	MM	
2/2/2018	13:45	B17-02662	12	2006	MM	
2/2/2018	13:45	B17-02662	13	2006	MM	
2/2/2018	13:45	B17-02662	3	2006	MM	
2/5/2018	14:45	SNC163	QA	QA	MM	
2/5/2018	14:45	Background	N/A	N/A	MM	
2/5/2018	14:45	B17-02742	1	1828	MM	
2/5/2018	14:45	B17-02742	2	1828	MM	
2/5/2018	14:45	B17-02742	3	1828	MM	
2/5/2018	14:45	B17-02742	4	1828	MM	
2/5/2018	14:45	B17-02742	5	1828	MM	
2/5/2018	14:45	B17-02742	6	1828	MM	
2/5/2018	14:45	B17-02742	7	1828	MM	
2/5/2018	14:45	B17-02742	8	1828	MM	
2/5/2018	14:45	B17-02742	9	1828	MM	
2/5/2018	14:45	B17-02742	10	1828	MM	
2/8/2018	16:00	SNC163	QA	QA	MM	
2/8/2018	16:00	Background	N/A	N/A	MM	
2/8/2018	16:00	B17-02743	1	1056	MM	
2/8/2018	16:00	B17-02743	2	1056	MM	
2/8/2018	16:00	B17-02743	3	1056	MM	
2/8/2018	16:00	B17-02743	4	1056	MM	
2/8/2018	16:00	B17-02743	5	1056	MM	
2/8/2018	16:00	B17-02743	6	1056	MM	
2/12/2018	17:00	SNC163	QA	QA	MM	
2/12/2018	17:00	Background	N/A	N/A	MM	
2/12/2018	17:00	B17-02877	1	1105	MM	
2/12/2018	17:00	B17-02877	2	1105	MM	
2/12/2018	17:00	B17-02877	3	1105	MM	
2/12/2018	17:00	B17-02877	4	1105	MM	
2/12/2018	17:00	B17-02877	5	1105	MM	
2/12/2018	17:00	B17-02877	6	1105	MM	
2/12/2018	17:00	B17-02877	7	1105	MM	
2/14/2018	15:30	SNC163	QA	QA	MM	
2/14/2018	15:30	Background	N/A	1812	MM	
2/14/2018	15:30	B18-00121	1	1812	MM	
2/14/2018	15:30	B18-00121	2	1812	MM	
2/14/2018	15:30	B18-00121	3	1812	MM	
2/14/2018	15:30	B18-00121	4	1812	MM	
2/14/2018	15:30	B18-00121	5	1812	MM	
2/14/2018	15:30	B18-00121	6	1812	MM	
2/14/2018	15:30	B18-00121	7	1812	MM	
2/16/2018	12:30	SNC163	QA	QA	MM	
2/16/2018	12:30	Background	N/A	1735	MM	
2/16/2018	12:30	B18-00339	9	1735	MM	
2/16/2018	12:30	B18-00339	10	1735	MM	

Beta Liquid Scintillation Counter Log Book

Date	Time	ARS Sample I.D. Number	Batch Number	Liquid Scintillation File Number	Technician Initials
12-13-18	1247	B18-00315-06	B18-00315	1428 1527	
L	↓	↓ 07	L	L	
12-13-18	1547	SNC 5	QA	QA	
↓	↓	Background	B18-00308	1703	mm
↓	↓	B18-00308-04		↓	mm
↓	↓	↓ -05		↓	mm
↓	↓	↓ -06		↓	mm
↓	↓	↓ -07		↓	mm
↓	↓	↓ -08		↓	mm
12-13-18	1600	SNC 5	QA	QA	mm
↓	↓	B18-00339-04	B18-00339	0746	mm
↓	↓	↓ -05	↓	↓	mm
↓	↓	↓ -06	↓	↓	mm
↓	↓	↓ -07	↓	↓	mm
↓	↓	↓ -08	↓	↓	mm
↓	↓	↓ -09	↓	↓	mm
↓	↓	↓ -10	↓	↓	mm
↓	↓	↓ -11	↓	↓	mm
↓	↓	↓ -12	↓	↓	mm
↓	↓	Background	↓	↓	mm

*

Background placed ahead of B1800339-04



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory

**Tritium-Screening
by
Low Level Liquid
Scintillation Counting**


Control Charts

3H Efficiency

Total # pts : 88
Valid # pts : 31
Mean : 60.81
SD : 0.20

Date	Value	Include
Dec 20, 2017	60.46	X
Dec 22, 2017	60.94	X
Dec 27, 2017	61.04	X
Dec 31, 2017	61.03	X
Jan 03, 2018	61.20	X
Jan 04, 2018	60.96	X
Jan 06, 2018	61.23	X
Jan 08, 2018	60.86	X
Jan 10, 2018	60.87	X
Jan 11, 2018	61.00	X
Jan 14, 2018	60.93	X
Jan 15, 2018	60.63	X
Jan 15, 2018	60.92	X
Jan 16, 2018	60.62	X
Jan 18, 2018	60.74	X
Jan 20, 2018	60.85	X
Jan 22, 2018	61.03	X
Jan 26, 2018	60.76	X
Jan 26, 2018	60.91	X
Feb 01, 2018	60.89	X
Feb 02, 2018	60.87	X
Feb 02, 2018	60.59	X
Feb 05, 2018	60.60	X
Feb 06, 2018	60.84	X
Feb 09, 2018	60.75	X
Feb 11, 2018	60.50	X
Feb 12, 2018	60.56	X
Feb 13, 2018	60.56	X
Feb 14, 2018	60.57	X
Feb 14, 2018	60.75	X
Feb 16, 2018	60.79	X

3H Efficiency : 88
Total # pts : 31
Valid # pts : 60.81
Mean : 60.81
SD : 0.20




3H Background

Total # pts : 86
Valid # pts : 31
Mean : 1.85
SD : 0.19

Date	Value	Include
Dec 20, 2017	1.98	X
Dec 22, 2017	1.57	X
Dec 27, 2017	1.72	X
Dec 31, 2017	1.83	X
Jan 03, 2018	1.88	X
Jan 04, 2018	1.85	X
Jan 06, 2018	2.07	X
Jan 08, 2018	2.02	X
Jan 10, 2018	1.42	X
Jan 11, 2018	2.05	X
Jan 14, 2018	2.20	X
Jan 15, 2018	1.62	X
Jan 15, 2018	2.18	X
Jan 16, 2018	1.82	X
Jan 18, 2018	1.88	X
Jan 20, 2018	2.08	X
Jan 22, 2018	2.13	X
Jan 26, 2018	1.55	X
Jan 26, 2018	1.93	X
Feb 01, 2018	1.77	X
Feb 02, 2018	1.72	X
Feb 02, 2018	1.75	X
Feb 05, 2018	1.82	X
Feb 06, 2018	1.62	X
Feb 09, 2018	1.70	X
Feb 11, 2018	1.93	X
Feb 12, 2018	1.88	X
Feb 13, 2018	1.75	X
Feb 14, 2018	2.03	X
Feb 14, 2018	1.80	X
Feb 16, 2018	1.92	X

3H Background
Total # pts : 86
Valid # pts : 31
Mean : 1.85
SD : 0.19





2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

American Radiation Services Analytical Reports

for

Los Alamos National Laboratory

Low Level Liquid Scintillation Counting

Calibration Information



QUALITY CONTROL PROGRAM
AMERICAN RADIATION SERVICES
RADIOACTIVE REFERENCE SOLUTIONS
ANNUAL ACTIVITY VERIFICATION

VERIFICATION DATE **4/27/2017 0:00** date counted
 STANDARD REFERENCE # **[REDACTED]**

Principal Radionuclide

H-3

ENTER -->

Half Life, Years

1.232E+01
1.232E+01

OR -->

Half Life, Days

4.4999E+03
4.4999E+03
Radionuclide **H-3**Dilution Reference Date **4/27/2017 0:00**Dilution Activity **2.59** pCi per gram ==> dpm/g**5.75**Verif. Date Decay Corrected **2.59** pCi per gram ==> dpm/g**5.75****Minimum of 3 Required**

Trial ID	Sample Counts	Count Time (min)	Detector	Efficiency	Bkg. (cpm)	Net Weight	Decay Corrected Activity Result (dpm/g)	Decay Corrected Activity Result (pCi/g)
S-0324-V2	17.76	1	LSC	0.3007	9.00	5.027	5.79	2.61
S-0324-V3	18.32	1	LSC	0.3033	9.00	5.071	6.06	2.73
S-0324-V4	17.74	1	LSC	0.3041	9.00	5.019	5.73	2.58
S-0324-V5	18.62	1	LSC	0.3036	9.00	5.008	6.33	2.85

10% Max**PASS**

Standard Deviation percent of known concentration

5% Max**PASS**

Target Activity

% Diff

Average

Two Sigma Uncertainty

5.98**0.54****4.77%****5.75****3.93%****2.69****0.24****4.77%****2.59****3.93%**Verification Expiration Date: **April 27, 2018**Prepared & Counted By [Signature]Date: **4/27/2017 0:00**Verified & Approved By [Signature]Date: **4-28-17**QC Approval [Signature]Date: **04-28-17****S-0324****H-3**Verified **4/28/17****SL****Expires 4/28/18****Manufacturer****NIST SRM 4927F****Sol Matrix****H2O****Ref No****NIST SRM 4927F****Tech****Unknown****Parent ID****S-0316****RADIOACTIVE STANDARDS -- BATON ROUGE LABORATORY**



Add / Edit Secondary Standards

Planning		Parent Standard Data			
Planning Comments	Create H-3 LCS standard	Parent Solution Reference #	NIST SRM 4927F		
Target dpm/g (on dil. date)	6	Parent Solution #	S-0316		
Target Final Volume mL	2000.00	Parent Principal Radionuclide	H-3	Half Life (Days)	4499.8
Appx mass g of Parent Sol'n	5.23779991812081	Parent Reference Date	08/10/2016 14:49		
Appx vol ml of Parent Sol'n	5.24724495904709	Parent Certified Act	2384.43044412127	Cert Act/Vol Units	dpm g
Expected Addition for Analysis g	5	Parent Cert Act Uncert 1 Sigma	0.0036		
Standards Preparation / Dilution		Parent Sp. Gravity G/ML	0.9982		
Secondary Solution #	S-0324	Parent Supplier	NIST SRM 4927F		
Dilution Date (New Ref Date)	04/27/2017 0:00	Parent Date Recvd	01/01/00		
Ampoule, Empty (g)		Parent Received By	Unknown		
Ampoule/Solution Gross (g)		Parent Cert Exp Date			
Net Wt Removed (g)		Parent Matrix	H2O		
Transfer Container, empty (g)	17.2688	Certified dpm/g At Ref Date	607764.948573606		
Container Plus Solution(g)	22.2799	Certified dpm/g On 04/27/2017 0:00	583960.313234318		
Net Wt Transferred (g)	5.0111	Parent Comments	Intermediate level H-3 standard for creating LCS solutions and matrix spikes. Dilution performed as stated above by Jacob Byrd- JPB 08/10/2016		
DPM Xferred On 04/27/2017 0:00	11480.6218145069	Parent Tech	Unknown		
Diluent/matrix	DI Water	Is Primary	FALSE		
Diluent Density Cont, empty (g)	1E-05	Is LCS	TRUE		
Test Mass of 5 ml of Diluent (g)		Is Tracer	FALSE		
Diluent Density Test - (g/mL)		Is Calib	FALSE		
Dilution Empty Container Mass (g)	415.17				
Dilution Full Cont g (if measured)	2411.11				
Dilution Final Volume ml (if measured)	2000				
Final Dilution Density (g/mL)	0.99797				
Final Dilution Measured Mass g	1995.94				
Comments					
Final Dilution dpm/g	5.75198744176021				
Final Dil New Ref Date/Time	04/27/2017 0:00				

Protocol# 11 - H3 Normal Lvl 2.lsa

User: ARS

Assay Definition

Assay Description:
H3 Normal Lvl
Assay Type: DPM (Single)
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\ARS\H3 Normal Lvl 2\20170427_1250
Raw Results Path: C:\Packard\Tricarb\Results\ARS\H3 Normal Lvl 2\20170427_1250\20170427_1250.results
RTF File Name: C:\Packard\Tricarb\Results\ARS\H3 Normal Lvl 2\20170427_1250\H3 Results.rtf
Comma-Delimited File Name: C:\Packard\Tricarb\Results\ARS\H3 Normal Lvl 2\20170427_1250\H3 Results.csv
Assay File Name: C:\Packard\Tricarb\Assays\H3 Normal Lvl 2.lsa

Count Conditions

Nuclide: Standard H3
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s
Pre-Count Delay (min): 0.00
Quench Set:
Low Energy: PE UG STD H3
Count Time (min): 120.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: On - Any Region

Regions	LL	UL	2Sigma % Terminator
A	2.0	18.6	0.50
B	0.0	2000.0	0.00
C	0.0	2000.0	0.00

Count Corrections

Static Controller: On Luminescence Correction: n/a
Colored Samples: Off Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Cycle 1 Results											
P#	S#	SMPL_ID	CPMA	DPM1	tsIE	Eff Nucl	In A	Count Time	DATE	TIME	MESSAGES
11	1	BACKGROUND	9.00	29.71	381.35		30.28	120.00	4/27/2017	12:51:17 PM	
11	2	S-0234-V1	18.75	62.31	378.81		30.09	120.00	4/27/2017	3:03:56 PM	
11	3	S-0234-V2	17.76	59.04	378.52		30.07	120.00	4/27/2017	5:16:35 PM	
11	4	S-0234-V3	18.32	60.43	381.90		30.33	120.00	4/27/2017	7:29:15 PM	
11	5	S-0234-V4	17.74	58.34	382.99		30.41	120.00	4/27/2017	9:41:55 PM	
11	6	S-0234-V5	18.62	61.34	382.40		30.36	120.00	4/27/2017	11:54:37 PM	

0324

JB

04-28-17

0324 JB 04-28-17

S-0234 Verification Weights

Tech:	JPB
Pipette:	FJ15820
Scale ID:	12332539
Standard ID:	S-0234

Sample ID	Std. Weight(g)
S-0234-V1	5.0073
S-0234-V2	5.0271
S-0234-V3	5.0171
S-0234-V4	5.0189
S-0234-V5	5.0077

JB 0324
04-28-17



2609 North River Road • Port Allen, Louisiana 70767

1 (800) 401-4277 • Fax (225) 381-2996

**American Radiation Services
Analytical Reports**

for

Los Alamos National Laboratory

Folder Duplicate



Report Compilation Checklist

ARS SDG: 18-00398Client Name: LANLSample Matrix: AQ**LEVEL 1 COMPONENTS**

	1st Reviewer			
1) Cover Page Complete and Accurate (see ARS-059)?	Yes	No	N/A	
2) Technical Review Checklist(s) Complete and Accurate?	Yes	No	N/A	
3) Case Narrative Complete and Accurate (see ARS-059)? <small>Include subcontractor name and information</small>	Yes	No	N/A	
4) Form 1s Present for all Samples and Tests? <small>Note: Ensure original Subcontract Forms 1s included if applicable.</small>	Yes	No	N/A	
5) Client Specific Components are Present and Complete?	Yes	No	N/A	

LEVEL 2 COMPONENTS

	1st Reviewer			
6) Batch Quality Control Report is Present and Accurate? <small>Include subcontractor QC reports if applicable</small>	Yes	No	N/A	
7) DQO Report is Present and Accurate?	Yes	No	N/A	
8) Client Specific Batch QC Components are Present and Complete?	Yes	No	N/A	

LEVEL 3 COMPONENTSEnsure all original subcontractor information is included, if applicable

	1st Reviewer			
9) Efficiencies are Present?	Yes	No	N/A	
10) Calibrations are Present?	Yes	No	N/A	
11) Backgrounds are Present?	Yes	No	N/A	
12) Spectrum Analysis is Present?	Yes	No	N/A	
13) Spectral Plots are Present?	Yes	No	N/A	
14) Plateaus are Present?	Yes	No	N/A	
15) Control Charts are Present?	Yes	No	N/A	
16) Other:	Yes	No	N/A	

LEVEL 4 COMPONENTSEnsure all original subcontractor information is included, if applicable

	1st Reviewer			
17) Preparation Raw Data Present and Complete?	Yes	No	N/A	
18) Instrument Raw Data Present and Complete?	Yes	No	N/A	
19) Calibration Certificates Present?	Yes	No	N/A	
20) Copies of Log Book Pages Present?	Yes	No	N/A	
21) Sample Receiving Documentation Present?	Yes	No	N/A	
22) LIMS Reports Present?	Yes	No	N/A	
23) Applicable Correspondence Present?	Yes	No	N/A	
24) Other:	Yes	No	N/A	

Sch
Report Generator Signature

4-18-18
Date

[Signature]
Management Review Signature

4-23-18
Date



LSC Technical Review Checklist

ARS SDG ARS1-17-00398

Sample Matrix: AQ Aliquot (Circle One): Dry As Received ☒ Filtered Other: _____

Required QC Samples (Mark all that apply): _____ Blank _____ LCS _____ LCSD _____ Sample Dup _____ MS _____ MSD _____

ARS A. Batch ID(s): Batch A: B17-00339 Batch B: N/A Batch C: N/A

Test Method(s): LSC-LLH3/SC-AQ N/A N/A

A. RADIOCHEMICAL PREPARATION REVIEW

	Chemist Review			Verifier Review		
1) 100% of Manual Transcriptions Verified?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
2) 100% of Manual Calculations Verified?	<input type="radio"/> Yes	<input type="radio"/> No	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input checked="" type="radio"/> N/A
3) Blank Composition/Configuration Matches Calibration?	<input type="radio"/> Yes	<input type="radio"/> No	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input checked="" type="radio"/> N/A
4) Deviations from procedure are documented and verified?	<input type="radio"/> Yes	<input type="radio"/> No	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input checked="" type="radio"/> N/A
5) Appropriate Cocktail Selected?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
6) Sample Prep Anomaly? <input checked="" type="checkbox"/> No <input type="checkbox"/> Yes (See Tech Notes) NCR # (If initiated): _____						
Melisa Morgan 2-13-18 Chemist Signature Date			2-13-18 Verifier Review Signature Date			

B. ANALYSIS REVIEW

	Analyst Review			QA Officer Review		
1) Calibrations Valid and Current?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
2) Backgrounds Valid and Current?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
3) Source Checks Completed and Acceptable?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
			2-16-18 QA Officer Signature Date			
	Analyst Review			Technical Review		
4) Background Checks Complete and Acceptable?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
5) 100% of Manually Entered Parameters Verified Accurate?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
6) Appropriate QC samples initiated at required frequency?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
6) Test/Sample Specific Parameters (See ARS-059 for details)						
a) Analysis Parameters Checked and Correct and Peak Shapes are Acceptable?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
b) Spectra show no Evidence of Interferences?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
c) Sample Quench for All Samples within Range of Quench Curve?	<input checked="" type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A	<input type="radio"/> Yes	<input type="radio"/> No	<input type="radio"/> N/A
7) Analysis Anomaly? <input checked="" type="checkbox"/> No <input type="checkbox"/> Yes (See Comments) NCR # (If initiated): _____						
Melisa Morgan 2-16-18 Analyst Signature Date			N/A Technical Reviewer Signature Date			



C. BATCH QC VALIDATION

GENERAL COMMENTS

This image shows a single sheet of white paper with horizontal blue or grey ruling lines. The lines are evenly spaced and run across the width of the page. There is no handwriting or other markings on the paper.



LSC Technical Review Checklist

ARS SDG ARS1-18-00398Sample Matrix: AQ Aliquot (Circle One): Dry As Received ☒ Filtered Other: _____Required QC Samples (Mark all that apply): ☒ Blank ☒ LQS ☒ LQSD Sample Dup MS MSDARS A. Batch ID(s): Batch A: B18-00376 Batch B: N/A Batch C: N/ATest Method(s): LSC-LLH3-AQ N/A N/A

A. RADIOCHEMICAL PREPARATION REVIEW

	Chemist Review	Verifier Review
1) 100% of Manual Transcriptions Verified?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
2) 100% of Manual Calculations Verified?	Yes No <input checked="" type="radio"/> N/A	Yes No <input checked="" type="radio"/> N/A
3) Blank Composition/Configuration Matches Calibration?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
4) Deviations from procedure are documented and verified?	Yes No <input checked="" type="radio"/> N/A	Yes No <input checked="" type="radio"/> N/A
5) Appropriate Cocktail Selected?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
6) Sample Prep Anomaly? <input checked="" type="checkbox"/> No <input type="checkbox"/> Yes (See Tech Notes) NCR # (If initiated): _____		
<u>Melisa Morgan</u> 4-10-18 Chemist Signature Date	<u>[Signature]</u> 4-10-18 Verifier Review Signature Date	

B. ANALYSIS REVIEW

	Analyst Review	QA Officer Review
1) Calibrations Valid and Current?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
2) Backgrounds Valid and Current?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
3) Source Checks Completed and Acceptable?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
<u>[Signature]</u> 4-13-18 QA Officer Signature Date		
	Analyst Review	Technical Review
4) Background Checks Complete and Acceptable?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
5) 100% of Manually Entered Parameters Verified Accurate?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
6) Appropriate QC samples initiated at required frequency?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
6) Test/Sample Specific Parameters (See ARS-059 for details)		
a) Analysis Parameters Checked and Correct and Peak Shapes are Acceptable?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
b) Spectra show no Evidence of Interferences?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
c) Sample Quench for All Samples within Range of Quench Curve?	<input checked="" type="radio"/> Yes No N/A	<input checked="" type="radio"/> Yes No N/A
7) Analysis Anomaly? <input checked="" type="checkbox"/> No <input type="checkbox"/> Yes (See Comments) NCR # (If initiated): _____		
<u>Melisa Morgan</u> 4-13-2018 Analyst Signature Date	<u>[Signature]</u> 4-13-18 Technical Reviewer Signature Date	

DQO Report for SDG
ARS1-18-00398

Client Name: Los Alamos National Laboratory

Profile Name: Nita Patel (Site Alias)

Report Level: 4

Analysis Code	Prep Type	Units	Aliquot	Prep Code	Procedure	Count Time						
LSC-LLH3/SC-AQ	WRAD	pCi	L		ARS-054							
		Analyte		RDL	LCS LL/UL	MS LL/UL	Rdy LL/UL	Grav LL/UL	RER	RPD	Surr LL/UL	
H-3				150 pCi/L	75/125	60/140	30/120	40/110	1	25	N/A	
LSC-LLH3-AQ	WRAD	pCi	L	N/A	ARS-040							
		Analyte		RDL	LCS LL/UL	MS LL/UL	Rdy LL/UL	Grav LL/UL	RER	RPD	Surr LL/UL	
Enriched H-3				3.221 pCi/L	80/120	60/140	30/120	40/110	1	25	N/A	

DQO Report for SDG

ARS1-18-00398

Analysis Code	Fraction	Units	Aliquot	Conductivity	Analyte Count
LSC-LLH3/SC-AQ	001	pCi	L	N/A	1
LSC-LLH3/SC-AQ	002	pCi	L	N/A	1
LSC-LLH3/SC-AQ	003	pCi	L	N/A	1
LSC-LLH3/SC-AQ	004	pCi	L	N/A	1
LSC-LLH3-AQ	001	pCi	L	N/A	1
LSC-LLH3-AQ	002	pCi	L	N/A	1
LSC-LLH3-AQ	003	pCi	L	N/A	1
LSC-LLH3-AQ	004	pCi	L	N/A	1

SDG Report - Samples and Containers

SDG Specific Data														
SDG		ARS1-18-00398			TAT Days		40		Project Type		Environmental			
Sample Count		Rpt Level			4		Date Received		2/12/2018		COC Number		2018-1648	
Client		Los Alamos National Laboratory					Client Deadline		3/24/2018		PO Number			
Client Code		114					Internal Deadline		3/23/2018		Job Number			
Profile Number		PN-00094					Lab Deadline		3/21/2018		Job Location			
Comment														

Samples and Containers Checked In Thus Far													
FR	Name	Matrix	Start Date	End Date	Disp	pH Orig	pH Final	Hold	Arch	Storage	Conductivity	Comments	
001	CAWA-18-72	AQ	2/7/2018 11:37 AM	2/7/2018 11:37 AM	H			90	5	K6			
	IC_ID	Cnt	Volume (mL)	Container Type		pH Orig	pH Final	CPM	uR Hr		VOA	Head	Temp (C)
	286468	1	1000.00	HDP Container		7	7		80	18	N	N/A	0
002	CAWA-18-75	AQ	2/7/2018 11:32 AM	2/7/2018 11:32 AM	H			90	5	K6			
	IC_ID	Cnt	Volume (mL)	Container Type		pH Orig	pH Final	CPM	uR Hr		VOA	Head	Temp (C)
	286469	1	1000.00	HDP Container		7	7		80	20	N	N/A	0
003	CAWA-18-78	AQ	2/7/2018 1:16 PM	2/7/2018 1:16 PM	H			90	5	K6			
	IC_ID	Cnt	Volume (mL)	Container Type		pH Orig	pH Final	CPM	uR Hr		VOA	Head	Temp (C)
	286470	1	1000.00	HDP Container		7	7		90	20	N	N/A	0
004	CAWA-18-129	AQ	2/7/2018 1:16 PM	2/7/2018 1:16 PM	H			90	5	K6			
	IC_ID	Cnt	Volume (mL)	Container Type		pH Orig	pH Final	CPM	uR Hr		VOA	Head	Temp (C)
	286471	1	1000.00	HDP Container		7	7		80	22	N	N/A	0

SDG Report - Analysis Assignments

SDG	ARS1-18-00398	Sample Count	
Client	Los Alamos National Laboratory	Analysis Count	2-8

Sample Count Totals Per Analysis			
Analysis Code	Analysis Description	In/Out	Samples Count
LSC-LLH3/SC-AQ	Low Level Tritium Screen in (Aqueous)	I	4
LSC-LLH3-AQ	Low Level Tritium by Enrichment Process in (Aqueous [AQ])	I	4

Analyses Assigned Per Fraction		
Fraction	Analysis Code	X = Assigned
001	LSC-LLH3/SC-AQ	X
001	LSC-LLH3-AQ	X
002	LSC-LLH3/SC-AQ	X
002	LSC-LLH3-AQ	X
003	LSC-LLH3/SC-AQ	X
003	LSC-LLH3-AQ	X
004	LSC-LLH3/SC-AQ	X
004	LSC-LLH3-AQ	X

ARS FILE TRACKING SHEET

SDG: ARS1-18-00398

Task	Date / Time	Initials
Date & Time Samples Received	2/12/18 08:00	MC
ICOC Initiated/Storage Location: <u>K6</u>	2/12/18 08:51	MC
Technical Checks Performed		
Report Written / EDD Generated <u>4-13-18 / 14081</u> <u>SDR</u>	<u>4-13-18 / 1440</u>	<u>SDR</u>
Report / EDD Reviewed for accuracy and completeness	<u>4-13-18 15:15</u>	<u>EPD/Londel</u>
Quality Assurance Checks Performed on Report	<u>4-23-18</u>	
Management Checks Performed on Report	<u>9:40</u>	<u>JK</u>
Preliminary Report Scan	<u>Na</u>	
Report E-mailed/Faxed		
Invoice Completed Invoice #: _____		
Requires Report Mailed Yes / No		
Requires Original COC mailed Yes / No		
Report Reviewed and Imaged		

SPECIAL REQUIREMENTS

Requirement	Yes	No
3 Hour Rush	<input type="checkbox"/>	<input checked="" type="checkbox"/>
24 Hour Rush	<input type="checkbox"/>	<input checked="" type="checkbox"/>
48 Hour Rush	<input type="checkbox"/>	<input checked="" type="checkbox"/>
3 Day Rush	<input type="checkbox"/>	<input checked="" type="checkbox"/>
5 Day Rush	<input type="checkbox"/>	<input checked="" type="checkbox"/>
10 Day Rush	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Standard Oil/Gas Client (5 Day)	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Standard Turnaround	<input checked="" type="checkbox"/>	<input type="checkbox"/>

NOTES

SDG: AP51-18-00398

SHIPPING CONTAINER

Good Condition	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No
Radioactive	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No
UN2810	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No
Sec. Seals	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No
Seals intact	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	<input type="checkbox"/> N/A
Air Bill	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No

COC PRESENT WITH SAMPLES

COC ☒ Yes ☐ No

SAMPLE CONTAINER(S)

Good Condition ☒ Yes ☐ No
 Sec. Seal ☒ Yes ☐ No
 Seal intact ☒ Yes ☐ No ☐ N/A
 Radioactive ☐ Yes ☒ No

Marked Radioactive

Samples Rev

Matrix

AF, **AQ**, BI, FE, LT, SI, SO, UF, VG]

External and Internal Surveys

Exposure Rate Meter: M3 269264

Serial No.: PR 256427

Calibration Due Date: **3/13/18**

Count Rate Meter: M2 154859

Serial No.: **PR 121649**

Calibration Due Date: **3/9/18**

Background Exposure Rate
($\mu\text{R/hr}$)

Max. Exposure Rate on
Shipping Containers Externals
(Plus Bkgd) 20 $\mu\text{R/hr}$

Background Count Rate (cpm)

Max. Removable Count Rate on
Shipping Containers Externals
(Plus Bkgd) **80** cpm

Max. Removable Count Rate on Shipping Containers Internals (Plus Bkqd) 80 cpm

pH \leq 2 is Acceptable

Acceptance Limits

$$<500 \mu R/hr \quad <100 \text{ cpm/cm}^2$$
[illegible]

Surveyors' Name:

Date/Time Surveyed:

2-12-18

2825

[illegible]