

The order of this data package is as follows:

1. Chain-of-Custody/Lab Request
2. Copies of field COCs
3. Validation Report
4. Laboratory analysis

Comments:

SAMPLE COLLECTION LOG/FIELD CHAIN OF CUSTODY

EVENT ID: 11366

EVENT NAME: Mortandad/Sandia (Cr Inv) MY2017 Q4

SAMPLE ID: CAMO-17-141977

WORK ORDER:

	AS PLANNED	AS COLLECTED		AS PLANNED	AS COLLECTED
Date Collected (MM/DD/YYYY):	08/01/2017	OK	FIELD MATRIX:	WG	OK
TIME COLLECTED (HH:MM):	1233		MEDIA:	UA	
PRS ID:	NA		SAMPLE TECH CODE:	GSP	
LOCATION ID:	R-33 S1		FIELD PREP:	F	
LOCATION TYPE:	NA		FIELD QC TYPE:	REG	
TOP DEPTH:	↓	↓	SAMPLE USAGE:	INV	↓
BOTTOM DEPTH:	↓	↓	EXCAVATED:		YES / NO / (NA)

PRIORITY	ORDER	CONTAINER	#	PRESERVATIVE	COLLECTED Y/N	SPECIAL INSTRUCTIONS
NA	WSP-All Metals	1 LITER POLY	1	HNO3 ICE	Y	NA
↓	WSP- GENINORG+PerChlorate	1 LITER POLY	1	ICE	↓	↓
↓	WSP- NH3+NO3/NO2	500 ML AMBER GLASS	1	H2SO4	↓	↓

SAMPLE COMMENTS:

LOCATION COMMENTS:

FIELD PARAMETERS:

TV 8-1-17

Sample Time	_____	HH:MM	Dissolved Oxygen	_____	Flow (in gpm)	_____
Oxidation-Reduction Potential	_____		pH	_____	Specific Conductance	_____
Temperature	_____		Turbidity	_____		

COLLECTED BY (PRINT): A. Vigil, K. Tow

RELINQUISHED BY (Printed Name) Andrew Vigil (Signature) [Signature]	Date/Time 8-1-2017 1335	RECEIVED BY (Printed Name) K. Green (Signature) [Signature]	Date/Time 8/1/17 1:35
RELINQUISHED BY (Printed Name) (Signature)	Date/Time	RECEIVED BY (Printed Name) (Signature)	Date/Time

Report Date: 07/24/2017

SAMPLE COLLECTION LOG/FIELD CHAIN OF CUSTODY

EVENT ID: 11366

EVENT NAME: Mortandad/Sandia (Cr Inv) MY2017 Q4

SAMPLE ID: CAMO-17-141978

WORK ORDER:

	<u>AS PLANNED</u>	<u>AS COLLECTED</u>		<u>AS PLANNED</u>	<u>AS COLLECTED</u>
Date Collected (MM/DD/YYYY):	08/01/2017	OK	FIELD MATRIX:	WG	OK
TIME COLLECTED (HH:MM):	1034		MEDIA:	UA	
PRS ID:	NA		SAMPLE TECH CODE:	GSP	
LOCATION ID:	R-33 S2		FIELD PREP:	F	
LOCATION TYPE:	NA		FIELD QC TYPE:	REG	
TOP DEPTH:			SAMPLE USAGE:	INV	
BOTTOM DEPTH:			EXCAVATED:		YES / NO / NA

PRIORITY	ORDER	CONTAINER	#	PRESERVATIVE	COLLECTED Y/N	SPECIAL INSTRUCTIONS
NA	WSP-All Metals	1 LITER POLY	1	HNO3 ICE	Y	NA
	WSP- GENINORG+PerChlorat e	1 LITER POLY	1	ICE		
	WSP- NH3+NO3/NO2	500 ML AMBER GLASS	1	H2SO4		

SAMPLE COMMENTS:

LOCATION COMMENTS:

FIELD PARAMETERS:

Sample Time _____ HH:MM _____ TV 8-1-2017
 Dissolved Oxygen _____ Flow (in gpm) _____
 Oxidation-Reduction Potential _____ pH _____ Specific Conductance _____
 Temperature _____ Turbidity _____

COLLECTED BY (PRINT): K. Tow, A. Vigil

RELINQUISHED BY (Printed Name) ANDREW VIGIL (Signature) <i>Andrew Vigil</i>	Date/Time 8-1-2017 1335	RECEIVED BY (Printed Name) K. Tow (Signature) <i>K. Tow</i>	Date/Time 8/1/17 1:35
RELINQUISHED BY (Printed Name) (Signature)	Date/Time	RECEIVED BY (Printed Name) (Signature)	Date/Time

Report Date: 07/24/2017

SAMPLE COLLECTION LOG/FIELD CHAIN OF CUSTODY

EVENT ID: 11366

EVENT NAME: Mortandad/Sandia (Cr Inv) MY2017 Q4

SAMPLE ID: CAMO-17-141993

WORK ORDER:

	AS PLANNED	AS COLLECTED		AS PLANNED	AS COLLECTED
Date Collected (MM/DD/YYYY):	08/01/2017	OK	FIELD MATRIX:	WG	OK
TIME COLLECTED (HH:MM):	1233		MEDIA:	UA	
PRS ID:	NA		SAMPLE TECH CODE:	GSP	
LOCATION ID:	R-33 S1		FIELD PREP:	UF	
LOCATION TYPE:	NA		FIELD QC TYPE:	REG	
TOP DEPTH:			SAMPLE USAGE:	INV	
BOTTOM DEPTH:			EXCAVATED:		YES / NO / <u>NA</u>

PRIORITY	ORDER	CONTAINER	#	PRESERVATIVE	COLLECTED Y/N	SPECIAL INSTRUCTIONS
NA	MSGP-Hg	1000 500 ML POLY AS 8/1/17	1	HNO3	Y	NA
	WSP-CN(T)	250 ML POLY	1	NAOH		
	WSP-TKN+TOC	500 ML AMBER GLASS	1	H2SO4		

SAMPLE COMMENTS: Sampled with running diesel generator ~30 ft away.

LOCATION COMMENTS:

none

FIELD PARAMETERS:

Sample Time	1233	HH:MM	Dissolved Oxygen	5.31	Flow (in gpm)	3.30
Oxidation-Reduction Potential	118.3		pH	7.51	Specific Conductance	144.3
Temperature	22.6		Turbidity	0.23		

COLLECTED BY (PRINT): A. Vigil, K. Tow

RELINQUISHED BY (Printed Name) <u>ANDREW VIGIL</u> (Signature) <u>[Signature]</u>	Date/Time 8-1-2017 1335	RECEIVED BY (Printed Name) <u>K. Tow</u> (Signature) <u>[Signature]</u>	Date/Time 8/1/17 1:35
RELINQUISHED BY (Printed Name) (Signature)	Date/Time	RECEIVED BY (Printed Name) (Signature)	Date/Time

Report Date: 07/24/2017

SAMPLE COLLECTION LOG/FIELD CHAIN OF CUSTODY

EVENT ID: 11366

EVENT NAME: Mortandad/Sandia (Cr Inv) MY2017 Q4

SAMPLE ID: CAMO-17-141994

WORK ORDER:

	AS PLANNED	AS COLLECTED		AS PLANNED	AS COLLECTED
Date Collected (MM/DD/YYYY):	08/01/2017	OK	FIELD MATRIX:	WG	OK
TIME COLLECTED (HH:MM):	1034		MEDIA:	UA	
PRS ID:	NA		SAMPLE TECH CODE:	GSP	
LOCATION ID:	R-33 S2		FIELD PREP:	UF	
LOCATION TYPE:	NA		FIELD QC TYPE:	REG	
TOP DEPTH:	↓		SAMPLE USAGE:	INV	↓
BOTTOM DEPTH:	↓	↓	EXCAVATED:		YES / NO / <u>NA</u>

PRIORITY	ORDER	CONTAINER	#	PRESERVATIVE	COLLECTED Y/N	SPECIAL INSTRUCTIONS
NA	MSGP-Hg	1000 500 ML POLY AS 8/1/17	1	HNO3	Y	NA
↓	WSP-CN(T)	250 ML POLY	1	NAOH	↓	↓
↓	WSP-TKN+TOC	500 ML AMBER GLASS	1	H2SO4	↓	↓

SAMPLE COMMENTS: Sampled with running diesel generator ~ 30 ft away.

LOCATION COMMENTS: none

FIELD PARAMETERS:

Sample Time	1034	HH:MM	Dissolved Oxygen	6.48	Flow (in gpm)	2.80
Oxidation-Reduction Potential	132.8		pH	7.53	Specific Conductance	139.9
Temperature	21.5		Turbidity	0.29		

COLLECTED BY (PRINT): K. Tow, A. Vigil

RELINQUISHED BY (Printed Name) ANDREW VIGIL (Signature) Andrew Vigil	Date/Time 8-1-2017 1335	RECEIVED BY (Printed Name) K. Green (Signature) [Signature]	Date/Time 8/1/17 1:35
RELINQUISHED BY (Printed Name) (Signature)	Date/Time	RECEIVED BY (Printed Name) (Signature)	Date/Time

Report Date: 07/24/2017

DATA VALIDATION REPORT

Chain Of Custody No. 2017-2284

1. Distribution Of Samples In EDD.

SDG	Analytical Method	Regular Samples	Field Duplicates	Trip Blanks	Field Blanks	Equipment Blanks
429570	EPA:120.1	2				
429570	EPA:150.1	2				
429570	EPA:160.1	2				
429570	EPA:170.0	4				
429570	EPA:245.2	4				
429570	EPA:300.0	2				
429570	EPA:310.1	2				
429570	EPA:335.4	2				
429570	EPA:350.1	2				
429570	EPA:351.2	2				
429570	EPA:353.2	2				
429570	EPA:365.4	2				
429570	SM:A2340B	2				
429570	SW-846:6010C	2				
429570	SW-846:6020	2				
429570	SW-846:6850	2				
429570	SW-846:9060	2				

SDG	Analytical Method	Analysis Lot ID	Prep Lot ID	Regular Samples	Field Duplicates	Trip Blanks	Field Blanks	Equipment Blanks	Method Blanks	Matrix Spikes	Matrix Spike Dups	Analytical Spikes	Post-Digestion Spikes	Lab Control Samples	Lab Control Sample Dups	Blank Spike	Blank Spike Dups	Lab Duplicates	Storage Blanks	Preparation Blanks	Reagent Blanks
429570	EPA:120.1	1688310	1688310	2										1			2				
429570	EPA:150.1	1690702	1690702	2										1			1				
429570	EPA:160.1	1688433	1688433	2					1					1			1				
429570	EPA:170.0	NA	NA	4																	
429570	EPA:245.2	1694628	1694626	4					1	1				1			1				
429570	EPA:300.0	1691711	1691711	2					1					1			1				
429570	EPA:310.1	1690697	1690697	2						1				1			1				

DATA VALIDATION REPORT

SDG	Analytical Method	Analysis Lot ID	Prep Lot ID	Regular Samples	Field Duplicates	Trip Blanks	Field Blanks	Equipment Blanks	Method Blanks	Matrix Spikes	Matrix Spike Dups	Analytical Spikes	Post-Digestion Spikes	Lab Control Samples	Lab Control Sample Dups	Blank Spike	Blank Spike Dups	Lab Duplicates	Storage Blanks	Preparation Blanks	Reagent Blanks
429570	EPA:335.4	1688106	1688105	2					1	1				1				1			
429570	EPA:350.1	1692774	1692772	2					1	2				1				2			
429570	EPA:351.2	1692759	1692758	2					1	1				1				1			
429570	EPA:353.2	1689327	1689327	2					1					1				2			
429570	EPA:365.4	1692781	1692780	2					1	2				1				2			
429570	SM:A2340B	1695942	1695942	2																	
429570	SW-846:6010C	1688319	1688317	2					1	1				1				1			
429570	SW-846:6020	1688337	1688336	2					1	1				1				1			
429570	SW-846:6850	1689374	1689372	2					1	1	1			1							
429570	SW-846:9060	1688478	1688478	2					1					1				2			

2. Distribution Of Analytes In EDD.

Analytical Method	Analytical Method Category	Field Sample ID	Lab Sample ID	Sample Purpose	Target Analytes	Surrogates	Spiked Compounds	TICS
EPA:120.1	GENERAL CHEMISTRY	CAMO-17-141977	1203849533	DUP	1	0	0	0
EPA:120.1	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	1	0	0	0
EPA:120.1	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:120.1	GENERAL CHEMISTRY	CASA-17-142030	1203845788	DUP	1	0	0	0
EPA:120.1	GENERAL CHEMISTRY	LCS	1203845787	LCS	0	0	1	0
EPA:150.1	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	1	0	0	0
EPA:150.1	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:150.1	GENERAL CHEMISTRY	CASA-17-142006	1203851538	DUP	1	0	0	0
EPA:150.1	GENERAL CHEMISTRY	LCS	1203851537	LCS	0	0	1	0
EPA:160.1	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	1	0	0	0
EPA:160.1	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:160.1	GENERAL CHEMISTRY	CAMO-17-142780	1203846057	DUP	1	0	0	0
EPA:160.1	GENERAL CHEMISTRY	LCS	1203846056	LCS	0	0	1	0
EPA:160.1	GENERAL CHEMISTRY	MB	1203846055	MB	1	0	0	0
EPA:170.0	VOC	CAMO-17-141977	429570001	REG	1	0	0	0
EPA:170.0	VOC	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:170.0	VOC	CAMO-17-141993	429570002	REG	1	0	0	0
EPA:170.0	VOC	CAMO-17-141994	429570004	REG	1	0	0	0

DATA VALIDATION REPORT

Analytical Method	Analytical Method Category	Field Sample ID	Lab Sample ID	Sample Purpose	Target Analytes	Surrogates	Spiked Compounds	TICS
EPA:245.2	INORGANIC	CAMO-17-141977	429570001	REG	1	0	0	0
EPA:245.2	INORGANIC	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:245.2	INORGANIC	CAMO-17-141993	429570002	REG	1	0	0	0
EPA:245.2	INORGANIC	CAMO-17-141994	429570004	REG	1	0	0	0
EPA:245.2	INORGANIC	CASA-17-142006	1203860688	DUP	1	0	0	0
EPA:245.2	INORGANIC	CASA-17-142006	1203860689	MS	0	0	1	0
EPA:245.2	INORGANIC	LCS	1203860687	LCS	0	0	1	0
EPA:245.2	INORGANIC	MB	1203860686	MB	1	0	0	0
EPA:300.0	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	4	0	0	0
EPA:300.0	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	4	0	0	0
EPA:300.0	GENERAL CHEMISTRY	CASA-17-142013	1203854053	DUP	4	0	0	0
EPA:300.0	GENERAL CHEMISTRY	LCS	1203854052	LCS	0	0	4	0
EPA:300.0	GENERAL CHEMISTRY	MB	1203854051	MB	4	0	0	0
EPA:310.1	GENERAL CHEMISTRY	CAMO-17-141977	1203851524	DUP	2	0	0	0
EPA:310.1	GENERAL CHEMISTRY	CAMO-17-141977	1203851528	MS	0	0	1	0
EPA:310.1	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	2	0	0	0
EPA:310.1	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	2	0	0	0
EPA:310.1	GENERAL CHEMISTRY	LCS	1203851522	LCS	0	0	1	0
EPA:335.4	GENERAL CHEMISTRY	CAMO-17-141993	429570002	REG	1	0	0	0
EPA:335.4	GENERAL CHEMISTRY	CAMO-17-141994	429570004	REG	1	0	0	0
EPA:335.4	GENERAL CHEMISTRY	CASA-17-142016	1203845292	DUP	1	0	0	0
EPA:335.4	GENERAL CHEMISTRY	CASA-17-142016	1203845295	MS	0	0	1	0
EPA:335.4	GENERAL CHEMISTRY	LCS	1203845290	LCS	0	0	1	0
EPA:335.4	GENERAL CHEMISTRY	MB	1203845289	MB	1	0	0	0
EPA:350.1	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	1	0	0	0
EPA:350.1	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:350.1	GENERAL CHEMISTRY	CAMO-17-142070	1203856523	DUP	1	0	0	0
EPA:350.1	GENERAL CHEMISTRY	CAMO-17-142070	1203856525	MS	0	0	1	0
EPA:350.1	GENERAL CHEMISTRY	LCS	1203856521	LCS	0	0	1	0
EPA:350.1	GENERAL CHEMISTRY	MB	1203856520	MB	1	0	0	0
EPA:350.1	GENERAL CHEMISTRY	WT_SEP-PO-17-141444	1203856522	DUP	1	0	0	0
EPA:350.1	GENERAL CHEMISTRY	WT_SEP-PO-17-141444	1203856524	MS	0	0	1	0
EPA:351.2	GENERAL CHEMISTRY	CAMO-17-141993	429570002	REG	1	0	0	0
EPA:351.2	GENERAL CHEMISTRY	CAMO-17-141994	429570004	REG	1	0	0	0
EPA:351.2	GENERAL CHEMISTRY	CASA-17-142037	1203856480	DUP	1	0	0	0
EPA:351.2	GENERAL CHEMISTRY	CASA-17-142037	1203856481	MS	0	0	1	0
EPA:351.2	GENERAL CHEMISTRY	LCS	1203856479	LCS	0	0	1	0
EPA:351.2	GENERAL CHEMISTRY	MB	1203856478	MB	1	0	0	0
EPA:353.2	GENERAL CHEMISTRY	CAMO-17-141977	1203848237	DUP	1	0	0	0
EPA:353.2	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	1	0	0	0

DATA VALIDATION REPORT

Analytical Method	Analytical Method Category	Field Sample ID	Lab Sample ID	Sample Purpose	Target Analytes	Surrogates	Spiked Compounds	TICS
EPA:353.2	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:353.2	GENERAL CHEMISTRY	CAMO-17-142236	1203848236	DUP	1	0	0	0
EPA:353.2	GENERAL CHEMISTRY	LCS	1203848235	LCS	0	0	1	0
EPA:353.2	GENERAL CHEMISTRY	MB	1203848234	MB	1	0	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-17-141977	429570001	REG	1	0	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-17-141978	429570003	REG	1	0	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-17-141979	1203856559	DUP	1	0	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-17-141979	1203856561	MS	0	0	1	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-17-141985	1203856560	DUP	1	0	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-17-141985	1203856562	MS	0	0	1	0
EPA:365.4	GENERAL CHEMISTRY	LCS	1203856558	LCS	0	0	1	0
EPA:365.4	GENERAL CHEMISTRY	MB	1203856557	MB	1	0	0	0
SM:A2340B	INORGANIC	CAMO-17-141977	429570001	REG	1	0	0	0
SM:A2340B	INORGANIC	CAMO-17-141978	429570003	REG	1	0	0	0
SW-846:6010C	INORGANIC	CAMO-17-141977	1203845802	DUP	17	0	0	0
SW-846:6010C	INORGANIC	CAMO-17-141977	1203845803	MS	0	0	17	0
SW-846:6010C	INORGANIC	CAMO-17-141977	429570001	REG	17	0	0	0
SW-846:6010C	INORGANIC	CAMO-17-141978	429570003	REG	17	0	0	0
SW-846:6010C	INORGANIC	LCS	1203845801	LCS	0	0	17	0
SW-846:6010C	INORGANIC	MB	1203845800	MB	17	0	0	0
SW-846:6020	INORGANIC	CAMO-17-141977	1203845845	DUP	11	0	0	0
SW-846:6020	INORGANIC	CAMO-17-141977	1203845846	MS	0	0	11	0
SW-846:6020	INORGANIC	CAMO-17-141977	429570001	REG	11	0	0	0
SW-846:6020	INORGANIC	CAMO-17-141978	429570003	REG	11	0	0	0
SW-846:6020	INORGANIC	LCS	1203845844	LCS	0	0	11	0
SW-846:6020	INORGANIC	MB	1203845843	MB	11	0	0	0
SW-846:6850	LCMS/MS PERCHLORATE	CAMO-17-141977	1203848365	MS	0	0	1	0
SW-846:6850	LCMS/MS PERCHLORATE	CAMO-17-141977	1203848366	MSD	0	0	1	0
SW-846:6850	LCMS/MS PERCHLORATE	CAMO-17-141977	429570001	REG	1	0	0	0
SW-846:6850	LCMS/MS PERCHLORATE	CAMO-17-141978	429570003	REG	1	0	0	0
SW-846:6850	LCMS/MS PERCHLORATE	LCS	1203848364	LCS	0	0	1	0
SW-846:6850	LCMS/MS PERCHLORATE	MB	1203848363	MB	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	CAMO-17-141993	429570002	REG	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	CAMO-17-141994	429570004	REG	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	CAMO-17-142312	1203846822	DUP	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	LCS	1203846821	LCS	0	0	1	0
SW-846:9060	GENERAL CHEMISTRY	MB	1203846820	MB	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	WT_LAP-17-133623	1203846823	DUP	1	0	0	0

DATA VALIDATION REPORT

3. Are any analytes missing?

No.

4. Were any holding times exceeded?

No.

5. Any contaminants in blanks?

Blank FS ID	Blank Lab Sample	Blank Type	Analytical Method	Sample	Parameter Name	Blank Lab Result	Lab Qualifier	Blank Lab Units	Blank Lab Detection Limit
MB	1203845800	METHOD BLANK	SW-846:6010C	W	Sodium	-216	J	ug/L	300
MB	1203854051	METHOD BLANK	EPA:300.0	W	Sulfate	0.175	J	mg/L	0.400

Field Sample ID	Blank Lab	Blank Type	Analytical Method	Parameter Name	Blank Lab Result	Blank Lab Units	Lab Result	Lab Qualifier	Lab Detection Limit	Detect Flag	Detect to Nondetect Factor	Detect to Estimated Factor	Use Factors
CAMO-17-141977	1203854051	METHOD BLANK	EPA:300.0	Sulfate	0.175	mg/L	3.44		0.400	Y	5	100	Y
CAMO-17-141978	1203854051	METHOD BLANK	EPA:300.0	Sulfate	0.175	mg/L	2.32		0.400	Y	5	100	Y
CAMO-17-141977	1203845800	METHOD BLANK	SW-846:6010C	Sodium	-216	ug/L	11700		300	Y			
CAMO-17-141978	1203845800	METHOD BLANK	SW-846:6010C	Sodium	-216	ug/L	11000		300	Y			

6. Any surrogate recoveries outside the control limits?

No.

DATA VALIDATION REPORT

7. Any MS/MSD recoveries or RPDs outside the control limits?

Field Sample ID	MS Lab Sample ID	MSD Lab Sample ID	Analytical Method	Parameter Name	Analysis Lot ID	Analysis Date	Sample Matrix	MS Spike Recovery	MSD Spike Recovery	MS Upper Limit	MS Lower Limit	MS Reject Limit	RPD	RPD Limit
CASA-17-142037	1203856481		EPA:351.2	Total Kjeldahl Nitrogen	1692758	08-24-2017	W	63.2		110	90	10		

8. Any LCS/LCSD or BS/BSD recoveries or RPDs outside the control limits?

No.

9. Any Field Duplicate RPDs outside the desired limits?

No.

10. Any Lab Duplicate RPDs outside the desired limits?

No.

11. Any required reporting limits exceeded?

No.

12. Additional Validator's Comments.

13. Display Flagged Data.

Location ID	COC Number	Field Sample ID	Sample Purpose	Analysis Type Code	Analytical Suite	Analytical Method	Parameter Name	Lab Qualifier	Validation Qualifier	Validation Reason Codes	Detect Flag	Lab Result	Lab Units	Report Result	Report Units	Report MDA	Report Uncertainty	Lab Matrix	Sample Date	Percent	Analysis Lot ID	Validation Status Code	Use Flag
R-33 S1	2017-2284	CAMO-17-141977	REG	INIT	GENERAL CHEMISTRY	EPA:300.0	Sulfate		J+	I4a	Y	3.44	mg/L	3.44	mg/L			W	08/01/2017		1691711	VAL	Y

DATA VALIDATION REPORT

Location ID	COC Number	Field Sample ID	Sample Purpose	Analysis Type Code	Analytical Suite	Analytical Method	Paramter Name	Lab Qualifier	Validation Qualifier	Validation Reason Codes	Detect Flag	Lab Result	Lab Units	Report Result	Report Units	Report MDA	Report Uncertainty	Lab Matrix	Sample Date	Percent	Analysis Lot ID	Validation Status Code	Use Flag
R-33 S2	2017-2284	CAMO-17-141978	REG	INIT	GENERAL CHEMISTRY	EPA:300.0	Sulfate	J+	I4a	Y	2.32	mg/L	2.32	mg/L			W	08/01/2017	1691711	VAL	Y		

Reason Code

Description

I4a	The affected analytes are considered estimated and biased high because this analyte was identified in the method blank but was >5x
J_LAB	The analytical laboratory qualified the detected result as estimated (J) because the result was less the PQL but greater than the MDL
NQ	The analytical laboratory did not qualify the analyte as not detected and/or any other standard qualify. The analyte is detected in the sample.
U_LAB	The analytical laboratory qualified the analyte as not detected.

14. Usable Result Count.

Field Sample ID	Location ID	Sample Purpose	Analytical Method	No. Unuseable Records	Total Records
CAMO-17-141977	R-33 S1	REG	EPA:120.1	0	1
CAMO-17-141977	R-33 S1	REG	EPA:150.1	0	1
CAMO-17-141977	R-33 S1	REG	EPA:160.1	0	1
CAMO-17-141977	R-33 S1	REG	EPA:170.0	0	1
CAMO-17-141977	R-33 S1	REG	EPA:245.2	0	1
CAMO-17-141977	R-33 S1	REG	EPA:300.0	0	4
CAMO-17-141977	R-33 S1	REG	EPA:310.1	0	2
CAMO-17-141977	R-33 S1	REG	EPA:350.1	0	1
CAMO-17-141977	R-33 S1	REG	EPA:353.2	0	1
CAMO-17-141977	R-33 S1	REG	EPA:365.4	0	1
CAMO-17-141977	R-33 S1	REG	SM:A2340B	0	1
CAMO-17-141977	R-33 S1	REG	SW-846:6010C	0	17
CAMO-17-141977	R-33 S1	REG	SW-846:6020	0	11
CAMO-17-141977	R-33 S1	REG	SW-846:6850	0	1
CAMO-17-141978	R-33 S2	REG	EPA:120.1	0	1
CAMO-17-141978	R-33 S2	REG	EPA:150.1	0	1
CAMO-17-141978	R-33 S2	REG	EPA:160.1	0	1
CAMO-17-141978	R-33 S2	REG	EPA:170.0	0	1
CAMO-17-141978	R-33 S2	REG	EPA:245.2	0	1
CAMO-17-141978	R-33 S2	REG	EPA:300.0	0	4

DATA VALIDATION REPORT

Field Sample ID	Location ID	Sample Purpose	Analytical Method	No. Unuseable Records	Total Records
CAMO-17-141978	R-33 S2	REG	EPA:310.1	0	2
CAMO-17-141978	R-33 S2	REG	EPA:350.1	0	1
CAMO-17-141978	R-33 S2	REG	EPA:353.2	0	1
CAMO-17-141978	R-33 S2	REG	EPA:365.4	0	1
CAMO-17-141978	R-33 S2	REG	SM:A2340B	0	1
CAMO-17-141978	R-33 S2	REG	SW-846:6010C	0	17
CAMO-17-141978	R-33 S2	REG	SW-846:6020	0	11
CAMO-17-141978	R-33 S2	REG	SW-846:6850	0	1
CAMO-17-141993	R-33 S1	REG	EPA:170.0	0	1
CAMO-17-141993	R-33 S1	REG	EPA:245.2	0	1
CAMO-17-141993	R-33 S1	REG	EPA:335.4	0	1
CAMO-17-141993	R-33 S1	REG	EPA:351.2	0	1
CAMO-17-141993	R-33 S1	REG	SW-846:9060	0	1
CAMO-17-141994	R-33 S2	REG	EPA:170.0	0	1
CAMO-17-141994	R-33 S2	REG	EPA:245.2	0	1
CAMO-17-141994	R-33 S2	REG	EPA:335.4	0	1
CAMO-17-141994	R-33 S2	REG	EPA:351.2	0	1
CAMO-17-141994	R-33 S2	REG	SW-846:9060	0	1



August 25, 2017

gel.com

Ms. Nita Patel
Los Alamos National Laboratory
TA-00, SM1237, Rm104C
Los Alamos, New Mexico 87545

Re: LANL- WQH Water Samples
Work Order: 429570
SDG: 2017-2284

Dear Ms. Patel:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the following analytical results for the sample(s) we received on August 03, 2017, and analyzed for General Chemistry, Metals and Perchlorates by LCMSMS. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4485.

Sincerely,

Katrina Hiott for
Valerie Davis
Project Manager

Chain of Custody: 2017-2284
Enclosures



ARS International, LLC (ARS-LANS-MTOA6-25093-GEL)
LANL- WQH Water Samples
Work Order #: 429570
SDG: 2017-2284

Table of Contents

Case Narrative.....	1
Chain of Custody and Supporting Documentation.....	5
Data Review Qualifier Flag Definition Sheet.....	10
Perchlorates by LCMSMS Analysis.....	13
Case Narrative.....	14
Sample Data Summary.....	20
Quality Control Summary.....	23
Quality Control Data.....	26
Metals Analysis.....	32
Case Narrative.....	33
Sample Data Summary.....	39
Quality Control Summary.....	48
General Chem Analysis.....	62
Case Narrative.....	63
Sample Data Summary.....	93
Quality Control Summary.....	100

Case Narrative

**Case Narrative for
ARS International, LLC (ARS-LANS-MTOA6-25093-GEL)
LANL- WQH Water Samples
Workorder #: 429570
SDG # : 2017-2284**

August 25, 2017

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt The samples arrived at GEL Laboratories LLC, Charleston, South Carolina on August 03, 2017 for analysis. The samples were delivered with proper chain of custody documentation and signatures. The samples were screened according to GEL Standard Operating Procedure. All sample containers arrived without any visible signs of tampering or breakage. Containers were checked for pH, where appropriate, and matched the preservative as documented on the accompanying chain of custody. Shipping container temperature was within specification (0 - 6C). Shipping container temperatures were checked, documented, and within specifications. There are no additional comments concerning sample receipt.

Sample Identification The laboratory received the following samples:

<u>Laboratory ID</u>	<u>Client ID</u>
429570001	CAMO-17-141977
429570002	CAMO-17-141993
429570003	CAMO-17-141978
429570004	CAMO-17-141994

Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: Case Narrative, Chain of Custody, Cooler Receipt Checklist, Data Package Qualifier Definitions and data from the following fractions: General Chemistry, Metals and Perchlorates by LCMSMS.

I certify that this data report is in compliance with the terms and conditions of the subcontract and task order, both technically and for completeness, for other than the conditions detailed in the attached case narrative.


Katrina Hiott for
Valerie Davis
Project Manager

List of current GEL Certifications as of 25 August 2017

State	Certification
Alaska	UST-0110
Arkansas	88-0651
CLIA	42D0904046
California	2940
Colorado	SC00012
Connecticut	PH-0169
Delaware	SC00012
DoD ELAP/ ISO17025 A2LA	2567.01
Florida NELAP	E87156
Foreign Soils Permit	P330-15-00283, P330-15-00253
Georgia	SC00012
Georgia SDWA	967
Hawaii	SC00012
Idaho Chemistry	SC00012
Idaho Radiochemistry	SC00012
Illinois NELAP	200029
Indiana	C-SC-01
Kansas NELAP	E-10332
Kentucky SDWA	90129
Kentucky Wastewater	90129
Louisiana NELAP	03046 (AI33904)
Louisiana SDWA	LA170010
Maryland	270
Massachusetts	M-SC012
Michigan	9976
Mississippi	SC00012
Nebraska	NE-OS-26-13
Nevada	SC000122018-1
New Hampshire NELAP	205415
New Jersey NELAP	SC002
New Mexico	SC00012
New York NELAP	11501
North Carolina	233
North Carolina SDWA	45709
North Dakota	R-158
Oklahoma	9904
Pennsylvania NELAP	68-00485
Puerto Rico	SC00012
S.Carolina Radchem	10120002
South Carolina Chemistry	10120001
Tennessee	TN 02934
Texas NELAP	T104704235-17-12
Utah NELAP	SC000122017-23
Vermont	VT87156
Virginia NELAP	460202
Washington	C780
West Virginia	997404

Chain of Custody and Supporting Documentation

SAMPLE RECEIPT & REVIEW FORM

Client: ESHL		SDG/AR/COC/Work Order: 429570	
Received By: ZKW		Date Received: 8/3/17	
Carrier and Tracking Number		Circle Applicable: FedEx Express <input checked="" type="checkbox"/> FedEx Ground <input type="checkbox"/> UPS <input type="checkbox"/> Field Services <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/> 5908 1782 4682 - 4C 5908 1782 4649 - 4C 5908 1782 4671 - 4C 5908 1782 4660 - 4C 5908 1782 4650 - 4C 5908 1782 4638 - 23C	
Suspected Hazard Information		Yes	No
Shipped as a DOT Hazardous?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
COC/Samples marked or classified as radioactive?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Is package, COC, and/or Samples marked HAZ?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Hazard Class Shipped:		UN#:	
Maximum Net Counts Observed* (Observed Counts - Area Background Counts):		Classified as: Rad 1 Rad 2 Rad 3	
If yes, select Hazards below, and contact the GEL Safety Group.		PCB's Flammable Foreign Soil RCRA Asbestos Beryllium Other:	
Sample Receipt Criteria		Yes	NA
Comments/Qualifiers (Required for Non-Conforming Items)			
1	Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2	Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
3	Samples requiring cold preservation within (0 ≤ 6 deg. C)?*	<input checked="" type="checkbox"/>	<input type="checkbox"/>
4	Daily check performed and passed on IR temperature gun?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
5	Sample containers intact and sealed?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
6	Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
7	Do any samples require Volatile Analysis?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
8	Samples received within holding time?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
9	Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
10	Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
11	Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
12	Are sample containers identifiable as GEL provided?	<input type="checkbox"/>	<input checked="" type="checkbox"/>
13	COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Comments (Use Continuation Form if needed): * Metals Cont. for -143205 rec'd unpreserved * ISOPU Conts. for -134282 rec'd unpreserved, preserved w/ HNO3 upon arrival (Lot# 170530)			

ORIGIN ID:SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB.
TA00 BLDG 1237 DPU 03

SHIP DATE: 02AUG17
ACTWGT: 49.0 LB MAN
CAD: 0014176/CAFE2916

LOS ALAMOS, NM 87545
UNITED STATES US



BILL SENDER

TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD

CHARLESTON SC 29407

(843) 656-8171

REF: 21PD0ASRGW04BAGWEO



FedEx
Express



ORIGIN ID:SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB.
TA00 BLDG 1237 DPU 03

J2AUG17
0 LB MAN
176/CAFE291

LOS ALAMOS, NM 87545
UNITED STATES US



BILL SENDER

TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD

CHARLESTON SC 29407

(843) 656-8171

REF: 21PD0ASRSW12CAWC00



FedEx
Express



2 of 2

THU - 03 AUG 10:30A
PRIORITY OVERNIGHT

MPS# 5908 1782 4650

Mstr# 5908 1782 4649

0201

X7 RBWA

29407

SC-US CHS

0251/80211111



1 of 2

THU - 03 AUG 10:30A
PRIORITY OVERNIGHT

TRK# 5908 1782 4660

MASTER

X7 RBWA

29407

SC-US CHS



X7 RBWA

2 of 2

MPS# 5908 1782 4671

Mstr# 5908 1782 4660


0201

THU - 03 AUG 10:30A
PRIORITY OVERNIGHT

29407

SC-US

CHS



FedEx
Express

0151015061201111

CHARLESTON SC 29407

(843) 656-8171

REF: 21PD0ASRSW12CAWC00

TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD

ORIGIN ID:SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB.
TA00 BLDG 1237 DPU 03
LOS ALAMOS, NM 87545
UNITED STATES US

SHIP DATE: 02AUG17
ACTWGT: 44.0 LB MAN
CAD: 0014176/CAFE2916

BILL SENDER

ORIGIN ID:SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB.
TA00 BLDG 1237 DPU 03

LOS ALAMOS, NM 87545
UNITED STATES US

SHIP DATE: 02AUG17
ACTWGT: 65.0 LB MAN
CAD: 0014176/CAFE2916

BILL SENDER


TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD

CHARLESTON SC 29407

(843) 556-8171

REF: 5W000AJHWC0000000

230



FedEx
Express



TRK# 5908 1782 4638
0201

THU - 03 AUG 10:30A
PRIORITY OVERNIGHT

X7 RBWA

29407
SC-US CHS



ORIGIN ID:SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB.
TA00 BLDG 1237 DPU 03

LOS ALAMOS, NM 87545
UNITED STATES US

SHIP DATE: 02AUG17
ACTWGT: 62.0 LB MAN
CAD: 0014176/CAFE2916

BILL SENDER


TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD

CHARLESTON SC 29407

(843) 556-8171

REF: 21PD0ASRSW12CAWC00

60c



FedEx
Express



TRK# 5908 1782 4682
0201

THU - 03 AUG 10:30A
PRIORITY OVERNIGHT

X7 RBWA

29407
SC-US CHS



SHIP DATE: 02AUG17
ACTWGT: 42.0 LB MAN
CAD: 0014176/CAFE2916

BILL SENDER

ORIGIN ID:SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB.
TA00 BLDG 1237 DPU 03

LOS ALAMOS, NM 87545
UNITED STATES US

TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD



CHARLESTON SC 29407

(843) 556-8171

REF: 21PD0ASRGW04BAGWEO

40c

FedEx
Express




THU - 03 AUG 10:30A
PRIORITY OVERNIGHT

TRK# 5908 1782 4649
0201

MASTER

X7 RBWA

29407
SC-US CHS



09-434 HIT2 08/15 3.3

Data Review Qualifier Flag Definition Sheet

Data Review Qualifier Definitions

Qualifier	Explanation
-----------	-------------

*	A quality control analyte recovery is outside of specified acceptance criteria
**	Analyte is a surrogate compound
<	Result is less than value reported
>	Result is greater than value reported
^	RPD of sample and duplicate evaluated using +/-RL. Concentrations are <5X the RL
A	The TIC is a suspected aldol-condensation product
B	Target analyte was detected in the associated blank
B	Metals-Either presence of analyte detected in the associated blank, or MDL/IDL < sample value < PQL
BD	Results are either below the MDC or tracer recovery is low
C	Analyte has been confirmed by GC/MS analysis
D	Results are reported from a diluted aliquot of the sample
d	5-day BOD-The 2:1 depletion requirement was not met for this sample
E	Organics-Concentration of the target analyte exceeds the instrument calibration range
E	Metals-%difference of sample and SD is >10%. Sample concentration must meet flagging criteria
H	Analytical holding time was exceeded
h	Preparation or preservation holding time was exceeded
J	Value is estimated
N	Metals-The Matrix spike sample recovery is not within specified control limits
N	Organics-Presumptive evidence based on mass spectral library search to make a tentative identification of the analyte (TIC). Quantitation is based on nearest internal standard response factor
N/A	Spike recovery limits do not apply. Sample concentration exceeds spike concentration by 4X or more
ND	Analyte concentration is not detected above the reporting limit
UI	Gamma Spectroscopy-Uncertain identification
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
Y	QC Samples were not spiked with this compound
Z	Paint Filter Test-Particulates passed through the filter, however no free liquids were observed.

P Organics-The concentrations between the primary and confirmation columns/detectors is >40% difference.
For HPLC, the difference is >70%.

U Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.

Perchlorates by LCMSMS Analysis

Case Narrative

**Perchlorates by LCMSMS
Technical Case Narrative
ARS International, LLC (ARSL)
SDG #: 2017-2284
Work Order #: 429570**

Method/Analysis Information

Procedure: **Definitive Low Level Perchlorate Analysis Utilizing Liquid Chromatography/Mass Spectrometry/Mass Spectrometry (LC/MS/MS) by EPA Method 6850 Modified (6850M)**

Analytical Method: SW-846:6850

Prep Method: SW-846:6850

Analytical Batch Number: 1689374

Prep Batch Number: 1689372

Sample Analysis

Sample ID	Client ID
429570001	429570001 (CAMO-17-141977)
429570003	429570003 (CAMO-17-141978)
1203848618	Interference Check Sample (ICS)
1203848363	Method Blank (MB)
1203848364	Laboratory Control Sample (LCS)
1203848365	429570001(CAMO-17-141977) Matrix Spike (MS)
1203848366	429570001(CAMO-17-141977) Matrix Spike Duplicate (MSD)

The samples in this SDG were analyzed on an "as received" basis.

Preparation/Analytical Method Verification

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-OA-E-067 REV# 14.

Calibration Information

Initial Calibration

All initial calibration requirements have been met for this SDG. Due to software constraints, all Initial Calibration Blanks must be designated as IPB001.

ICV Requirements

All associated initial calibration verification standard(s) (ICV) met the acceptance criteria.

CCB Requirements

All continuing calibration blanks (CCB) bracketing the analyses associated with this batch were within acceptance criteria.

CCV Requirements

All continuing calibration checks (CCV) requirements were met by all bracketing CCV standards.

Low Level Standard (CRI) Requirements

All low level calibration verification (CRI) requirements were met by all bracketing CRI standards.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Interference Check Sample (ICS)

The ICS spike recoveries met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recoveries met the acceptance limits.

QC Sample Designation

Client sample 429570001 (CAMO-17-141977) was chosen for matrix spike and matrix spike duplicate analysis.

Matrix Spike (MS) Recovery Statement

The MS recoveries were within the established acceptance limits.

MS/MSD Relative Percent Difference (RPD) Statement

The RPDs between the MS and MSD met the acceptance limits.

Internal Standard Area Acceptance

The internal standard areas were within the required acceptance criteria for all samples and QC.

Retention Time

During the analysis of Perchlorate by LC/MS/MS, retention time shifts are commonly observed. These retention time shifts, which are caused by fouling of the column by the sample matrices, are problematic when the retention time is used as one of the criterion for confirmation. To overcome this problem, a known amount of O(18) labeled Perchlorate was added to each sample as a retention time standard. The presence of Perchlorate was confirmed by the relative retention time (RRT) of the Perchlorate peak and the O(18) standard. A RRT window of 0.98 to 1.02, as required by DOD QSM 5.0, has been used. In addition to the isotopic ratio, the presence of Perchlorate in the samples associated with this data package have been confirmed using the relative retention criteria stated above, not the absolute retention time.

Technical Information

Holding Time Specifications

All samples in this SDG in this analytical batch met the specified holding time. GEL assigns holding times based

on the associated methodology, which assigns the date and time from sample collection of sample receipt. Those holding times expressed in hours are calculated in the AlphaLIMS system. Those holding times expressed as days expire at midnight on the day of expiration.

Preparation/Analytical Method Verification

All procedures were performed as stated in the SOP.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-extraction/Re-analysis

Re-extractions or re-analyses were not required in this SDG.

Miscellaneous Information

Manual Integrations

Manual integrations were not required for any data file associated with this SDG.

Method Comments

The samples in this SDG were not originally analyzed using EPA Method 314.0.

Additional Comments

The Perchlorate Isotope Ratio on the Form I may differ slightly from the ratio on the corresponding raw data due to rounding rules and/or significant figures or due to software limitations when there are manual integrations, dilutions or other factors. The ratio value of the Form I is the correct value. The retention time marker, Perchlorate-O (18), is added to all samples, instrument blanks, and standards prior to injection. It is used to verify the retention time of Perchlorate and Perchlorate-101 and to insure an accurate injection occurred. Due to various anions affecting the recovery of Perchlorate-O (18) and not Perchlorate and Perchlorate-101, the calibration curves of Perchlorate and Perchlorate-101 are internally corrected for using Perchlorate-O (18).

Perchlorate Isotope Ratio

The Perchlorate isotope ratio met acceptance criteria for all samples and QC samples. Please see the isotope ratio criteria in the Miscellaneous Section.

System Configuration

The laboratory utilizes a Waters LC 2795 liquid chromatography instrument for Perchlorate analysis. It is coupled with a Micromass Quattro Ultima Mass Spectrometer/Mass Spectrometer. It is designated as LCMSMS #2. It is fitted with an electrospray probe that is operated in the negative electrospray ionization mode for Perchlorate analysis. The laboratory may also utilize an Agilent 1100 liquid chromatography instrument for Perchlorate analysis. It is coupled with an Applied Biosystems 4000 Mass Spectrometer/Mass Spectrometer, designated as LCMSMS #3 or LCMSMS #4. It is also fitted with an electrospray probe that is operated in the negative electrospray ionization mode for Perchlorate analysis.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. An electronic signature page inserted after the case narrative will include the data validator's signature and title. The signature page also includes the data qualifiers used in the fractional package. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Chromatographic Columns

The LC-MS/MS Perchlorate analysis was performed on a Quatro Ultima LC/MS/MS.

Chromatographic separation of Perchlorate is accomplished through analysis on the following anion column:

Dionex: IonPac AG-16 2 x 50 mm.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report for

ARSL004 ARS International, LLC (ARS-LANS-MTOA6-25093-GEL)

Client SDG: 2017-2284 GEL Work Order: 429570


The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- ** Analyte is a surrogate compound
- J Value is estimated
- U Analyte was analyzed for, but not detected above the MDL, MDA, MDC or LOD.
- DL Indicates that sample is diluted.
- RA Indicates that sample is re-analyzed without re-extraction.
- RE Indicates that sample is re-extracted.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature: 

Name: Charles Wilson

Date: 15 AUG 2017

Title: Analyst II

Sample Data Summary

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: SW846 6850 ModifiedMatrix: WATERExtraction Batch ID: 1689372Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

CAMO-17-141977Date Received: 03-AUG-17GEL Job No (SDG): 2017-2284GEL Sample ID: 429570001Date Filtered: 07-AUG-17Injection Volume (uL): 20%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.418	ug/L		1	12-AUG-17 03:25	per0811070a
	Perchlorate Isotope Ratio			3.07			1	12-AUG-17 03:25	per0811070a
14797-73-0	Perchlorate-101	.05	.2	0.399	ug/L		1	12-AUG-17 03:25	per0811070a
	Perchlorate-O(18)			0.446	ug/L		1	12-AUG-17 03:25	per0811070a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: SW846 6850 ModifiedMatrix: WATERExtraction Batch ID: 1689372Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

CAMO-17-141978Date Received: 03-AUG-17GEL Job No (SDG): 2017-2284GEL Sample ID: 429570003Date Filtered: 07-AUG-17Injection Volume (uL): 20%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.362	ug/L		1	12-AUG-17 04:05	per0811073a
	Perchlorate Isotope Ratio			3.08			1	12-AUG-17 04:05	per0811073a
14797-73-0	Perchlorate-101	.05	.2	0.344	ug/L		1	12-AUG-17 04:05	per0811073a
	Perchlorate-O(18)			0.440	ug/L		1	12-AUG-17 04:05	per0811073a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Quality Control Summary

Perchlorate Laboratory Control Sample

Lab Name: General Engineering Laboratories

Lab Code: GEL

GEL Job No. (SDG): 2017-2284

Extract Batch Code: 1689372

Date Filtered: 07-AUG-17

Matrix: WATER

Sample ID: 1203848364

Analyte^	True	Found	Units	%Rec	Q	Control Limits
Perchlorate	0.200	.199	ug/L	99		85 - 115
Perchlorate Isotope Ratio		3.01				-
Perchlorate-101	0.200	.194	ug/L	97		85 - 115
Perchlorate-O(18)		.443	ug/L			-

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

Perchlorate Spike/Spike Duplicate Summary

Lab Name: General Engineering Laboratories

Lab Code: GEL

GEL Job No (SDG): 2017-2284

Extract Batch Code: 1689372

Date Extracted: 07-AUG-17

GEL MS/PS ID: 1203848365

Client ID: CAMO-17-141977

GEL MSD/PSD ID: 1203848366

QC Type: MS

Compound^	Spike Added	Sample Conc	Units	MS Conc	MS Rec #	MSD Conc	MSD Rec #	RPD #	RPD Limit	Recovery Limit
Perchlorate	0.200	0.418	ug/L	0.602	92	.614	98	2	30	75 - 125
Perchlorate Isotope Ratio	0	3.07		2.96		2.95		0		-
Perchlorate-101	0.200	0.399	ug/L	0.596	99	.61	105	2	30	75 - 125
Perchlorate-O(18)	0	0.446	ug/L	0.438		.434		1		-

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

Quality Control Data

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: EPA 6850 ModifiedMatrix: WATERExtraction Batch ID: 1689372Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

MBDate Received: 07-AUG-17GEL Job No (SDG): 2017-2284GEL Sample ID: 1203848363Date Filtered: 07-AUG-17Injection Volume (uL): 20%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.050	ug/L	U	1	12-AUG-17 02:04	per0811064a
	Perchlorate Isotope Ratio						1	12-AUG-17 02:04	per0811064a
14797-73-0	Perchlorate-101	.05	.2	0.050	ug/L	U	1	12-AUG-17 02:04	per0811064a
	Perchlorate-O(18)			0.429	ug/L		1	12-AUG-17 02:04	per0811064a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: EPA 6850 ModifiedMatrix: WATERExtraction Batch ID: 1689372Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

LCSDate Received: 07-AUG-17GEL Job No (SDG): 2017-2284GEL Sample ID: 1203848364Date Filtered: 07-AUG-17Injection Volume (uL): 20%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.199	ug/L	J	1	12-AUG-17 02:18	per0811065a
	Perchlorate Isotope Ratio			3.01			1	12-AUG-17 02:18	per0811065a
14797-73-0	Perchlorate-101	.05	.2	0.194	ug/L	J	1	12-AUG-17 02:18	per0811065a
	Perchlorate-O(18)			0.443	ug/L		1	12-AUG-17 02:18	per0811065a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: SW846 6850 ModifiedMatrix: WATERExtraction Batch ID: 1689372Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

ICS

Date Received:

GEL Job No (SDG): 2017-2284GEL Sample ID: 1203848618Date Filtered: 07-AUG-17Injection Volume (uL): 20

%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.230	ug/L		1	12-AUG-17 02:31	per0811066a
	Perchlorate Isotope Ratio			2.78			1	12-AUG-17 02:31	per0811066a
14797-73-0	Perchlorate-101	.05	.2	0.243	ug/L		1	12-AUG-17 02:31	per0811066a
	Perchlorate-O(18)			0.442	ug/L		1	12-AUG-17 02:31	per0811066a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: SW846 6850 ModifiedMatrix: WATERExtraction Batch ID: 1689372Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

CAMO-17-141977MSDate Received: 03-AUG-17GEL Job No (SDG): 2017-2284GEL Sample ID: 1203848365Date Filtered: 07-AUG-17Injection Volume (uL): 20%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.602	ug/L		1	12-AUG-17 03:38	per0811071a
	Perchlorate Isotope Ratio			2.96			1	12-AUG-17 03:38	per0811071a
14797-73-0	Perchlorate-101	.05	.2	0.596	ug/L		1	12-AUG-17 03:38	per0811071a
	Perchlorate-O(18)			0.438	ug/L		1	12-AUG-17 03:38	per0811071a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: SW846 6850 ModifiedMatrix: WATERExtraction Batch ID: 1689372Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

CAMO-17-141977MSDDate Received: 03-AUG-17GEL Job No (SDG): 2017-2284GEL Sample ID: 1203848366Date Filtered: 07-AUG-17Injection Volume (uL): 20%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.614	ug/L		1	12-AUG-17 03:52	per0811072a
	Perchlorate Isotope Ratio			2.95			1	12-AUG-17 03:52	per0811072a
14797-73-0	Perchlorate-101	.05	.2	0.610	ug/L		1	12-AUG-17 03:52	per0811072a
	Perchlorate-O(18)			0.434	ug/L		1	12-AUG-17 03:52	per0811072a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Metals Analysis

Case Narrative

Metals
Technical Case Narrative
ARS International, LLC (ARSL)
SDG #: 2017-2284
Work Order #: 429570

Sample ID	Client ID
429570001	CAMO-17-141977
429570002	CAMO-17-141993
429570003	CAMO-17-141978
429570004	CAMO-17-141994
1203845800	Method Blank (MB) ICP
1203845801	Laboratory Control Sample (LCS)
1203845804	429570001(CAMO-17-141977L) Serial Dilution (SD)
1203845802	429570001(CAMO-17-141977D) Sample Duplicate (DUP)
1203845803	429570001(CAMO-17-141977S) Matrix Spike (MS)
1203845843	Method Blank (MB) ICP-MS
1203845844	Laboratory Control Sample (LCS)
1203845847	429570001(CAMO-17-141977L) Serial Dilution (SD)
1203845845	429570001(CAMO-17-141977D) Sample Duplicate (DUP)
1203845846	429570001(CAMO-17-141977S) Matrix Spike (MS)
1203860686	Method Blank (MB) CVAA
1203860687	Laboratory Control Sample (LCS)
1203860690	429620001(CASA-17-142006L) Serial Dilution (SD)
1203860688	429620001(CASA-17-142006D) Sample Duplicate (DUP)
1203860689	429620001(CASA-17-142006S) Matrix Spike (MS)

Sample Analysis

Samples 429570001,002,003 and 004 in this SDG were analyzed for metals and mercury on an "as received" basis.

Method/Analysis Information

Analytical Batch:	1688319, 1688337, 1694628 and 1695942
Prep Batch :	1688317, 1688336 and 1694626
Standard Operating Procedures:	GL-MA-E-013 REV# 29, GL-MA-E-006 REV# 13, GL-MA-E-014 REV# 31, GL-MA-E-010 REV# 35 and GL-GC-E-107 REV# 10
Analytical Method:	SW846 3005A/6010C, SW846 3005A/6020A, EPA 245.2 1974 and SM:A2340B
Prep Method :	SW846 3005A and EPA 245.1/245.2 Prep

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

System Configuration

The Hardness as CaCO₃ is calculated from Calcium and Magnesium results.

The Metals analysis-ICP was performed on a P E 5300 Optima radial/axial-viewing inductively coupled plasma atomic emission spectrometer. The instrument is equipped with an ESI SC-FAST introduction, cyclonic spray chamber, and yttrium or scandium internal standard.

The Metals analysis-Mercury was performed on a Perkin-Elmer Flow Injection Mercury System (FIMS-100) automated mercury analyzer. The instrument consists of a cold vapor atomic absorption spectrometer set to detect mercury at a wavelength of 253.7 nm.

The Metals analysis - ICPMS was performed on a PerkinElmer NexION 350X ICPMS. The instrument is equipped with a ESI PFA-ST nebulizer, quadrupole mass spectrometer, dual mode electron multiplier detector, and Kinetic Energy Discrimination (KED) technology. Internal standards of scandium, germanium, indium, tantalum, and/or lutetium were utilized to cover the mass spectrum.

Calibration Information

Instrument Calibration

All initial calibration requirements have been met for this sample delivery group (SDG).

CRDL/PQL Requirements

The PQL standard recoveries for SW846 6010C or 6010D met the control limits with the exception of potassium, silica and sodium. Client sample concentrations were less than the MDL or greater than two times the PQL; therefore the data were not adversely affected. 429570001 (CAMO-17-141977) and 429570003 (CAMO-17-141978)-ICP.

ICSA/ICSAB Statement

All interference check samples (ICSA and ICSAB) associated with this SDG met the established acceptance criteria. For the ICP-MS analysis, the ICSA solution contains analyte concentrations which are verified trace impurities indigenous to the purchased standard.

Continuing Calibration Blanks (CCB) Requirements

All continuing calibration blanks (CCB) bracketing this batch met the established acceptance criteria.

Continuing Calibration Verification (CCV) Requirements

All continuing calibration verifications (CCV) bracketing this SDG met the acceptance criteria.

Quality Control (QC) Information

Method Blank (MB) Statement

The MBs analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recoveries met the acceptance limits.

Quality Control (QC) Sample Statement

The following samples were selected as the quality control (QC) samples for this SDG: 429570001 (CAMO-17-141977)-ICP and ICP-MS and 429620001 (CASA-17-142006)-CVAA.

Matrix Spike (MS/MSD) Recovery Statement

The percent recoveries (%R) obtained from the MS/MSD analyses are evaluated when the sample concentration

is less than four times (4X) the spike concentration added. The matrix spike met the recommended quality control acceptance criteria for percent recoveries for all applicable analytes.

Duplicate Relative Percent Difference (RPD) Statement

The RPD obtained from the designated sample duplicate (DUP) is evaluated based on acceptance criteria of 20% when the sample is >5X the contract required reporting limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control of +/-RL is used to evaluate the DUP results. The relative percent differences (RPD) between the sample and its duplicate (DUP) were within acceptable limits for all applicable analytes.

Serial Dilution % Difference Statement

All applicable analytes in the serial dilution (SDILT) demonstrated acceptable correlation to its associated sample and met the established acceptance percent difference criteria.

Technical Information

Holding Time Specifications

GEL assigns holding times based on the associated methodology. Holding time is measured by comparison of the date and time of sample collection to the date and time of sample preparation and analysis. Those holding times expressed in hours are calculated in the AlphaLIMS system. Those holding times expressed as days expire at midnight on the day of expiration. All samples in this SDG met the specified holding time.

Preparation/Analytical Method Verification

All procedures were performed as stated in the SOP.

Sample Dilutions

The samples in this SDG did not require dilutions.

Preparation Information

The samples in this SDG were not diluted and were prepared according to the cited SOP.

Miscellaneous Information

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. An electronic signature page inserted after the case narrative will include the data validator's signature and title. The signature page also includes the data qualifiers used in the fractional package. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Additional Comments

Total Hardness by Calculation is determined using the results of Total Calcium (Ca) and Total Magnesium (Mg) determined by ICP or ICP-MS.

$$\text{Hardness} = 2.497 (\text{Ca}) + 4.118 (\text{Mg})$$

Please refer to the Total Ca and Total Mg data to validate results appearing on the Hardness Summary sheet. Both results are in the Inorganic/metals section of the package. There is no Batch QC for calculated results, and thus no QC Summary for the Hardness by Calculation Batch. The MDLs and PQLs are calculated using the higher of the two calculated values of Ca or Mg.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report for

ARSL004 ARS International, LLC (ARS-LANS-MTOA6-25093-GEL)

Client SDG: 2017-2284 GEL Work Order: 429570

The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- J Value is estimated
- U Analyte was analyzed for, but not detected above the MDL, MDA, MDC or LOD.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature:



Name: Nik-Cole Elmore

Date: 29 AUG 2017

Title: Data Validator

Sample Data Summary

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284**CONTRACT:** ESHL00114**METHOD TYPE:** EPA**SAMPLE ID:** 429570001**BASIS:** As Received**DATE COLLECTED** 01-AUG-17**CLIENT ID:** CAMO-17-141977**LEVEL:** Low**DATE RECEIVED** 03-AUG-17**MATRIX:** W**%SOLIDS:** 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
7439-97-6	Mercury	0.067	ug/L	U	0.067	0.2	0.2	1	AV	MTM1	08/24/17 10:38	082417W1-3	1694628

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284

CONTRACT: ESHL00114

METHOD TYPE: SW846

SAMPLE ID: 429570001

BASIS: As Received

DATE COLLECTED 01-AUG-17

CLIENT ID: CAMO-17-141977

LEVEL: Low

DATE RECEIVED 03-AUG-17

MATRIX: W

%SOLIDS: 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
7429-90-5	Aluminum	68	ug/L	U	68	200	200	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-36-0	Antimony	1	ug/L	U	1	3	3	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-38-2	Arsenic	2	ug/L	U	2	5	5	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-39-3	Barium	31.7	ug/L		1	5	5	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-41-7	Beryllium	1	ug/L	U	1	5	5	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-42-8	Boron	15	ug/L	U	15	50	50	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-43-9	Cadmium	0.30	ug/L	U	0.3	1	1	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-70-2	Calcium	11700	ug/L		50	200	200	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-47-3	Chromium	5.19	ug/L	J	3	10	10	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-48-4	Cobalt	1	ug/L	U	1	5	5	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-50-8	Copper	3	ug/L	U	3	10	10	1	P	HSC	08/16/17 07:28	081617-1	1688319
7439-89-6	Iron	30	ug/L	U	30	100	100	1	P	HSC	08/16/17 07:28	081617-1	1688319
7439-92-1	Lead	0.50	ug/L	U	0.5	2	2	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7439-95-4	Magnesium	3870	ug/L		110	300	300	1	P	HSC	08/16/17 07:28	081617-1	1688319
7439-96-5	Manganese	2	ug/L	U	2	10	10	1	P	HSC	08/16/17 07:28	081617-1	1688319
7439-98-7	Molybdenum	1.18	ug/L		0.2	0.5	0.5	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-02-0	Nickel	1.31	ug/L	J	0.6	2	2	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-09-7	Potassium	1440	ug/L		50	150	150	1	P	HSC	08/16/17 07:28	081617-1	1688319
7782-49-2	Selenium	2	ug/L	U	2	5	5	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7631-86-9	Silica	73300	ug/L		53	213	213	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-22-4	Silver	0.30	ug/L	U	0.3	1	1	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-23-5	Sodium	11700	ug/L		100	300	300	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-24-6	Strontium	48.8	ug/L		1	5	5	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-28-0	Thallium	0.60	ug/L	U	0.6	2	2	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-31-5	Tin	3.29	ug/L	J	2.5	10	10	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-61-1	Uranium	0.830	ug/L		0.067	0.2	0.2	1	MS	BAJ	08/14/17 22:12	170814-2	1688337
7440-62-2	Vanadium	5.7	ug/L		1	5	5	1	P	HSC	08/16/17 07:28	081617-1	1688319
7440-66-6	Zinc	3.3	ug/L	U	3.3	10	10	1	P	HSC	08/16/17 07:28	081617-1	1688319

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284**CONTRACT:** ESHL00114**METHOD TYPE:****SAMPLE ID:** 429570001**BASIS:** As Received**DATE COLLECTED** 01-AUG-17**CLIENT ID:** CAMO-17-141977**LEVEL:** Low**DATE RECEIVED** 03-AUG-17**MATRIX:** W**%SOLIDS:** 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
	Hardness as CaCO3	45.3	mg/L		0.453	1.24	1.24	1		TXT1	08/25/17 14:47		1695942

Prep Information:

Analytical Batch	Prep Batch	Prep Method	Initial wt./vol.	Units	Final wt./vol.	Units	Date	Analyst
1688319	1688317	SW846 3005A	50	mL	50	mL	08/04/17	SXW1
1688337	1688336	SW846 3005A	50	mL	50	mL	08/04/17	SXW1
1694628	1694626	EPA 245.1/245.2 Prep	20	mL	20	mL	08/23/17	AXS5

***Analytical Methods:**

P SW846 3005A/6010C
MS SW846 3005A/6020A
AV EPA 245.2 1974

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284**CONTRACT:** ESHL00114**METHOD TYPE:** EPA**SAMPLE ID:** 429570002**BASIS:** As Received**DATE COLLECTED** 01-AUG-17**CLIENT ID:** CAMO-17-141993**LEVEL:** Low**DATE RECEIVED** 03-AUG-17**MATRIX:** W**%SOLIDS:** 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
7439-97-6	Mercury	0.067	ug/L	U	0.067	0.2	0.2	1	AV	MTM1	08/24/17 10:40	082417W1-3	1694628

Prep Information:

Analytical Batch	Prep Batch	Prep Method	Initial wt./vol.	Units	Final wt./vol.	Units	Date	Analyst
1694628	1694626	EPA 245.1/245.2 Prep	20	mL	20	mL	08/23/17	AXS5

***Analytical Methods:**

AV EPA 245.2 1974

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284**CONTRACT:** ESHL00114**METHOD TYPE:** EPA**SAMPLE ID:** 429570003**BASIS:** As Received**DATE COLLECTED** 01-AUG-17**CLIENT ID:** CAMO-17-141978**LEVEL:** Low**DATE RECEIVED** 03-AUG-17**MATRIX:** W**%SOLIDS:** 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
7439-97-6	Mercury	0.067	ug/L	U	0.067	0.2	0.2	1	AV	MTM1	08/24/17 10:41	082417W1-3	1694628

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284

CONTRACT: ESHL00114

METHOD TYPE: SW846

SAMPLE ID: 429570003

BASIS: As Received

DATE COLLECTED 01-AUG-17

CLIENT ID: CAMO-17-141978

LEVEL: Low

DATE RECEIVED 03-AUG-17

MATRIX: W

%SOLIDS: 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
7429-90-5	Aluminum	68	ug/L	U	68	200	200	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-36-0	Antimony	1	ug/L	U	1	3	3	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-38-2	Arsenic	2	ug/L	U	2	5	5	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-39-3	Barium	36.1	ug/L		1	5	5	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-41-7	Beryllium	1	ug/L	U	1	5	5	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-42-8	Boron	15	ug/L	U	15	50	50	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-43-9	Cadmium	0.30	ug/L	U	0.3	1	1	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-70-2	Calcium	11300	ug/L		50	200	200	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-47-3	Chromium	5.54	ug/L	J	3	10	10	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-48-4	Cobalt	1	ug/L	U	1	5	5	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-50-8	Copper	3	ug/L	U	3	10	10	1	P	HSC	08/16/17 07:25	081617-1	1688319
7439-89-6	Iron	30	ug/L	U	30	100	100	1	P	HSC	08/16/17 07:25	081617-1	1688319
7439-92-1	Lead	0.50	ug/L	U	0.5	2	2	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7439-95-4	Magnesium	4170	ug/L		110	300	300	1	P	HSC	08/16/17 07:25	081617-1	1688319
7439-96-5	Manganese	2	ug/L	U	2	10	10	1	P	HSC	08/16/17 07:25	081617-1	1688319
7439-98-7	Molybdenum	1.01	ug/L		0.2	0.5	0.5	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-02-0	Nickel	0.60	ug/L	U	0.6	2	2	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-09-7	Potassium	2400	ug/L		50	150	150	1	P	HSC	08/16/17 07:25	081617-1	1688319
7782-49-2	Selenium	2	ug/L	U	2	5	5	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7631-86-9	Silica	78500	ug/L		53	213	213	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-22-4	Silver	0.30	ug/L	U	0.3	1	1	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-23-5	Sodium	11000	ug/L		100	300	300	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-24-6	Strontium	47.7	ug/L		1	5	5	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-28-0	Thallium	0.60	ug/L	U	0.6	2	2	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-31-5	Tin	2.5	ug/L	U	2.5	10	10	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-61-1	Uranium	0.926	ug/L		0.067	0.2	0.2	1	MS	BAJ	08/14/17 22:28	170814-2	1688337
7440-62-2	Vanadium	5.42	ug/L		1	5	5	1	P	HSC	08/16/17 07:25	081617-1	1688319
7440-66-6	Zinc	3.3	ug/L	U	3.3	10	10	1	P	HSC	08/16/17 07:25	081617-1	1688319

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284**CONTRACT:** ESHL00114**METHOD TYPE:****SAMPLE ID:** 429570003**BASIS:** As Received**DATE COLLECTED** 01-AUG-17**CLIENT ID:** CAMO-17-141978**LEVEL:** Low**DATE RECEIVED** 03-AUG-17**MATRIX:** W**%SOLIDS:** 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
	Hardness as CaCO3	45.4	mg/L		0.453	1.24	1.24	1		TXT1	08/25/17 14:47		1695942

Prep Information:

Analytical Batch	Prep Batch	Prep Method	Initial wt./vol.	Units	Final wt./vol.	Units	Date	Analyst
1688319	1688317	SW846 3005A	50	mL	50	mL	08/04/17	SXW1
1688337	1688336	SW846 3005A	50	mL	50	mL	08/04/17	SXW1
1694628	1694626	EPA 245.1/245.2 Prep	20	mL	20	mL	08/23/17	AXS5

***Analytical Methods:**

P SW846 3005A/6010C
MS SW846 3005A/6020A
AV EPA 245.2 1974

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 2017-2284**CONTRACT:** ESHL00114**METHOD TYPE:** EPA**SAMPLE ID:** 429570004**BASIS:** As Received**DATE COLLECTED** 01-AUG-17**CLIENT ID:** CAMO-17-141994**LEVEL:** Low**DATE RECEIVED** 03-AUG-17**MATRIX:** W**%SOLIDS:** 0

CAS	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
7439-97-6	Mercury	0.067	ug/L	U	0.067	0.2	0.2	1	AV	MTM1	08/24/17 10:43	082417W1-3	1694628

Prep Information:

Analytical Batch	Prep Batch	Prep Method	Initial wt./vol.	Units	Final wt./vol.	Units	Date	Analyst
1694628	1694626	EPA 245.1/245.2 Prep	20	mL	20	mL	08/23/17	AXS5

***Analytical Methods:**

AV EPA 245.2 1974

Quality Control Summary

METALS
-3b-
PREPARATION BLANK SUMMARY

SDG NO. 2017-2284

Contract: ESHL00114

Matrix: W

<u>Sample ID</u>	<u>Analyte</u>	<u>Result</u>	<u>Units</u>	<u>Acceptance Window</u>	<u>Conc Qual</u>	<u>M*</u>	<u>MDL</u>	<u>RDL</u>
1203845800	Aluminum	68	ug/L	+/-200	U	P	68	200
	Barium	1	ug/L	+/-5	U	P	1	5
	Beryllium	1	ug/L	+/-5	U	P	1	5
	Boron	15	ug/L	+/-50	U	P	15	50
	Calcium	50	ug/L	+/-200	U	P	50	200
	Cobalt	1	ug/L	+/-5	U	P	1	5
	Copper	3	ug/L	+/-10	U	P	3	10
	Iron	30	ug/L	+/-100	U	P	30	100
	Magnesium	110	ug/L	+/-300	U	P	110	300
	Manganese	2	ug/L	+/-10	U	P	2	10
	Potassium	50	ug/L	+/-150	U	P	50	150
	Silica	53	ug/L	+/-213	U	P	53	213
	Sodium	-216	ug/L	+/-300	J	P	100	300
	Strontium	1	ug/L	+/-5	U	P	1	5
	Tin	2.5	ug/L	+/-10	U	P	2.5	10
	Vanadium	1	ug/L	+/-5	U	P	1	5
	Zinc	3.3	ug/L	+/-10	U	P	3.3	10
1203845843	Arsenic	2	ug/L	+/-5	U	MS	2	5
	Cadmium	0.3	ug/L	+/-1	U	MS	0.3	1
	Chromium	3	ug/L	+/-10	U	MS	3	10
	Lead	0.5	ug/L	+/-2	U	MS	0.5	2
	Molybdenum	0.2	ug/L	+/-0.5	U	MS	0.2	0.5
	Nickel	0.6	ug/L	+/-2	U	MS	0.6	2
	Antimony	1	ug/L	+/-3	U	MS	1	3
	Selenium	2	ug/L	+/-5	U	MS	2	5
	Silver	0.3	ug/L	+/-1	U	MS	0.3	1
	Thallium	0.6	ug/L	+/-2	U	MS	0.6	2
	Uranium	0.067	ug/L	+/-0.2	U	MS	0.067	0.2
1203860686	Mercury	0.067	ug/L	+/-0.2	U	AV	0.067	0.2

*Analytical Methods:

P SW846 3005A/6010C
MS SW846 3005A/6020A
AV EPA 245.1/245.2

METALS

-5a-

Matrix Spike Summary

SDG NO. 2017-2284

Client ID: CAMO-17-141977S

Contract: ESHL00114

Level: Low

Matrix: WATER

% Solids:

Sample ID: 429570001

Spike ID: 1203845803

<u>Analyte</u>	<u>Units</u>	<u>Acceptance Limit</u>	<u>Spiked Result</u>	<u>C</u>	<u>Sample Result</u>	<u>C</u>	<u>Spike Added</u>	<u>% Recovery</u>	<u>Qual</u>	<u>M*</u>
Aluminum	ug/L	75-125	4820		68	U	5000	96.2		P
Barium	ug/L	75-125	529		31.7		500	99.4		P
Beryllium	ug/L	75-125	503		1	U	500	101		P
Boron	ug/L	75-125	537		15	U	500	105		P
Calcium	ug/L	75-125	16400		11700		5000	92.9		P
Cobalt	ug/L	75-125	510		1	U	500	102		P
Copper	ug/L	75-125	517		3	U	500	103		P
Iron	ug/L	75-125	5150		30	U	5000	103		P
Magnesium	ug/L	75-125	8890		3870		5000	100		P
Manganese	ug/L	75-125	495		2	U	500	98.8		P
Potassium	ug/L	75-125	6380		1440		5000	98.9		P
Silica	ug/L		82400		73300		10700	84.3	N/A	P
Sodium	ug/L	75-125	16800		11700		5000	101		P
Strontium	ug/L	75-125	533		48.8		500	96.8		P
Tin	ug/L	75-125	507		3.29	J	500	101		P
Vanadium	ug/L	75-125	512		5.7		500	101		P
Zinc	ug/L	75-125	494		3.3	U	500	98.3		P

*Analytical Methods:

P

SW846 3005A/6010C

METALS

-5a-

Matrix Spike Summary

SDG NO. 2017-2284 Client ID CAMO-17-141977S

Contract: ESHL00114 Level: Low

Matrix: WATER % Solids:

Sample ID: 429570001 Spike ID: 1203845846

<u>Analyte</u>	<u>Units</u>	<u>Acceptance Limit</u>	<u>Spiked Result</u>	<u>C</u>	<u>Sample Result</u>	<u>C</u>	<u>Spike Added</u>	<u>% Recovery</u>	<u>Qual</u>	<u>M*</u>
Antimony	ug/L	75-125	49		1	U	50	97.6		MS
Arsenic	ug/L	75-125	50.6		2	U	50	98		MS
Cadmium	ug/L	75-125	52.2		0.3	U	50	104		MS
Chromium	ug/L	75-125	55.6		5.19	J	50	101		MS
Lead	ug/L	75-125	52		0.5	U	50	104		MS
Molybdenum	ug/L	75-125	52		1.18		50	102		MS
Nickel	ug/L	75-125	52.4		1.31	J	50	102		MS
Selenium	ug/L	75-125	50		2	U	50	99.9		MS
Silver	ug/L	75-125	52.5		0.3	U	50	105		MS
Thallium	ug/L	75-125	48.6		0.6	U	50	96.9		MS
Uranium	ug/L	75-125	53		0.83		50	104		MS

*Analytical Methods:

MS SW846 3005A/6020A

METALS

-5a-

Matrix Spike Summary

SDG NO. 2017-2284 **Client ID:** CASA-17-142006S**Contract:** ESHL00114 **Level:** Low**Matrix:** WATER **% Solids:****Sample ID:** 429620001 **Spike ID:** 1203860689

<u>Analyte</u>	<u>Units</u>	<u>Acceptance Limit</u>	<u>Spiked Result</u>	<u>C</u>	<u>Sample Result</u>	<u>C</u>	<u>Spike Added</u>	<u>% Recovery</u>	<u>Qual</u>	<u>M*</u>
Mercury	ug/L	75-125	2.07		0.067	U	2	103		AV

*Analytical Methods:

AV EPA 245.1/245.2

Metals
-6-
Duplicate Sample Summary

SDG No.: 2017-2284

Lab Code: GEL

Contract: ESHL00114

Client ID: CAMO-17-141977D

Matrix: WATER

Level: Low

Sample ID: 429570001

Duplicate ID: 1203845802

Percent Solids for Dup: N/A

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M*
Aluminum	ug/L		68 U		68 U				P
Barium	ug/L	+/-20%	31.7		31.5		.754		P
Beryllium	ug/L		1 U		1 U				P
Boron	ug/L		15 U		15 U				P
Calcium	ug/L	+/-20%	11700		11800		.00851		P
Cobalt	ug/L		1 U		1 U				P
Copper	ug/L		3 U		3 U				P
Iron	ug/L		30 U		30 U				P
Magnesium	ug/L	+/-20%	3870		3830		.94		P
Manganese	ug/L		2 U		2 U				P
Potassium	ug/L	+/-20%	1440		1500		4.01		P
Silica	ug/L	+/-20%	73300		72600		1.04		P
Sodium	ug/L	+/-20%	11700		11300		3.62		P
Strontium	ug/L	+/-20%	48.8		48.2		1.35		P
Tin	ug/L		3.29 J		2.5 U		200		P
Vanadium	ug/L	+/-5	5.7		5.3		7.13		P
Zinc	ug/L		3.3 U		3.3 U				P

*Analytical Methods:

P SW846 3005A/6010C

Metals
-6-
Duplicate Sample Summary

SDG No.: 2017-2284

Lab Code: GEL

Contract: ESHL00114

Client ID: CAMO-17-141977D

Matrix: WATER

Level: Low

Sample ID: 429570001

Duplicate ID: 1203845845

Percent Solids for Dup: N/A

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M*
Antimony	ug/L		1 U		1 U				MS
Arsenic	ug/L		2 U		2 U				MS
Cadmium	ug/L		0.3 U		0.3 U				MS
Chromium	ug/L	+/-10	5.19 J		5.38 J		3.48		MS
Lead	ug/L		0.5 U		0.5 U				MS
Molybdenum	ug/L	+/- .5	1.18		1.2		1.77		MS
Nickel	ug/L	+/-2	1.31 J		1.32 J		.687		MS
Selenium	ug/L		2 U		2 U				MS
Silver	ug/L		0.3 U		0.3 U				MS
Thallium	ug/L		0.6 U		0.6 U				MS
Uranium	ug/L	+/- .2	0.83		0.853		2.73		MS

*Analytical Methods:

MS SW846 3005A/6020A

Metals
–6–
Duplicate Sample Summary

SDG No.: 2017–2284**Lab Code:** GEL**Contract:** ESHL00114**Client ID:** CASA–17–142006D**Matrix:** WATER**Level:** Low**Sample ID:** 429620001**Duplicate ID:** 1203860688**Percent Solids for Dup:** N/A

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M*
Mercury	ug/L		0.067	U	0.067	U			AV

*Analytical Methods:

AV EPA 245.1/245.2

METALS

-7-

Laboratory Control Sample Summary

SDG NO. 2017-2284

Contract: ESHL00114

Aqueous LCS Source:OS2I

Solid LCS Source:

<u>Sample ID</u>	<u>Analyte</u>	<u>Units</u>	<u>True Value</u>	<u>Result</u>	<u>C</u>	<u>% Recovery</u>	<u>Acceptance Limit</u>	<u>M*</u>
1203845801								
	Aluminum	ug/L	5000	5080		102	80-120	P
	Barium	ug/L	500	507		101	80-120	P
	Beryllium	ug/L	500	501		100	80-120	P
	Boron	ug/L	500	517		103	80-120	P
	Calcium	ug/L	5000	5110		102	80-120	P
	Cobalt	ug/L	500	521		104	80-120	P
	Copper	ug/L	500	508		102	80-120	P
	Iron	ug/L	5000	5190		104	80-120	P
	Magnesium	ug/L	5000	5320		106	80-120	P
	Manganese	ug/L	500	504		101	80-120	P
	Potassium	ug/L	5000	5140		103	80-120	P
	Silica	ug/L	10700	10400		97.2	80-120	P
	Sodium	ug/L	5000	5010		100	80-120	P
	Strontium	ug/L	500	495		98.9	80-120	P
	Tin	ug/L	500	508		102	80-120	P
	Vanadium	ug/L	500	506		101	80-120	P
	Zinc	ug/L	500	495		98.9	80-120	P

*Analytical Methods:

P SW846 3005A/6010C

METALS

-7-

Laboratory Control Sample Summary

SDG NO. 2017-2284

Contract: ESHL00114

Aqueous LCS Source:O2Si

Solid LCS Source:

<u>Sample ID</u>	<u>Analyte</u>	<u>Units</u>	<u>True Value</u>	<u>Result</u>	<u>C</u>	<u>% Recovery</u>	<u>Acceptance Limit</u>	<u>M*</u>
1203845844								
	Cadmium	ug/L	50	53.2		106	80-120	MS
	Chromium	ug/L	50	49.9		99.8	80-120	MS
	Lead	ug/L	50	52.2		104	80-120	MS
	Molybdenum	ug/L	50	52.2		104	80-120	MS
	Nickel	ug/L	50	52.5		105	80-120	MS
	Selenium	ug/L	50	51.2		102	80-120	MS
	Silver	ug/L	50	54.3		109	80-120	MS
	Thallium	ug/L	50	48.8		97.5	80-120	MS
	Uranium	ug/L	50	50.8		102	80-120	MS
	Antimony	ug/L	50	50.1		100	80-120	MS
	Arsenic	ug/L	50	52		104	80-120	MS

*Analytical Methods:

MS SW846 3005A/6020A

METALS

-7-

Laboratory Control Sample Summary

SDG NO. 2017-2284

Contract: ESHL00114

Aqueous LCS Source: GEL

Solid LCS Source:

<u>Sample ID</u>	<u>Analyte</u>	<u>Units</u>	<u>True Value</u>	<u>Result</u>	<u>C</u>	<u>% Recovery</u>	<u>Acceptance Limit</u>	<u>M*</u>
1203860687	Mercury	ug/L	2	2.06		103	85-115	AV

*Analytical Methods:

AV EPA 245.1/245.2

METALS

-9-

Serial Dilution Sample Summary

SDG NO. 2017-2284

Client ID: CAMO-17-141977L

Contract: ESHL00114

Matrix: LIQUID

Level: Low

Sample ID: 429570001

Serial Dilution ID: 1203845804

<u>Analyte</u>	<u>Initial Value</u> ug/L	<u>C</u>	<u>Serial Value</u> ug/L	<u>C</u>	<u>% Difference</u>	<u>Qual</u>	<u>Acceptance Limit</u>	<u>M*</u>
Aluminum	68	U	340	U				P
Barium	31.7		31.2		1.647			P
Beryllium	1	U	5	U				P
Boron	15	U	75	U				P
Calcium	11700		11700		.115		10	P
Cobalt	1	U	5	U				P
Copper	3	U	15	U				P
Iron	30	U	150	U				P
Magnesium	3870		4010		3.584			P
Manganese	2	U	10	U				P
Potassium	1440		1460		1.632			P
Silica	73300		74200		1.204		10	P
Sodium	11700		12600		8.164		10	P
Strontium	48.8		46.5		4.694			P
Tin	3.29	J	12.5	U	22.685			P
Vanadium	5.7		7.18	J	26.029			P
Zinc	3.3	U	16.5	U				P

*Analytical Methods:

P SW846 3005A/6010C

METALS

-9-

Serial Dilution Sample Summary

SDG NO. 2017-2284

Client ID: CAMO-17-141977L

Contract: ESHL00114

Matrix: LIQUID

Level: Low

Sample ID: 429570001

Serial Dilution ID: 1203845847

<u>Analyte</u>	<u>Initial Value ug/L</u>	<u>C</u>	<u>Serial Value ug/L</u>	<u>C</u>	<u>% Difference</u>	<u>Qual</u>	<u>Acceptance Limit</u>	<u>M*</u>
Antimony	1	U	5	U				MS
Arsenic	2	U	10	U				MS
Cadmium	.3	U	1.5	U				MS
Chromium	5.19	J	15	U	3.543			MS
Lead	.5	U	2.5	U				MS
Molybdenum	1.18		1.25	J	5.957			MS
Nickel	1.31	J	3	U	2.374			MS
Selenium	2	U	10	U				MS
Silver	.3	U	1.5	U				MS
Thallium	.6	U	3	U				MS
Uranium	.83		.86	J	3.614			MS

*Analytical Methods:

MS SW846 3005A/6020A

METALS

-9-

Serial Dilution Sample Summary

SDG NO. 2017-2284 **Client ID:** CASA-17-142006L**Contract:** ESHL00114**Matrix:** LIQUID **Level:** Low**Sample ID:** 429620001 **Serial Dilution ID:** 1203860690

<u>Analyte</u>	<u>Initial Value ug/L</u>	<u>C</u>	<u>Serial Value ug/L</u>	<u>C</u>	<u>% Difference</u>	<u>Qual</u>	<u>Acceptance Limit</u>	<u>M*</u>
Mercury	.067	U	.335	U				AV

*Analytical Methods:

AV EPA 245.1/245.2

General Chem Analysis

Case Narrative

**General Chemistry
Technical Case Narrative
ARS International, LLC (ARSL)
SDG #: 2017-2284
Work Order #: 429570**

Method/Analysis Information

Product: Carbon and Total Organic

Analytical Batch: 1688478

Method: SW 9060 Total Organic Carbon

Sample Analysis

The following samples were analyzed using the analytical protocol as established in SW-846:9060:

Sample ID	Client ID
429570002	CAMO-17-141993
429570004	CAMO-17-141994
1203846820	Method Blank (MB)
1203846821	Laboratory Control Sample (LCS)
1203846822	429324003(CAMO-17-142312) Sample Duplicate (DUP)
1203846823	429607002(NonSDG) Sample Duplicate (DUP)
1203846824	429324003(CAMO-17-142312) Post Spike (PS)
1203846825	429607002(NonSDG) Post Spike (PS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-093 REV# 15.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Carbon analysis was performed on a O-I Analytical 1030W Carbon Analyzer.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Quality Control (QC) Information**Method Blank (MB) Statement**

The MBs analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recoveries met the acceptance limits.

Quality Control (QC) Designation

Samples 429324003 (CAMO-17-142312) and 429607002 (NonSDG) were selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits where applicable.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The following samples 1203846823 (Non SDG 429607002DUP) and 1203846825 (Non SDG 429607002PS) in this sample group were diluted due to matrix interference. Dilutions may be required for many reasons, including to minimize matrix interferences or to bring over range target analyte concentrations into the linear calibration range.

Sample Re-analysis

Samples were re-analyzed due to CCV failure. The reanalysis data with passing instrument QC was reported.

Miscellaneous Information**Additional Comments**

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an

effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product:	Cyanide and Total		
Analytical Batch:	1688106	Method:	WSP-CN(T)
Prep Batch :	1688105	Method:	EPA 335.4

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 335.4 1993:

Sample ID	Client ID
429570002	CAMO-17-141993
429570004	CAMO-17-141994
1203845289	Method Blank (MB)
1203845290	Laboratory Control Sample (LCS)
1203845292	429092002(CASA-17-142016) Sample Duplicate (DUP)
1203845295	429092002(CASA-17-142016) Matrix Spike (MS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-095 REV# 20.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Flow Injection analysis was performed on a Lachat QuickChem FIA+ 8000 Series.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Sample 429092002 (CASA-17-142016) was selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recovery for this sample set was within the required acceptance limits where applicable.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information**Additional Comments**

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Ion Chromatography

Analytical Batch: 1691711

Method: WSP-ANIONS

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA:300.0:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203854051	Method Blank (MB)
1203854052	Laboratory Control Sample (LCS)
1203854053	429458003(CASA-17-142013) Sample Duplicate (DUP)
1203854054	429458003(CASA-17-142013) Post Spike (PS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-086 REV# 25.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Ion Chromatography analysis was performed on a Dionex ICS-3000 Ion Chromatograph.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Sample 429458003 (CASA-17-142013) was selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits where applicable.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Dilutions

The following samples 1203854053 (CASA-17-142013DUP) and 1203854054 (CASA-17-142013PS) were diluted because target analyte concentrations exceeded the calibration range. Dilutions may be required for many reasons, including to minimize matrix interferences or to bring over range target analyte concentrations into the linear calibration range.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Manual Integrations

Samples 1203854053 (CASA-17-142013DUP), 1203854054 (CASA-17-142013PS), 429570001 (CAMO-17-141977) and 429570003 (CAMO-17-141978) were manually integrated to correctly position the baseline as set in the calibration standards.

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages

electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product:	Ammonia Nitrogen		
Analytical Batch:	1692774	Method:	NH3
Prep Batch :	1692772	Method:	EPA 350.1 Prep

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA:350.1:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203856520	Method Blank (MB)
1203856521	Laboratory Control Sample (LCS)
1203856522	429712005(NonSDG) Sample Duplicate (DUP)
1203856523	429317001(CAMO-17-142070) Sample Duplicate (DUP)
1203856524	429712005(NonSDG) Matrix Spike (MS)
1203856525	429317001(CAMO-17-142070) Matrix Spike (MS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-106 REV# 9.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8000 Series.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within

acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information**Method Blank (MB) Statement**

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Samples 429317001 (CAMO-17-142070) and 429712005 (NonSDG) were selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits where applicable.

Duplicate Relative Percent Difference (RPD) Statement

The Relative Percent Difference (RPD) between the sample and duplicate falls outside of the established acceptance limits because of the heterogeneous matrix of the sample:

Analyte	Sample	Value
Nitrogen, Ammonia	1203856523 (CAMO-17-142070DUP)	abs(.0818 - .0235)* (+/- .05 mg/L)

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product:	Total Kjeldahl Nitrogen		
Analytical Batch:	1692759	Method:	TKN
Prep Batch :	1692758	Method:	EPA 351.2 Prep

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA:351.2:

Sample ID	Client ID
429570002	CAMO-17-141993
429570004	CAMO-17-141994
1203856478	Method Blank (MB)
1203856479	Laboratory Control Sample (LCS)
1203856480	429266002(CASA-17-142037) Sample Duplicate (DUP)
1203856481	429266002(CASA-17-142037) Matrix Spike (MS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-104 REV# 14.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8000 Series.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Calibration Verification Information

All associated calibration verification standard(s) (ICV or CCV) met the acceptance criteria.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Sample 429266002 (CASA-17-142037) was selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The percent recoveries (%R) obtained from the spike analyses are evaluated when the sample concentration is less than four times (4X) the spike concentration added. The matrix spike recovered outside of the established acceptance limits due to matrix interference and/or non-homogeneity. 1203856481 (CASA-17-142037MS).

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

Samples 1203856478 (MB), 1203856479 (LCS), 1203856480 (CASA-17-142037DUP) and 1203856481 (CASA-17-142037MS) were re-analyzed due to CCV failure. The reanalysis data with passing instrument QC was reported.

Miscellaneous Information

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Nitrate Nitrite by Cadmium Reduction

Analytical Batch: 1689327

Method: NO3NO2

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA:353.2:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203848234	Method Blank (MB)
1203848235	Laboratory Control Sample (LCS)
1203848236	429449001(CAMO-17-142236) Sample Duplicate (DUP)
1203848237	429570001(CAMO-17-141977) Sample Duplicate (DUP)
1203848238	429449001(CAMO-17-142236) Post Spike (PS)
1203848239	429570001(CAMO-17-141977) Post Spike (PS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-128 REV# 9.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8500 Series.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Calibration Verification Information

All associated calibration verification standard(s) (ICV or CCV) met the acceptance criteria.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Samples 429449001 (CAMO-17-142236) and 429570001 (CAMO-17-141977) were selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits where applicable.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an

effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product:	Total Phosphorus		
Analytical Batch:	1692781	Method:	PO4
Prep Batch :	1692780	Method:	EPA 365.4 Prep

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 365.4 1974:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203856557	Method Blank (MB)
1203856558	Laboratory Control Sample (LCS)
1203856559	429324001(CAMO-17-141979) Sample Duplicate (DUP)
1203856560	429324005(CAMO-17-141985) Sample Duplicate (DUP)
1203856561	429324001(CAMO-17-141979) Matrix Spike (MS)
1203856562	429324005(CAMO-17-141985) Matrix Spike (MS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-103 REV# 10.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8000 Series.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within

acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Samples 429324001 (CAMO-17-141979) and 429324005 (CAMO-17-141985) were selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits where applicable.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages

electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Solids and Total Dissolved

Analytical Batch: 1688433

Method: TDS

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA:160.1:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203846055	Method Blank (MB)
1203846056	Laboratory Control Sample (LCS)
1203846057	429324004(CAMO-17-142780) Sample Duplicate (DUP)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-001 REV# 15.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Solids analysis was performed on a Sartorius Balance BAL216. Solids lab

Initial Calibration

All initial calibration requirements have been met for this SDG.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Consecutive Weight Checks

All consecutive weight checks were met.

Quality Control (QC) Designation

Sample 429324004 (CAMO-17-142780) was selected for QC analysis.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information**Additional Comments**

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Specific Conductivity

Analytical Batch: 1688310

Method: EPA120.1 Specific Conductivity

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA:120.1:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203845787	Laboratory Control Sample (LCS)
1203845788	429266001(CASA-17-142030) Sample Duplicate (DUP)
1203849533	429570001(CAMO-17-141977) Sample Duplicate (DUP)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-009 REV# 14.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Titration and Ion analysis was performed on a Orion 160 Conductivity Meter.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Initial Standardization

The titrant was properly standardized

Quality Control (QC) Information

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Samples 429266001 (CASA-17-142030) and 429570001 (CAMO-17-141977) were selected for QC analysis.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information**Additional Comments**

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: pH
Analytical Batch: 1690702 **Method:** PH

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 150.1 1982:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203851537	Laboratory Control Sample (LCS)
1203851538	429620001(CASA-17-142006) Sample Duplicate (DUP)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-008 REV# 22.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Titration and Ion analysis was performed on a Thermo Orion Star A111. Immediates

Initial Calibration

All initial calibration requirements have been met for this SDG.

Initial Standardization

The titrant was properly standardized

Quality Control (QC) Information

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Sample 429620001 (CASA-17-142006) was selected for QC analysis.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

Samples (See Below) were received by the laboratory outside of the method specified holding time. The data is qualified.

Sample	Analyte	Value
1203851538 (CASA-17-142006DUP)	pH	Received 03-AUG-17, out of holding 01-AUG-17
429570001 (CAMO-17-141977)	pH	Received 03-AUG-17, out of holding 01-AUG-17
429570003 (CAMO-17-141978)	pH	Received 03-AUG-17, out of holding 01-AUG-17

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Alkalinity

Analytical Batch: 1690697 **Method:** EPA 310.1 Total Alkalinity

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA:310.1:

Sample ID	Client ID
429570001	CAMO-17-141977
429570003	CAMO-17-141978
1203851522	Laboratory Control Sample (LCS)
1203851524	429570001(CAMO-17-141977) Sample Duplicate (DUP)
1203851528	429570001(CAMO-17-141977) Matrix Spike (MS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-033 REV# 13.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Titration and Ion analysis was performed on a Electronic bottle-top buret.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Initial Standardization

The titrant was properly standardized

Quality Control (QC) Information

Laboratory Control Sample Duplicate (LCSD)

An LCSD was not used in place of matrix QC.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

Sample 429570001 (CAMO-17-141977) was selected for QC analysis.

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recovery for this sample set was within the required acceptance limits where applicable.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information**Additional Comments**

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report for

ARSL004 ARS International, LLC (ARS-LANS-MTOA6-25093-GEL)

Client SDG: 2017-2284 GEL Work Order: 429570


The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- H Analytical holding time was exceeded
- J Value is estimated
- U Analyte was analyzed for, but not detected above the MDL, MDA, MDC or LOD.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature: 

Name: Kristen Mizzell

Date: 25 AUG 2017

Title: Analyst I

Sample Data Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: August 25, 2017

Company : Los Alamos National Laboratory
Address : TA-00, SM1237, Rm104C

Los Alamos, New Mexico 87545
Contact: Ms. Nita Patel
Project: LANL- WQH Water Samples

Client SDG: 2017-2284

Client Sample ID: CAMO-17-141977
Sample ID: 429570001
Matrix: W
Collect Date: 01-AUG-17 12:33
Receive Date: 03-AUG-17
Collector: Client

Project: ESHL00114
Client ID: ARSL004

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Ion Chromatography												
WSP-ANIONS "As Received"												
Bromide	U	ND	0.067	0.200	mg/L		1	MXL2	08/15/17	0616	1691711	1
Chloride		2.23	0.067	0.200	mg/L		1					
Fluoride		0.140	0.033	0.100	mg/L		1					
Sulfate		3.44	0.133	0.400	mg/L		1					
Nutrient Analysis												
NH3 "As Received"												
Nitrogen, Ammonia	U	ND	0.017	0.050	mg/L	1.00	1	KLP1	08/22/17	1218	1692774	2
NO3NO2 "As Received"												
Nitrogen, Nitrate/Nitrite		0.544	0.017	0.050	mg/L		1	AXH3	08/10/17	0656	1689327	3
PO4 "As Received"												
Phosphorus, Total as P	U	ND	0.020	0.050	mg/L	1.00	1	KLP1	08/23/17	1108	1692781	4
Solids Analysis												
TDS "As Received"												
Total Dissolved Solids		157	3.40	14.3	mg/L			KLP1	08/04/17	1513	1688433	5
Titration and Ion Analysis												
EPA 310.1 Total Alkalinity "As Received"												
Alkalinity, Total as CaCO3		66.4	1.45	4.00	mg/L			RXB5	08/11/17	1628	1690697	6
Carbonate alkalinity (CaCO3)	U	ND	1.45	4.00	mg/L							
EPA120.1 Specific Conductivity "As Received"												
Conductivity		156	1.00	1.00	umhos/cm		1	VH1	08/08/17	1508	1688310	7
PH "As Received"												
pH at Temp 12.0C	H	7.84	0.010	0.100	SU		1	RXB5	08/11/17	1631	1690702	8

The following Prep Methods were performed:

Method	Description	Analyst	Date	Time	Prep Batch
EPA 350.1 Prep	EPA 350.1 Ammonia Nitrogen Prep	AXH3	08/22/17	0952	1692772
EPA 365.4 Prep	EPA 365.4 Phosphorus, Total in liquid PR	KLP1	08/22/17	1700	1692780

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: August 25, 2017

Company : Los Alamos National Laboratory
Address : TA-00, SM1237, Rm104C

Los Alamos, New Mexico 87545
Contact: Ms. Nita Patel
Project: LANL- WQH Water Samples

Client SDG: 2017-2284

Client Sample ID: CAMO-17-141977
Sample ID: 429570001

Project: ESHL00114
Client ID: ARSL004

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
The following Analytical Methods were performed:												
Method	Description		Analyst Comments									
1	EPA:300.0											
2	EPA:350.1											
3	EPA:353.2											
4	EPA 365.4 1974											
5	EPA:160.1											
6	EPA:310.1											
7	EPA:120.1											
8	EPA 150.1 1982											

Notes:

Column headers are defined as follows:

DF: Dilution Factor	Lc/LC: Critical Level
DL: Detection Limit	PF: Prep Factor
MDA: Minimum Detectable Activity	RL: Reporting Limit
MDC: Minimum Detectable Concentration	SQL: Sample Quantitation Limit

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: August 25, 2017

Company : Los Alamos National Laboratory
Address : TA-00, SM1237, Rm104C

Los Alamos, New Mexico 87545
Contact: Ms. Nita Patel
Project: LANL- WQH Water Samples

Client SDG: 2017-2284

Client Sample ID: CAMO-17-141993
Sample ID: 429570002
Matrix: W
Collect Date: 01-AUG-17 12:33
Receive Date: 03-AUG-17
Collector: Client

Project: ESHL00114
Client ID: ARSL004

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Carbon Analysis												
SW 9060 Total Organic Carbon "As Received"												
Total Organic Carbon Average	U	ND	0.330	1.00	mg/L		1	TSM	08/16/17	0330	1688478	1
Flow Injection Analysis												
WSP-CN(T) "As Received"												
Cyanide, Total	U	ND	1.67	5.00	ug/L	1.00	1	AXH3	08/04/17	1003	1688106	2
Nutrient Analysis												
TKN "As Received"												
Nitrogen, Total Kjeldahl	U	ND	0.033	0.100	mg/L	1.00	1	KLP1	08/24/17	1535	1692759	3

The following Prep Methods were performed:

Method	Description	Analyst	Date	Time	Prep Batch
EPA 335.4	EPA 335.4 Total Cyanide	AXH3	08/04/17	0850	1688105
EPA 351.2 Prep	EPA 351.2 Total Kjeldahl Nitrogen Prep	KLP1	08/23/17	1700	1692758

The following Analytical Methods were performed:

Method	Description	Analyst Comments
1	SW-846:9060	
2	EPA 335.4 1993	
3	EPA:351.2	

Notes:

Column headers are defined as follows:

DF: Dilution Factor	Lc/LC: Critical Level
DL: Detection Limit	PF: Prep Factor
MDA: Minimum Detectable Activity	RL: Reporting Limit
MDC: Minimum Detectable Concentration	SQL: Sample Quantitation Limit

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: August 25, 2017

Company : Los Alamos National Laboratory
Address : TA-00, SM1237, Rm104C

Los Alamos, New Mexico 87545
Contact: Ms. Nita Patel
Project: LANL- WQH Water Samples

Client SDG: 2017-2284

Client Sample ID: CAMO-17-141978
Sample ID: 429570003
Matrix: W
Collect Date: 01-AUG-17 10:34
Receive Date: 03-AUG-17
Collector: Client

Project: ESHL00114
Client ID: ARSL004

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Ion Chromatography												
WSP-ANIONS "As Received"												
Bromide	U	ND	0.067	0.200	mg/L		1	MXL2	08/15/17	0645	1691711	1
Chloride		1.93	0.067	0.200	mg/L		1					
Fluoride		0.128	0.033	0.100	mg/L		1					
Sulfate		2.32	0.133	0.400	mg/L		1					
Nutrient Analysis												
NH3 "As Received"												
Nitrogen, Ammonia		0.0533	0.017	0.050	mg/L	1.00	1	KLP1	08/22/17	1219	1692774	2
NO3NO2 "As Received"												
Nitrogen, Nitrate/Nitrite		0.350	0.017	0.050	mg/L		1	AXH3	08/10/17	0700	1689327	3
PO4 "As Received"												
Phosphorus, Total as P	U	ND	0.020	0.050	mg/L	1.00	1	KLP1	08/23/17	1109	1692781	4
Solids Analysis												
TDS "As Received"												
Total Dissolved Solids		136	3.40	14.3	mg/L			KLP1	08/04/17	1513	1688433	5
Titration and Ion Analysis												
EPA 310.1 Total Alkalinity "As Received"												
Alkalinity, Total as CaCO3		69.8	1.45	4.00	mg/L			RXB5	08/11/17	1635	1690697	6
Carbonate alkalinity (CaCO3)	U	ND	1.45	4.00	mg/L							
EPA120.1 Specific Conductivity "As Received"												
Conductivity		152	1.00	1.00	umhos/cm		1	VH1	08/08/17	1508	1688310	7
PH "As Received"												
pH at Temp 13.1C	H	7.93	0.010	0.100	SU		1	RXB5	08/11/17	1634	1690702	8

The following Prep Methods were performed:

Method	Description	Analyst	Date	Time	Prep Batch
EPA 350.1 Prep	EPA 350.1 Ammonia Nitrogen Prep	AXH3	08/22/17	0952	1692772
EPA 365.4 Prep	EPA 365.4 Phosphorus, Total in liquid PR	KLP1	08/22/17	1700	1692780

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: August 25, 2017

Company : Los Alamos National Laboratory
Address : TA-00, SM1237, Rm104C

Los Alamos, New Mexico 87545
Contact: Ms. Nita Patel
Project: LANL- WQH Water Samples

Client SDG: 2017-2284

Client Sample ID: CAMO-17-141978
Sample ID: 429570003

Project: ESHL00114
Client ID: ARSL004

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
The following Analytical Methods were performed:												
Method	Description					Analyst Comments						
1	EPA:300.0											
2	EPA:350.1											
3	EPA:353.2											
4	EPA 365.4 1974											
5	EPA:160.1											
6	EPA:310.1											
7	EPA:120.1											
8	EPA 150.1 1982											

Notes:

Column headers are defined as follows:

DF: Dilution Factor	Lc/LC: Critical Level
DL: Detection Limit	PF: Prep Factor
MDA: Minimum Detectable Activity	RL: Reporting Limit
MDC: Minimum Detectable Concentration	SQL: Sample Quantitation Limit

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: August 25, 2017

Company : Los Alamos National Laboratory
Address : TA-00, SM1237, Rm104C

Los Alamos, New Mexico 87545

Contact: Ms. Nita Patel

Client SDG: 2017-2284

Project: LANL- WQH Water Samples

Client Sample ID: CAMO-17-141994

Project: ESHL00114

Sample ID: 429570004

Client ID: ARSL004

Matrix: W

Collect Date: 01-AUG-17 10:34

Receive Date: 03-AUG-17

Collector: Client

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Carbon Analysis												
SW 9060 Total Organic Carbon "As Received"												
Total Organic Carbon Average	U	ND	0.330	1.00	mg/L		1	TSM	08/16/17	0416	1688478	1
Flow Injection Analysis												
WSP-CN(T) "As Received"												
Cyanide, Total	U	ND	1.67	5.00	ug/L	1.00	1	AXH3	08/04/17	1004	1688106	2
Nutrient Analysis												
TKN "As Received"												
Nitrogen, Total Kjeldahl	U	ND	0.033	0.100	mg/L	1.00	1	KLP1	08/24/17	1536	1692759	3

The following Prep Methods were performed:

Method	Description	Analyst	Date	Time	Prep Batch
EPA 335.4	EPA 335.4 Total Cyanide	AXH3	08/04/17	0850	1688105
EPA 351.2 Prep	EPA 351.2 Total Kjeldahl Nitrogen Prep	KLP1	08/23/17	1700	1692758

The following Analytical Methods were performed:

Method	Description	Analyst Comments
1	SW-846:9060	
2	EPA 335.4 1993	
3	EPA:351.2	

Notes:

Column headers are defined as follows:

DF: Dilution Factor

DL: Detection Limit

MDA: Minimum Detectable Activity

MDC: Minimum Detectable Concentration

Lc/LC: Critical Level

PF: Prep Factor

RL: Reporting Limit

SQL: Sample Quantitation Limit

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: August 25, 2017

Page 1 of 7

Los Alamos National Laboratory
TA-00, SM1237, Rm104C
Los Alamos, New Mexico

Contact: Ms. Nita Patel

Workorder: 429570

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Carbon Analysis											
Batch	1688478										
QC1203846822	429324003	DUP									
Total Organic Carbon Average		U	ND	U	ND	mg/L	N/A		TSM	08/15/17	20:51
QC1203846823	429607002	DUP									
Total Organic Carbon Average			17.2		17.1	mg/L	0.175 ^	(+/-10.0)		08/16/17	06:37
QC1203846821	LCS										
Total Organic Carbon Average	10.0				9.54	mg/L		95.4 (80%-120%)		08/15/17	19:52
QC1203846820	MB										
Total Organic Carbon Average			U		ND	mg/L				08/15/17	19:41
QC1203846824	429324003	PS									
Total Organic Carbon Average	10.0	U	ND		10.3	mg/L		100 (75%-125%)		08/15/17	21:38
QC1203846825	429607002	PS									
Total Organic Carbon Average	10.0		1.72		11.8	mg/L		100 (75%-125%)		08/16/17	07:24
Flow Injection Analysis											
Batch	1688106										
QC1203845292	429092002	DUP									
Cyanide, Total		U	ND	U	ND	ug/L	N/A		AXH3	08/04/17	09:44
QC1203845290	LCS										
Cyanide, Total	50.0				51.8	ug/L		104 (90%-110%)		08/04/17	09:31
QC1203845289	MB										
Cyanide, Total			U		ND	ug/L				08/04/17	09:30
QC1203845295	429092002	MS									
Cyanide, Total	100	U	ND		107	ug/L		107 (90%-110%)		08/04/17	09:45

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 429570

Page 2 of 7

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Ion Chromatography											
Batch	1691711										
QC1203854053	429458003	DUP									
Bromide		0.664		0.662	mg/L	0.317	^	(+/-0.200)	MXL2	08/15/17	05:17
Chloride		69.7		69.7	mg/L	0.023		(0%-20%)		08/15/17	15:27
Fluoride		0.107		0.104	mg/L	3.32	^	(+/-0.100)		08/15/17	05:17
Sulfate		86.7		86.7	mg/L	0.068		(0%-20%)		08/15/17	15:27
QC1203854052	LCS										
Bromide	1.25			1.27	mg/L			102	(80%-120%)	08/15/17	04:18
Chloride	5.00			4.65	mg/L			93	(80%-120%)		
Fluoride	2.50			2.38	mg/L			95.2	(80%-120%)		
Sulfate	10.0			9.61	mg/L			96.1	(80%-120%)		
QC1203854051	MB										
Bromide		U		ND	mg/L					08/15/17	03:49
Chloride		U		ND	mg/L						
Fluoride		U		ND	mg/L						
Sulfate		J		0.175	mg/L						
QC1203854054	429458003	PS									
Bromide	1.25	0.664		1.87	mg/L			96.1	(75%-125%)	08/15/17	05:46
Chloride	5.00	6.97		12.5	mg/L			110	(75%-125%)	08/15/17	15:57

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 429570

Page 3 of 7

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Ion Chromatography											
Batch	1691711										
Fluoride	2.50	0.107		2.35	mg/L		89.6	(75%-125%)	MXL2	08/15/17	05:46
Sulfate	10.0	8.67		19.2	mg/L		105	(75%-125%)		08/15/17	15:57
Nutrient Analysis											
Batch	1689327										
QC1203848236	429449001	DUP									
Nitrogen, Nitrate/Nitrite		0.693		0.694	mg/L	0.144		(0%-20%)	AXH3	08/10/17	06:54
QC1203848237	429570001	DUP									
Nitrogen, Nitrate/Nitrite		0.544		0.540	mg/L	0.738		(0%-20%)		08/10/17	06:58
QC1203848235	LCS										
Nitrogen, Nitrate/Nitrite	1.00			0.999	mg/L		99.9	(90%-110%)		08/10/17	06:52
QC1203848234	MB										
Nitrogen, Nitrate/Nitrite			U	ND	mg/L					08/10/17	06:46
QC1203848238	429449001	PS									
Nitrogen, Nitrate/Nitrite	1.00	0.693		1.69	mg/L		99.7	(90%-110%)		08/10/17	06:55
QC1203848239	429570001	PS									
Nitrogen, Nitrate/Nitrite	1.00	0.544		1.55	mg/L		101	(90%-110%)		08/10/17	06:59
Batch	1692759										
QC1203856480	429266002	DUP									
Nitrogen, Total Kjeldahl		U	ND	U	ND	mg/L	N/A		KLP1	08/24/17	15:22
QC1203856479	LCS										
Nitrogen, Total Kjeldahl	1.00			1.02	mg/L		102	(90%-110%)		08/24/17	15:21
QC1203856478	MB										
Nitrogen, Total Kjeldahl			U	ND	mg/L					08/24/17	15:20

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 429570

Page 4 of 7

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Nutrient Analysis											
Batch	1692759										
QC1203856481	429266002	MS									
Nitrogen, Total Kjeldahl	1.00	U	ND	0.632	mg/L		63.2 *	(90%-110%)	KLP1	08/24/17	15:23
Batch	1692774										
QC1203856522	429712005	DUP									
Nitrogen, Ammonia			1.45	1.35	mg/L	7.14		(0%-20%)	KLP1	08/22/17	12:23
QC1203856523	429317001	DUP									
Nitrogen, Ammonia		J	0.0235	0.0818	mg/L	111 * ^		(+/-0.050)		08/22/17	12:08
QC1203856521	LCS										
Nitrogen, Ammonia	1.00			1.00	mg/L		100	(90%-110%)		08/22/17	12:07
QC1203856520	MB										
Nitrogen, Ammonia			U	ND	mg/L					08/22/17	12:06
QC1203856524	429712005	MS									
Nitrogen, Ammonia	1.00		1.45	2.47	mg/L		102	(90%-110%)		08/22/17	12:28
QC1203856525	429317001	MS									
Nitrogen, Ammonia	1.00	J	0.0235	1.02	mg/L		99.7	(90%-110%)		08/22/17	12:09
Batch	1692781										
QC1203856559	429324001	DUP									
Phosphorus, Total as P		U	ND	U	ND	mg/L	N/A		KLP1	08/23/17	10:57
QC1203856560	429324005	DUP									
Phosphorus, Total as P		U	ND	U	ND	mg/L	N/A			08/23/17	11:04
QC1203856558	LCS										
Phosphorus, Total as P	1.00			1.07	mg/L		107	(80%-124%)		08/23/17	10:55
QC1203856557	MB										
Phosphorus, Total as P			U	ND	mg/L					08/23/17	10:54

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 429570

Page 5 of 7

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Nutrient Analysis											
Batch	1692781										
QC1203856561	429324001	MS									
Phosphorus, Total as P	1.00	U	ND	1.07	mg/L		107	(63%-139%)	KLP1	08/23/17	10:58
QC1203856562	429324005	MS									
Phosphorus, Total as P	1.00	U	ND	1.06	mg/L		104	(63%-139%)		08/23/17	11:05
Solids Analysis											
Batch	1688433										
QC1203846057	429324004	DUP									
Total Dissolved Solids			177	181	mg/L	1.59		(0%-5%)	KLP1	08/04/17	15:13
QC1203846056	LCS										
Total Dissolved Solids	300			294	mg/L		98.1	(95%-105%)		08/04/17	15:13
QC1203846055	MB										
Total Dissolved Solids			U	ND	mg/L					08/04/17	15:13
Titration and Ion Analysis											
Batch	1688310										
QC1203845788	429266001	DUP									
Conductivity			193	198	umhos/cm	2.25		(0%-10%)	VH1	08/08/17	15:08
QC1203849533	429570001	DUP									
Conductivity			156	156	umhos/cm	0.128		(0%-10%)		08/08/17	15:08
QC1203845787	LCS										
Conductivity	1410			1450	umhos/cm		103	(95%-105%)		08/08/17	15:08
Batch	1690697										
QC1203851524	429570001	DUP									
Alkalinity, Total as CaCO3			66.4	66.6	mg/L	0.303		(0%-20%)	RXB5	08/11/17	16:30
Carbonate alkalinity (CaCO3)		U	ND	U	ND	mg/L	N/A				

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 429570

Page 6 of 7

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Titration and Ion Analysis											
Batch	1690697										
QC1203851522 LCS											
Alkalinity, Total as CaCO3	100			110	mg/L		110	(90%-110%)	RXB5	08/11/17	15:58
QC1203851528 429570001 MS											
Alkalinity, Total as CaCO3	100	66.4		171	mg/L		105	(80%-120%)		08/11/17	16:32
Batch	1690702										
QC1203851538 429620001 DUP											
pH		H	8.38 H	8.40	SU	0.238		(0%-5%)	RXB5	08/11/17	16:41
QC1203851537 LCS											
pH	7.00			6.99	SU		99.9	(99%-101%)		08/11/17	15:58

Notes:

- < Result is less than value reported
- > Result is greater than value reported
- B The target analyte was detected in the associated blank.
- E General Chemistry--Concentration of the target analyte exceeds the instrument calibration range
- H Analytical holding time was exceeded
- J Value is estimated
- N/A RPD or %Recovery limits do not apply.
- N1 See case narrative
- ND Analyte concentration is not detected above the detection limit
- NJ Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Q One or more quality control criteria have not been met. Refer to the applicable narrative or DER.
- R Per section 9.3.4.1 of Method 1664 Revision B, due to matrix spike recovery issues, this result may not be reported or used for regulatory compliance purposes.
- R Sample results are rejected
- U Analyte was analyzed for, but not detected above the MDL, MDA, MDC or LOD.
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Z Paint Filter Test--Particulates passed through the filter, however no free liquids were observed.
- ^ RPD of sample and duplicate evaluated using +/-RL. Concentrations are <5X the RL. Qualifier Not Applicable for Radiochemistry.
- d 5-day BOD--The 2:1 depletion requirement was not met for this sample
- e 5-day BOD--Test replicates show more than 30% difference between high and low values. The data is qualified per the method and can be used for reporting purposes

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 429570

Page 7 of 7

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
h	Preparation or preservation holding time was exceeded										

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

* Indicates that a Quality Control parameter was not within specifications.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.