

International Hydrogen Infrastructure Workshop

Validation of Hydrogen Meter Testing Device in
Europe

THIS DOCUMENT IS PUBLIC

Boston • 11.09.2018

Joachim Schütte • Air Liquide Advanced Technologies



Contents

1. Background Hydrogen Flow Metering
2. FCH-JU Study on Metering
3. Metering Testing Device
4. Test Campaign
5. Summary

1. Background Hydrogen Flow Metering

- Flow meters are not approved according to OIML R139 due to the absence of testing facilities in Europe (H₂, 700 bar,...)
- Existing OIML R139-2014 is not adapted for hydrogen dispensers
 - Accuracy, MMQ, Durability test, Tests at constant flow rate
→ *Currently in revision (expected for beginning of 2019)*
- Until beginning of 2018 no certified reference testing device in Europe to determine the global accuracy → *of meters and dispensers*
- Up to now the sale of H₂ without certified flow meter is *tolerated* by the authorities (demonstration projects, limited group of users)
- By entering the commercial phase with the extension of the HRS network the uncalibrated sale of H₂ cannot be tolerated anymore
- Pressure of the Offices of Weights and Measures (Eichämter) is currently increasing in Germany
→ **Therefore, short-term solution for the approval H₂ dispensers is necessary for the further ramp-up of the HRS network**

2. FCH-JU Study on Metering

FCH-JU Study “Development of a Metering Protocol for Hydrogen Refuelling Stations” (Contract No: FCH / OP / 196)

- **Objective:** Define an accelerated test protocol to quickly certify HRS without a certified meter
- **Scope:**
 - **Task 1:** Development of a testing protocol for HRS regarding compliance with OIML R-139
 - **Task 2:** Design and implementation of a test campaign
 - **Task 3:** Agreement from relevant national authorities/institutes
- **Organization:**
 - **Leader:** Air Liquide
 - **National Metrological Institutes:**
 - PTB (*Physikalisch-Technische Bundesanstalt*) in Germany
 - LNE (*Laboratoire national de métrologie et d'essais*) in France
 - NMI Certin in the Netherlands
 - **Laboratory expert in gaseous flow metering:** Cesame Exadebit

2. FCH-JU Study on Metering

Approach for the testing protocol:

- **OIML R139:2014** is the reference
- Main components of the measuring system (calculator and meter) are approved according to different standards:
 - OIML R139:2007
 - OIML R117-1:2007 or OIML R117-2:2014: “*Dynamic measuring systems for liquids other than water*”
 - OIML R137:2012: “*Gas meters*” (for meters only, not applicable to calculator)
- Assessment of deviations to OIML R139:2014 for these components for each category of tests required for Type Approval:
 - Electromagnetic compatibility (EMC)
 - Environment testing (climatic test, humidity, etc.)
 - Accuracy tests
 - Gas temperature accuracy tests
 - Durability tests
 - Software (WELMEC 7.2)

→ **Decision is made to require, or not, complete new tests or additional tests depending of the certification of the components**

2. FCH-JU Study on Metering

		Calculating & indicating device		Measurement transducer (electronics) & Measurement sensor	
		Certified according to: OIML R117-1:2007 or OIML R139:2007	Certified according to: OIML R117-2:2014 or OIML R139:2014	Certified according to: OIML R117-1:2007 or OIML R139:2007 or OIML R137:2012	Certified according to: OIML R117-2:2014 or OIML R139:2014
Type approval tests	EMC	2 (3)	3	2 (3)	3
	Environment testing (climatic test, humidity)	3		3	
	Mechanical test (vibration)	3 if M1		3 if M1	
	Accuracy test	3		1	
	Accuracy gas temperature tests	3		4	
	Software (WELMEC 7.2)	4	3	4	3
	Durability test	3		4 (3)	
Initial verification	Adjustment on site	1		1	

Legend:

1 = Complete new tests

2 = Additional test required

3 = No test required

4 = No test required, but under conditions

() = specific for existing HRS

3. Metering Testing Device

Characteristics:

- Fulfills metrological requirements acc. to OIML R139 → Uncertainty $U < \frac{1}{5}$ MPE = 0,3 %
- High Precision Scale: resolution 0.2 g, ATEX certified
- Composite tank of 104L (i.e.: 4.0 kg of hydrogen at 700 bar)
- Trailer walls, doors and roof serve as protection against wind forces
- Independant vent stack for depressurization of the tank
- Test Bench is now CE-marked



3. Metering Testing Device

- Certified as first **reference standard** for calibration, conformity assessment and verification of hydrogen refuelling dispensers by PTB in March 2018
- Test Bench is also accepted by LNE and NMI as reference standard



Bericht
Report

METERING TEST BENCH
for hydrogen refuelling station
in accordance to SAE J2601

Messeinrichtung
für Wasserstoffbetankungsanlagen
nach SAE J2601

der/of

AIR LIQUIDE
Advanced Technologies
2, rue de Clémenceire
F – 38360 Sassenage, France

Usage for the conformity assessment and verification
for legal metrology purposes
Einsatz für Konformitätsbewertungen und Eichungen
im Rahmen des gesetzlichen Messwesens

Anzahl der Seiten: 22
Number of pages:

Geschäftszeichen: 1.42-4087238
Reference No.:

Im Auftrag
On behalf of PTB


Dr. Roland Schmidt

Braunschweig, 2018-03-28

Siegel
Seal



Im Auftrag
On behalf of PTB


Dr. Rainer Kramer

4. Test Campaign

Criteria for HRS testing

- All technologies and/or specificities should be tested
- HRS from different manufacturers in Europe: Air Liquide, Linde and Nel
- HRS in operation in minimum 3 different countries of the European Union

Selected HRS

Air Liquide: compressed gas

Flow Meter located in the station, which can be far away from the dispenser

- HRS Kamen, Germany → short distance between Flow Meter and dispenser (6 m)
- HRS Koblenz, Germany → long distance between Flow Meter and dispenser (25 m)
- HRS Versailles, France
- HRS Rhoon (Rotterdam), Netherlands

Linde: liquid hydrogen & compressed gas

- HRS Hannover → liquid
- HRS Cologne → compressed gas

Nel: compressed gas → HRS Rostock Germany



Kamen - Air Liquide



Cologne - Linde



Rostock - Nel

4. Test Campaign

Test Program for Accuracy Tests:

- 3 x Full Filling: 20-700 bar
- 3 x Partial Filling: 20-350 bar
- 3 x Partial Filling: 350-700 bar
- 3 x MMQ (1kg)

Recording further data:

- Environmental conditions :
Ambiant temperature, wind velocity,
- Filling conditions: pressure ramp,
temperature, flow rate)
- Start/final pressure and mass

Duration for one Test Campaign:

- At least 3 days including installation due to long time for depressurization (1.5h for a full tank) are needed



4. Test Campaign

First Test Campaign in Kamen



5. Summary

Summary

- Good test results with a high repeatability
- Test Bench and test results are highly appreciated by PTB and the Offices of Weights and Measures
- Test results are good baseline for further discussions with the German Offices of Weights and Measures

Next steps

- Developed test protocol has to be endorsed by further NMIs
- Finalize the type approval approval process for new HRS and start the on-site verification
- Modify existing HRS to be fully compliant with metrological requirements and start the conformity assessment

→ **In the next years more than 100 HRS have to be certified!!!**



The image shows a white canopy structure with a logo on its edge. The logo consists of a blue square with a white circle inside, followed by the text "Air Liquide" in red. The canopy is supported by a white pillar.

Air Liquide

Thank you for your attention!
#TeamWasserstoff

A large white sign with a blue logo and red text. The logo is a blue square with a white circle inside. The text "Air Liquide" is written in red, slanted upwards to the right.

Air Liquide

**Wasserstoff-
Station**